



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032423\
Data File : VY013069.D
Acq On : 24 Mar 2023 09:43
Operator : KP/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Title : SW846 8260
Last Update : Fri Mar 03 01:13:13 2023



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.3	10048	PASS
75	95	30	60	52.7	26048	PASS
95	95	100	100	100.0	49429	PASS
96	95	5	9	6.5	3212	PASS
173	174	0.00	2	1.5	660	PASS
174	95	50	100	91.2	45064	PASS
175	174	5	9	7.7	3451	PASS
176	174	95	101	95.3	42939	PASS
177	176	5	9	6.8	2941	PASS