

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032425\
 Data File : VY021616.D
 Acq On : 24 Mar 2025 14:28
 Operator : SY/MD
 Sample : Q1619-05
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-1

Quant Time: Mar 25 01:30:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 05 12:42:45 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	2492	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.622	114	3569	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	1721	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.353	152	423	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	1693	64.277	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	= 128.560%	
35) Dibromofluoromethane	7.628	113	1349	57.503	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	= 115.000%	
50) Toluene-d8	10.109	98	3407	38.356	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 76.720%	
62) 4-Bromofluorobenzene	12.414	95	305	10.094	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 20.180%#	
Target Compounds						
52) Toluene	10.176	92	397	5.821	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032425\
Data File : VY021616.D
Acq On : 24 Mar 2025 14:28
Operator : SY/MD
Sample : Q1619-05
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-1

Quant Time: Mar 25 01:30:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 05 12:42:45 2025
Response via : Initial Calibration

