

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032425\
 Data File : VY021617.D
 Acq On : 24 Mar 2025 14:51
 Operator : SY/MD
 Sample : Q1614-01
 Misc : 5.10g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-04-032025RE

Quant Time: Mar 25 01:30:48 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 05 12:42:45 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	2755	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.615	114	4424	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	2655	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	504	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.054	65	2126	73.012	ug/l	-0.01
Spiked Amount	50.000	Range 50 - 163	Recovery	=	146.020%	
35) Dibromofluoromethane	7.628	113	1622	55.778	ug/l	-0.01
Spiked Amount	50.000	Range 54 - 147	Recovery	=	111.560%	
50) Toluene-d8	10.109	98	4755	43.186	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	=	86.380%	
62) 4-Bromofluorobenzene	12.407	95	581	15.513	ug/l	0.00
Spiked Amount	50.000	Range 30 - 143	Recovery	=	31.020%	
Target Compounds						
52) Toluene	10.170	92	474	5.607	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032425\
Data File : VY021617.D
Acq On : 24 Mar 2025 14:51
Operator : SY/MD
Sample : Q1614-01
Misc : 5.10g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-04-032025RE

Quant Time: Mar 25 01:30:48 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 05 12:42:45 2025
Response via : Initial Calibration

