

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032425\
 Data File : VY021618.D
 Acq On : 24 Mar 2025 15:14
 Operator : SY/MD
 Sample : Q1611-01
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RBR200030

Quant Time: Mar 25 01:31:01 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 05 12:42:45 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	5848	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.616	114	7577	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	3621	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	475	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	2940	47.565	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	95.140%
35) Dibromofluoromethane	7.640	113	2210	44.373	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	88.740%
50) Toluene-d8	10.103	98	7655	40.594	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	81.180%
62) 4-Bromofluorobenzene	12.414	95	672	10.476	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	20.960%#
Target Compounds						
52) Toluene	10.164	92	671	4.634	ug/l	Qvalue 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032425\
Data File : VY021618.D
Acq On : 24 Mar 2025 15:14
Operator : SY/MD
Sample : Q1611-01
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RBR200030

Quant Time: Mar 25 01:31:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 05 12:42:45 2025
Response via : Initial Calibration

