

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032425\
 Data File : VY021613.D
 Acq On : 24 Mar 2025 12:49
 Operator : SY/MD
 Sample : VY0324SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0324SBS01

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 03/26/2025
 Supervised By :Semsettin Yesilyurt 03/26/2025

Quant Time: Mar 25 01:29:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 05 12:42:45 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	184347	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	282510	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	248626	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	133143	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	102138	52.420	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 104.840%			
35) Dibromofluoromethane	7.634	113	97858	52.697	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 105.400%			
50) Toluene-d8	10.103	98	361919	51.474	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery = 102.940%			
62) 4-Bromofluorobenzene	12.401	95	126300	52.808	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery = 105.620%			
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.860	85	31517	18.631	ug/l	99
3) Chloromethane	2.068	50	44923	18.835	ug/l	96
4) Vinyl Chloride	2.202	62	53090	20.369	ug/l	100
5) Bromomethane	2.586	94	38269	20.315	ug/l	96
6) Chloroethane	2.726	64	36822	21.378	ug/l	94
7) Trichlorofluoromethane	3.055	101	83973	23.083	ug/l	98
8) Diethyl Ether	3.452	74	20229	20.065	ug/l	98
9) 1,1,2-Trichlorotrifluo...	3.805	101	43457	20.923	ug/l	98
10) Methyl Iodide	4.006	142	39856	18.877	ug/l	98
11) Tert butyl alcohol	4.860	59	13965	96.390	ug/l	95
12) 1,1-Dichloroethene	3.793	96	38466	20.313	ug/l	81
13) Acrolein	3.647	56	8663	87.412	ug/l	93
14) Allyl chloride	4.384	41	61012	19.937	ug/l	98
15) Acrylonitrile	5.055	53	47879	111.562	ug/l	98
16) Acetone	3.866	43	42612	100.270	ug/l	96
17) Carbon Disulfide	4.098	76	94886	15.989	ug/l	98
18) Methyl Acetate	4.384	43	21114	22.392	ug/l	98
19) Methyl tert-butyl Ether	5.110	73	100759	20.993	ug/l	98
20) Methylene Chloride	4.610	84	53205	23.851	ug/l	97
21) trans-1,2-Dichloroethene	5.116	96	41083	19.372	ug/l	95
22) Diisopropyl ether	6.018	45	139390	21.136	ug/l	98
23) Vinyl Acetate	5.957	43	394308	104.762	ug/l	99
24) 1,1-Dichloroethane	5.915	63	81303	20.888	ug/l	99
25) 2-Butanone	6.890	43	62785	106.802	ug/l	96
26) 2,2-Dichloropropane	6.884	77	71770	20.039	ug/l	100
27) cis-1,2-Dichloroethene	6.890	96	49342	20.481	ug/l	98
28) Bromochloromethane	7.244	49	31408	19.247	ug/l	100
29) Tetrahydrofuran	7.262	42	38964	107.363	ug/l	98
30) Chloroform	7.414	83	85602	21.179	ug/l	97
31) Cyclohexane	7.701	56	65123	17.846	ug/l	97
32) 1,1,1-Trichloroethane	7.616	97	77510	20.944	ug/l	100
36) 1,1-Dichloropropene	7.835	75	57155	19.663	ug/l	99
37) Ethyl Acetate	6.981	43	28572	22.239	ug/l	99
38) Carbon Tetrachloride	7.811	117	70796	21.177	ug/l	94
39) Methylcyclohexane	9.109	83	65608	18.401	ug/l	98
40) Benzene	8.079	78	175351	20.479	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032425\
 Data File : VY021613.D
 Acq On : 24 Mar 2025 12:49
 Operator : SY/MD
 Sample : VY0324SBSD01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0324SBSD01

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 03/26/2025
 Supervised By :Semsettin Yesilyurt 03/26/2025

Quant Time: Mar 25 01:29:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 05 12:42:45 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.219	41	14620	20.899	ug/l #	91
42) 1,2-Dichloroethane	8.158	62	51030	21.303	ug/l	99
43) Isopropyl Acetate	8.195	43	54802	21.973	ug/l #	86
44) Trichloroethene	8.859	130	45864	21.196	ug/l	98
45) 1,2-Dichloropropane	9.140	63	43310	21.252	ug/l	98
46) Dibromomethane	9.231	93	25450	22.444	ug/l	99
47) Bromodichloromethane	9.420	83	66516	22.089	ug/l	96
48) Methyl methacrylate	9.219	41	24039	20.990	ug/l	98
49) 1,4-Dioxane	9.231	88	5451	489.768	ug/l	92
51) 4-Methyl-2-Pentanone	9.999	43	142293	110.576	ug/l	100
52) Toluene	10.170	92	112531	20.845	ug/l	98
53) t-1,3-Dichloropropene	10.396	75	57871	21.813	ug/l	98
54) cis-1,3-Dichloropropene	9.853	75	66048	20.969	ug/l	98
55) 1,1,2-Trichloroethane	10.572	97	33471	22.566	ug/l	96
56) Ethyl methacrylate	10.438	69	40599	21.462	ug/l	100
57) 1,3-Dichloropropane	10.713	76	55693	21.664	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.713	63	99079	113.459	ug/l	100
59) 2-Hexanone	10.761	43	94596	111.173	ug/l	98
60) Dibromochloromethane	10.914	129	46157	22.699	ug/l	97
61) 1,2-Dibromoethane	11.017	107	30803	22.004	ug/l	98
64) Tetrachloroethene	10.645	164	49923	24.693	ug/l	98
65) Chlorobenzene	11.438	112	123750	21.077	ug/l	96
66) 1,1,1,2-Tetrachloroethane	11.511	131	45487	21.913	ug/l	99
67) Ethyl Benzene	11.517	91	206296	20.019	ug/l	99
68) m/p-Xylenes	11.627	106	162166	41.450	ug/l	99
69) o-Xylene	11.956	106	73810	20.470	ug/l	100
70) Styrene	11.968	104	126501	21.148	ug/l	100
71) Bromoform	12.133	173	26751	23.054	ug/l #	99
73) Isopropylbenzene	12.255	105	202176	19.277	ug/l	99
74) N-amyl acetate	12.066	43	46349	20.349	ug/l	96
75) 1,1,2,2-Tetrachloroethane	12.505	83	37558	20.293	ug/l	97
76) 1,2,3-Trichloropropane	12.554	75	28212m	21.028	ug/l	
77) Bromobenzene	12.529	156	49536	20.189	ug/l	97
78) n-propylbenzene	12.596	91	250447	19.662	ug/l	100
79) 2-Chlorotoluene	12.676	91	142785	19.605	ug/l	100
80) 1,3,5-Trimethylbenzene	12.737	105	168640	19.786	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.304	75	11709	19.840	ug/l	97
82) 4-Chlorotoluene	12.779	91	151121	20.038	ug/l	99
83) tert-Butylbenzene	12.999	119	152139	20.009	ug/l	99
84) 1,2,4-Trimethylbenzene	13.041	105	169692	20.052	ug/l	99
85) sec-Butylbenzene	13.175	105	227378	20.099	ug/l	100
86) p-Isopropyltoluene	13.291	119	186551	20.016	ug/l	100
87) 1,3-Dichlorobenzene	13.285	146	101479	20.777	ug/l	99
88) 1,4-Dichlorobenzene	13.364	146	98863	20.697	ug/l	99
89) n-Butylbenzene	13.614	91	175942	20.444	ug/l	99
90) Hexachloroethane	13.883	117	40836	20.538	ug/l	99
91) 1,2-Dichlorobenzene	13.657	146	88822	21.218	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.273	75	6232	22.479	ug/l	93
93) 1,2,4-Trichlorobenzene	14.919	180	50930	22.274	ug/l	98
94) Hexachlorobutadiene	15.023	225	33053	23.061	ug/l	98
95) Naphthalene	15.145	128	80284	22.756	ug/l	99
96) 1,2,3-Trichlorobenzene	15.328	180	43386	22.561	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032425\
Data File : VY021613.D
Acq On : 24 Mar 2025 12:49
Operator : SY/MD
Sample : VY0324SBSD01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0324SBSD01


Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 03/26/2025
Supervised By :Semsettin Yesilyurt 03/26/2025

Quant Time: Mar 25 01:29:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030425S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 05 12:42:45 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.707 min Scan# 91

Delta R.T. -0.006 min

Lab File: VY021613.D

Acq: 24 Mar 2025 12:49

Instrument :

MSVOA_Y

ClientSampleId :

VY0324SBS01

Tgt Ion:168 Resp: 18434

Ion Ratio Lower Upper

168 100





99 55.9 46.0 69.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/26/2025

Supervised By :Semsettin Yesilyurt 03/26/2025



#2

Dichlorodifluoromethane

Concen: 18.631 ug/l

RT: 1.860 min Scan# 23

Delta R.T. -0.006 min

Lab File: VY021613.D




Acq: 24 Mar 2025 12:49


Tgt Ion: 85 Resp: 31517

Ion Ratio Lower Upper

85 100


87 30.6 15.7 47.0





#3
Chloromethane
Concen: 18.835 ug/l
RT: 2.068 min Scan# 51
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49




Instrument :
MSVOA_Y
ClientSampleId :
VY0324SBSD01



Tgt Ion: 50 Resp: 4492
Ion Ratio Lower Upper
50 100
52 30.5 26.2 39.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2025
Supervised By :Semsettin Yesilyurt 03/26/2025



#4
Vinyl Chloride
Concen: 20.369 ug/l
RT: 2.202 min Scan# 79
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49


Tgt Ion: 62 Resp: 53090
Ion Ratio Lower Upper
62 100
64 32.8 26.2 39.2





#5
Bromomethane
Concen: 20.315 ug/l
RT: 2.586 min Scan# 145
Delta R.T. -0.018 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49




Instrument :
MSVOA_Y
ClientSampleId :
VY0324SBS01




Tgt Ion: 94 Resp: 38269
Ion Ratio Lower Upper
94 100
96 91.3 76.3 114.5

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 03/26/2025
Supervised By :Semsettin Yesilyurt 03/26/2025




#6
Chloroethane
Concen: 21.378 ug/l
RT: 2.726 min Scan# 165
Delta R.T. -0.012 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49



Tgt Ion: 64 Resp: 36822
Ion Ratio Lower Upper
64 100
66 29.9 26.7 40.1









#7
 Trichlorofluoromethane
 Concen: 23.083 ug/l
 RT: 3.055 min Scan# 219
 Delta R.T. -0.012 min
 Lab File: VY021613.D
 Acq: 24 Mar 2025 12:49

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0324SBSD01

Tgt Ion:101 Resp: 8397
 Ion Ratio Lower Upper
 101 100
 103 66.0 51.7 77.5




Manual Integrations
 APPROVED


Reviewed By :Mahesh Dadoda 03/26/2025
 Supervised By :Semsettin Yesilyurt 03/26/2025



#8
 Diethyl Ether
 Concen: 20.065 ug/l
 RT: 3.452 min Scan# 284
 Delta R.T. -0.006 min
 Lab File: VY021613.D
 Acq: 24 Mar 2025 12:49

Tgt Ion: 74 Resp: 20229
 Ion Ratio Lower Upper
 74 100
 45 97.3 49.7 149.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 20.923 ug/l
RT: 3.805 min Scan# 34
Delta R.T. -0.025 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49

Instrument :
MSVOA_Y
ClientSampleId :
VY0324SBS01

Tgt Ion:101 Resp: 4345
Ion Ratio Lower Upper
101 100
85 46.8 37.1 55.7
151 79.5 62.1 93.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2025
Supervised By :Semsettin Yesilyurt 03/26/2025



#10
Methyl Iodide
Concen: 18.877 ug/l
RT: 4.006 min Scan# 375
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49


Tgt Ion:142 Resp: 39856
Ion Ratio Lower Upper
142 100
127 47.6 39.1 58.7
141 15.0 11.8 17.6





#11
Tert butyl alcohol
Concen: 96.390 ug/l
RT: 4.860 min Scan# 517
Delta R.T. -0.012 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49




Instrument :
MSVOA_Y
ClientSampleId :
VY0324SBS01



Tgt Ion: 59 Resp: 1396
Ion Ratio Lower Upper
59 100
57 8.3 8.2 12.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2025
Supervised By :Semsettin Yesilyurt 03/26/2025



#12
1,1-Dichloroethene
Concen: 20.313 ug/l
RT: 3.793 min Scan# 340
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49


Tgt Ion: 96 Resp: 38466
Ion Ratio Lower Upper
96 100
61 139.5 135.2 202.8
98 54.0 51.4 77.2





#13
Acrolein
Concen: 87.412 ug/l
RT: 3.647 min Scan# 316
Delta R.T. -0.018 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49




Instrument :
MSVOA_Y
ClientSampleId :
VY0324SBS01



Tgt Ion: 56 Resp: 866
Ion Ratio Lower Upper
56 100
55 73.1 63.4 95.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2025
Supervised By :Semsettin Yesilyurt 03/26/2025



#14
Allyl chloride
Concen: 19.937 ug/l
RT: 4.384 min Scan# 437
Delta R.T. -0.012 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49


Tgt Ion: 41 Resp: 61012
Ion Ratio Lower Upper
41 100
39 72.6 59.3 88.9
76 37.9 29.8 44.8





#15
Acrylonitrile
Concen: 111.562 ug/l
RT: 5.055 min Scan# 549
Delta R.T. -0.012 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49




Instrument :
MSVOA_Y
ClientSampleId :
VY0324SBS01



Tgt Ion: 53 Resp: 47879
Ion Ratio Lower Upper
53 100
52 80.8 66.3 99.5
51 37.2 30.3 45.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2025
Supervised By :Semsettin Yesilyurt 03/26/2025



#16
Acetone
Concen: 100.270 ug/l
RT: 3.866 min Scan# 352
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49


Tgt Ion: 43 Resp: 42612
Ion Ratio Lower Upper
43 100
58 29.5 22.1 33.1





#17
Carbon Disulfide
Concen: 15.989 ug/l
RT: 4.098 min Scan# 391
Delta R.T. -0.018 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49




Instrument :
MSVOA_Y
ClientSampleId :
VY0324SBSD01



Tgt Ion: 76 Resp: 9488
Ion Ratio Lower Upper
76 100
78 10.1 7.4 11.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2025
Supervised By :Semsettin Yesilyurt 03/26/2025



#18
Methyl Acetate
Concen: 22.392 ug/l
RT: 4.384 min Scan# 437
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49


Tgt Ion: 43 Resp: 21114
Ion Ratio Lower Upper
43 100
74 25.0 19.4 29.0





#19
Methyl tert-butyl Ether
Concen: 20.993 ug/l
RT: 5.110 min Scan# 5110
Delta R.T. -0.012 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49




Instrument :
MSVOA_Y
ClientSampleId :
VY0324SBS01



Tgt Ion: 73 Resp: 100759
Ion Ratio Lower Upper
73 100
57 23.5 17.9 26.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2025
Supervised By :Semsettin Yesilyurt 03/26/2025



#20
Methylene Chloride
Concen: 23.851 ug/l
RT: 4.610 min Scan# 474
Delta R.T. -0.012 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49


Tgt Ion: 84 Resp: 53205
Ion Ratio Lower Upper
84 100
49 131.3 102.0 153.0
51 37.6 33.0 49.6
86 64.8 50.5 75.7





#21
trans-1,2-Dichloroethene
Concen: 19.372 ug/l
RT: 5.116 min Scan# 510
Delta R.T. -0.012 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49




Instrument :
MSVOA_Y
ClientSampleId :
VY0324SBSD01



Tgt Ion: 96 Resp: 4108
Ion Ratio Lower Upper
96 100
61 143.0 111.5 167.3
98 60.5 53.8 80.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2025
Supervised By :Semsettin Yesilyurt 03/26/2025



#22
Diisopropyl ether
Concen: 21.136 ug/l
RT: 6.018 min Scan# 705
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49

Tgt Ion: 45 Resp: 139390
Ion Ratio Lower Upper
45 100
43 59.4 48.7 73.1
87 27.6 21.5 32.3
59 11.6 9.3 13.9









#23
 Vinyl Acetate
 Concen: 104.762 ug/l
 RT: 5.957 min Scan# 697
 Delta R.T. -0.012 min
 Lab File: VY021613.D
 Acq: 24 Mar 2025 12:49

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0324SBS01

Tgt Ion: 43 Resp: 394308
 Ion Ratio Lower Upper
 43 100
 86 10.1 8.5 12.7




Manual Integrations
 APPROVED


Reviewed By :Mahesh Dadoda 03/26/2025
 Supervised By :Semsettin Yesilyurt 03/26/2025



#24
 1,1-Dichloroethane
 Concen: 20.888 ug/l
 RT: 5.915 min Scan# 688
 Delta R.T. -0.012 min
 Lab File: VY021613.D
 Acq: 24 Mar 2025 12:49

Tgt Ion: 63 Resp: 81303
 Ion Ratio Lower Upper
 63 100
 98 7.3 3.3 9.9
 100 4.5 2.1 6.5









#25
2-Butanone
Concen: 106.802 ug/l
RT: 6.890 min Scan# 849
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49

Instrument :
MSVOA_Y
ClientSampleId :
VY0324SBSD01

Tgt Ion: 43 Resp: 6278
Ion Ratio Lower Upper
43 100
72 24.7 21.4 32.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2025
Supervised By :Semsettin Yesilyurt 03/26/2025



#26
2,2-Dichloropropane
Concen: 20.039 ug/l
RT: 6.884 min Scan# 847
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49


Tgt Ion: 77 Resp: 71770
Ion Ratio Lower Upper
77 100
97 22.6 11.3 33.8





#27
 cis-1,2-Dichloroethene
 Concen: 20.481 ug/l
 RT: 6.890 min Scan# 849
 Delta R.T. -0.006 min
 Lab File: VY021613.D
 Acq: 24 Mar 2025 12:49




Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0324SBSD01



Tgt Ion: 96 Resp: 4934
 Ion Ratio Lower Upper
 96 100
 61 146.6 0.0 300.8
 98 63.9 0.0 130.0




Manual Integrations
 APPROVED


Reviewed By :Mahesh Dadoda 03/26/2025
 Supervised By :Semsettin Yesilyurt 03/26/2025



#28
 Bromochloromethane
 Concen: 19.247 ug/l
 RT: 7.244 min Scan# 906
 Delta R.T. -0.006 min
 Lab File: VY021613.D
 Acq: 24 Mar 2025 12:49

Tgt Ion: 49 Resp: 31408
 Ion Ratio Lower Upper
 49 100
 129 1.8 0.0 4.2
 130 74.3 59.6 89.4





#29

Tetrahydrofuran

Concen: 107.363 ug/l

RT: 7.262 min Scan# 909

Delta R.T. -0.006 min

Lab File: VY021613.D

Acq: 24 Mar 2025 12:49

Instrument :

MSVOA_Y

ClientSampleId :





VY0324SBSD01

Tgt Ion:	42	Resp:	3896
Ion Ratio	Lower	Upper	
42	100		
72	42.5	34.8	52.2
71	39.3	32.6	48.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/26/2025
Supervised By :Semsettin Yesilyurt 03/26/2025



#30

Chloroform

Concen: 21.179 ug/l




RT: 7.414 min Scan# 934


Delta R.T. -0.012 min

Lab File: VY021613.D

Acq: 24 Mar 2025 12:49


Tgt Ion:	83	Resp:	85602
Ion Ratio	Lower	Upper	
83	100		
85	64.0	53.1	79.7





#31
Cyclohexane
Concen: 17.846 ug/l
RT: 7.701 min Scan# 982
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49




Instrument :
MSVOA_Y
ClientSampleId :
VY0324SBSD01



Tgt Ion: 56 Resp: 6512
Ion Ratio Lower Upper
56 100
69 38.4 27.0 40.4
84 87.6 71.1 106.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2025
Supervised By :Semsettin Yesilyurt 03/26/2025



#32
1,1,1-Trichloroethane
Concen: 20.944 ug/l
RT: 7.616 min Scan# 967
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49


Tgt Ion: 97 Resp: 77510
Ion Ratio Lower Upper
97 100
99 64.6 51.7 77.5
61 44.1 35.7 53.5





#33
1,2-Dichloroethane-d4
Concen: 52.420 ug/l
RT: 8.061 min Scan# 1041
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49




Instrument :
MSVOA_Y
ClientSampleId :
VY0324SBS01




Tgt Ion: 65 Resp: 102138
Ion Ratio Lower Upper
65 100
67 52.5 0.0 102.8

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 03/26/2025
Supervised By :Semsettin Yesilyurt 03/26/2025




#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.615 min Scan# 1131
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49




Tgt Ion:114 Resp: 282510
Ion Ratio Lower Upper
114 100
63 20.3 0.0 40.8
88 15.7 0.0 30.8





#35
Dibromofluoromethane
Concen: 52.697 ug/l
RT: 7.634 min Scan# 91
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49




Instrument :
MSVOA_Y
ClientSampleId :
VY0324SBS01




Tgt Ion:113 Resp: 9785
Ion Ratio Lower Upper
113 100
111 102.9 82.0 123.0
192 19.7 15.9 23.9

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 03/26/2025
Supervised By :Semsettin Yesilyurt 03/26/2025




#36
1,1-Dichloropropene
Concen: 19.663 ug/l
RT: 7.835 min Scan# 1003
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49




Tgt Ion: 75 Resp: 57155
Ion Ratio Lower Upper
75 100
110 34.9 17.3 51.7
77 32.4 25.4 38.2





#37
Ethyl Acetate
Concen: 22.239 ug/l
RT: 6.981 min Scan# 801
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49




Instrument :
MSVOA_Y
ClientSampleId :
VY0324SBS01



Tgt Ion: 43 Resp: 28572
Ion Ratio Lower Upper
43 100
61 13.7 11.4 17.0
70 10.4 8.2 12.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2025
Supervised By :Semsettin Yesilyurt 03/26/2025



#38
Carbon Tetrachloride
Concen: 21.177 ug/l
RT: 7.811 min Scan# 999
Delta R.T. -0.012 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49


Tgt Ion:117 Resp: 70796
Ion Ratio Lower Upper
117 100
119 87.9 75.8 113.8
121 27.5 23.9 35.9





#39
Methylcyclohexane
Concen: 18.401 ug/l
RT: 9.109 min Scan# 11
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49




Instrument :
MSVOA_Y
ClientSampleId :
VY0324SBSD01



Tgt Ion: 83 Resp: 65608
Ion Ratio Lower Upper
83 100
55 74.3 60.7 91.1
98 47.0 38.6 57.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2025
Supervised By :Semsettin Yesilyurt 03/26/2025



#40
Benzene
Concen: 20.479 ug/l
RT: 8.079 min Scan# 1043
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49

Tgt Ion: 78 Resp: 175351
Ion Ratio Lower Upper
78 100
77 23.6 19.0 28.4









#41
Methacrylonitrile
Concen: 20.899 ug/l
RT: 7.219 min Scan# 902
Delta R.T. 0.000 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49

Instrument :
MSVOA_Y
ClientSampleId :
VY0324SBS01

Tgt Ion:	41	Resp:	14620
Ion Ratio	Lower	Upper	
41	100		
39	61.9	35.0	52.6
67	76.1	61.2	91.8
52	35.1	26.8	40.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2025
Supervised By :Semsettin Yesilyurt 03/26/2025



#42
1,2-Dichloroethane
Concen: 21.303 ug/l
RT: 8.158 min Scan# 1056
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49

Tgt Ion:	62	Resp:	51030
Ion Ratio	Lower	Upper	
62	100		
98	9.6	0.0	18.2









#43
Isopropyl Acetate
Concen: 21.973 ug/l
RT: 8.195 min Scan# 1062
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49

Instrument : MSVOA_Y
ClientSampleId : VY0324SBSD01

Tgt Ion: 43 Resp: 5480
Ion Ratio Lower Upper
43 100
61 18.5 23.4 35.0
87 12.5 10.5 15.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2025
Supervised By :Semsettin Yesilyurt 03/26/2025



#44
Trichloroethene
Concen: 21.196 ug/l
RT: 8.859 min Scan# 1171
Delta R.T. -0.012 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49

Tgt Ion:130 Resp: 45864
Ion Ratio Lower Upper
130 100
95 97.6 0.0 191.6





#45

1,2-Dichloropropane

Concen: 21.252 ug/l

RT: 9.140 min Scan# 1218

Delta R.T. -0.006 min

Lab File: VY021613.D

Acq: 24 Mar 2025 12:49

Instrument :

MSVOA_Y

ClientSampleId :

VY0324SBS01

Tgt Ion: 63 Resp: 43310

Ion Ratio Lower Upper

63 100





65 31.2 26.0 39.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/26/2025

Supervised By :Semsettin Yesilyurt 03/26/2025



#46

Dibromomethane

Concen: 22.444 ug/l

RT: 9.231 min Scan# 1232

Delta R.T. -0.006 min

Lab File: VY021613.D

Acq: 24 Mar 2025 12:49




Tgt Ion: 93 Resp: 25450


Ion Ratio Lower Upper

93 100

95 79.9 65.8 98.8

174 97.2 77.9 116.9









#47
Bromodichloromethane
Concen: 22.089 ug/l
RT: 9.420 min Scan# 1264
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49

Instrument :
MSVOA_Y
ClientSampleId :
VY0324SBS01

Tgt Ion: 83 Resp: 66510
Ion Ratio Lower Upper
83 100
85 60.8 51.1 76.7
127 8.9 6.7 10.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2025
Supervised By :Semsettin Yesilyurt 03/26/2025



#48
Methyl methacrylate
Concen: 20.990 ug/l
RT: 9.219 min Scan# 1230
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49


Tgt Ion: 41 Resp: 24039
Ion Ratio Lower Upper
41 100
69 85.3 67.8 101.8
39 61.0 46.4 69.6





#49
1,4-Dioxane
Concen: 489.768 ug/l
RT: 9.231 min Scan# 1232
Delta R.T. 0.000 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49




Instrument :
MSVOA_Y
ClientSampleId :
VY0324SBSD01



Tgt Ion: 88 Resp: 545
Ion Ratio Lower Upper
88 100
43 33.5 25.0 37.4
58 64.5 58.4 87.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2025
Supervised By :Semsettin Yesilyurt 03/26/2025



#50
Toluene-d8
Concen: 51.474 ug/l
RT: 10.103 min Scan# 1375
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49

Tgt Ion: 98 Resp: 361919
Ion Ratio Lower Upper
98 100
100 64.5 52.1 78.1









#51
4-Methyl-2-Pentanone
Concen: 110.576 ug/l
RT: 9.999 min Scan# 1358
Delta R.T. 0.000 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49

Instrument :
MSVOA_Y
ClientSampleId :
VY0324SBS01

Tgt Ion: 43 Resp: 14229
Ion Ratio Lower Upper
43 100
58 39.2 31.4 47.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2025
Supervised By :Semsettin Yesilyurt 03/26/2025



#52
Toluene
Concen: 20.845 ug/l
RT: 10.170 min Scan# 1386
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49

Tgt Ion: 92 Resp: 112531
Ion Ratio Lower Upper
92 100
91 175.0 138.2 207.2









#53
 t-1,3-Dichloropropene
 Concen: 21.813 ug/l
 RT: 10.396 min Scan# 1423
 Delta R.T. 0.000 min
 Lab File: VY021613.D
 Acq: 24 Mar 2025 12:49

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0324SBS01

Tgt Ion: 75 Resp: 5787
 Ion Ratio Lower Upper
 75 100
 77 30.2 25.0 37.6




Manual Integrations
 APPROVED


Reviewed By :Mahesh Dadoda 03/26/2025
 Supervised By :Semsettin Yesilyurt 03/26/2025



#54
 cis-1,3-Dichloropropene
 Concen: 20.969 ug/l
 RT: 9.853 min Scan# 1334
 Delta R.T. -0.006 min
 Lab File: VY021613.D
 Acq: 24 Mar 2025 12:49


Tgt Ion: 75 Resp: 66048
 Ion Ratio Lower Upper
 75 100
 77 31.9 24.6 37.0
 39 48.0 39.4 59.2





#55
1,1,2-Trichloroethane
Concen: 22.566 ug/l
RT: 10.572 min Scan# 1452
Delta R.T. 0.000 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49




Instrument :
MSVOA_Y
ClientSampleId :
VY0324SBSD01




Tgt Ion	Ratio	Lower	Upper
97	100		
83	87.0	70.0	105.0
85	60.6	44.3	66.5
99	59.2	51.4	77.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 03/26/2025
Supervised By :Semsettin Yesilyurt 03/26/2025




#56
Ethyl methacrylate
Concen: 21.462 ug/l
RT: 10.438 min Scan# 1430
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49



Tgt Ion	Ratio	Lower	Upper
69	100		
41	66.8	53.3	79.9
39	38.6	31.0	46.6





#57

1,3-Dichloropropane

Concen: 21.664 ug/l

RT: 10.713 min Scan# 1476

Delta R.T. -0.006 min

Lab File: VY021613.D

Acq: 24 Mar 2025 12:49

Instrument :

MSVOA_Y

ClientSampleId :

VY0324SBS01

Tgt Ion: 76 Resp: 5569

Ion Ratio Lower Upper

76 100





78 32.2 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/26/2025

Supervised By :Semsettin Yesilyurt 03/26/2025



#58

2-Chloroethyl Vinyl ether

Concen: 113.459 ug/l

RT: 9.713 min Scan# 1311

Delta R.T. 0.000 min

Lab File: VY021613.D




Acq: 24 Mar 2025 12:49


Tgt Ion: 63 Resp: 99079

Ion Ratio Lower Upper

63 100


106 27.8 22.1 33.1





#59
2-Hexanone
Concen: 111.173 ug/l
RT: 10.761 min Scan# 1483
Delta R.T. 0.000 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49




Instrument :
MSVOA_Y
ClientSampleId :
VY0324SBSD01




Tgt Ion: 43 Resp: 94590
Ion Ratio Lower Upper
43 100
58 53.4 27.3 81.8

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 03/26/2025
Supervised By :Semsettin Yesilyurt 03/26/2025




#60
Dibromochloromethane
Concen: 22.699 ug/l
RT: 10.914 min Scan# 1508
Delta R.T. 0.000 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49




Tgt Ion:129 Resp: 46157
Ion Ratio Lower Upper
129 100
127 75.6 39.1 117.3





#61
1,2-Dibromoethane
Concen: 22.004 ug/l
RT: 11.017 min Scan# 1525
Delta R.T. 0.000 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49




Instrument :
MSVOA_Y
ClientSampleId :
VY0324SBS01




Tgt Ion:107 Resp: 3080
Ion Ratio Lower Upper
107 100
109 92.4 75.8 113.8

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 03/26/2025
Supervised By :Semsettin Yesilyurt 03/26/2025




#62
4-Bromofluorobenzene
Concen: 52.808 ug/l
RT: 12.401 min Scan# 1752
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49




Tgt Ion: 95 Resp: 126300
Ion Ratio Lower Upper
95 100
174 82.9 0.0 165.0
176 79.3 0.0 160.0





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.414 min Scan# 11
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49




Instrument :
MSVOA_Y
ClientSampleId :
VY0324SBSD01



Tgt Ion:117 Resp: 248620
Ion Ratio Lower Upper
117 100
82 55.7 44.6 67.0
119 32.6 25.4 38.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2025
Supervised By :Semsettin Yesilyurt 03/26/2025



#64
Tetrachloroethene
Concen: 24.693 ug/l
RT: 10.645 min Scan# 1464
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49


Tgt Ion:164 Resp: 49923
Ion Ratio Lower Upper
164 100
166 124.6 100.4 150.6
129 89.4 73.8 110.8
131 91.1 71.5 107.3





#65
Chlorobenzene
Concen: 21.077 ug/l
RT: 11.438 min Scan# 1594
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49




Instrument :
MSVOA_Y
ClientSampleId :
VY0324SBSD01



Tgt Ion:112 Resp: 123750
Ion Ratio Lower Upper
112 100
114 30.5 26.1 39.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2025
Supervised By :Semsettin Yesilyurt 03/26/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 21.913 ug/l
RT: 11.511 min Scan# 1606
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49

Tgt Ion:131 Resp: 45487
Ion Ratio Lower Upper
131 100
133 95.7 47.6 142.9
119 63.0 32.3 96.9









#67
Ethyl Benzene
Concen: 20.019 ug/l
RT: 11.517 min Scan# 1607
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49

Instrument :
MSVOA_Y
ClientSampleId :
VY0324SBSD01

Tgt Ion: 91 Resp: 206290
Ion Ratio Lower Upper
91 100
106 31.0 24.6 36.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2025
Supervised By :Semsettin Yesilyurt 03/26/2025



#68
m/p-Xylenes
Concen: 41.450 ug/l
RT: 11.627 min Scan# 1625
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49

Tgt Ion:106 Resp: 162166
Ion Ratio Lower Upper
106 100
91 200.9 161.7 242.5









#69
o-Xylene
Concen: 20.470 ug/l
RT: 11.956 min Scan# 1679
Delta R.T. 0.000 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49

Instrument :
MSVOA_Y
ClientSampleId :
VY0324SBSD01

Tgt Ion:106 Resp: 73810
Ion Ratio Lower Upper
106 100
91 215.5 107.6 322.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2025
Supervised By :Semsettin Yesilyurt 03/26/2025



#70
Styrene
Concen: 21.148 ug/l
RT: 11.968 min Scan# 1681
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49


Tgt Ion:104 Resp: 126501
Ion Ratio Lower Upper
104 100
78 50.2 40.7 61.1
103 55.1 44.1 66.1





#71
Bromoform
Concen: 23.054 ug/l
RT: 12.133 min Scan# 1708
Delta R.T. 0.000 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49




Instrument :
MSVOA_Y
ClientSampleId :
VY0324SBS01




Tgt Ion:173 Resp: 2675
Ion Ratio Lower Upper
173 100
175 48.7 23.9 71.7
254 0.0 0.1 0.1

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 03/26/2025
Supervised By :Semsettin Yesilyurt 03/26/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.346 min Scan# 1907
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49




Tgt Ion:152 Resp: 133143
Ion Ratio Lower Upper
152 100
115 57.7 29.0 87.0
150 166.0 0.0 347.2





#73
Isopropylbenzene
Concen: 19.277 ug/l
RT: 12.255 min Scan# 1728
Delta R.T. 0.000 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49




Instrument :
MSVOA_Y
ClientSampleId :
VY0324SBSD01




Tgt Ion: 105 Resp: 202170
Ion Ratio Lower Upper
105 100
120 26.6 13.1 39.3

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 03/26/2025
Supervised By :Semsettin Yesilyurt 03/26/2025




#74
N-ethyl acetate
Concen: 20.349 ug/l
RT: 12.066 min Scan# 1697
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49




Tgt Ion: 43 Resp: 46349
Ion Ratio Lower Upper
43 100
70 45.6 34.6 51.8
55 29.4 21.1 31.7
61 23.5 19.4 29.0





#75
1,1,2,2-Tetrachloroethane
Concen: 20.293 ug/l
RT: 12.505 min Scan# 1769
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49




Instrument :
MSVOA_Y
ClientSampleId :
VY0324SBSD01



Tgt Ion: 83 Resp: 3755
Ion Ratio Lower Upper
83 100
131 10.0 5.3 16.1
85 65.9 31.9 95.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2025
Supervised By :Semsettin Yesilyurt 03/26/2025



#76
1,2,3-Trichloropropane
Concen: 21.028 ug/l m
RT: 12.554 min Scan# 1777
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49


Tgt Ion: 75 Resp: 28212
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0





#77
Bromobenzene
Concen: 20.189 ug/l
RT: 12.529 min Scan# 1773
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49




Instrument :
MSVOA_Y
ClientSampleId :
VY0324SBS01



Tgt Ion: 156 Resp: 49536
Ion Ratio Lower Upper
156 100
77 181.4 93.6 280.8
158 95.9 48.3 144.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2025
Supervised By :Semsettin Yesilyurt 03/26/2025



#78
n-propylbenzene
Concen: 19.662 ug/l
RT: 12.596 min Scan# 1784
Delta R.T. 0.000 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49

Tgt Ion: 91 Resp: 250447
Ion Ratio Lower Upper
91 100
120 22.7 11.3 33.8









#79
2-Chlorotoluene
Concen: 19.605 ug/l
RT: 12.676 min Scan# 1798
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49

Instrument :
MSVOA_Y
ClientSampleId :
VY0324SBS01

Tgt Ion: 91 Resp: 142785
Ion Ratio Lower Upper
91 100
126 34.1 17.2 51.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2025
Supervised By :Semsettin Yesilyurt 03/26/2025



#80
1,3,5-Trimethylbenzene
Concen: 19.786 ug/l
RT: 12.737 min Scan# 1807
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49


Tgt Ion:105 Resp: 168640
Ion Ratio Lower Upper
105 100
120 49.2 24.6 73.6





#81
trans-1,4-Dichloro-2-butene
Concen: 19.840 ug/l
RT: 12.304 min Scan# 1736
Delta R.T. 0.000 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49




Instrument :
MSVOA_Y
ClientSampleId :
VY0324SBS01



Tgt Ion: 75 Resp: 11709
Ion Ratio Lower Upper
75 100
53 98.0 81.0 121.4
89 47.6 36.9 55.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2025
Supervised By :Semsettin Yesilyurt 03/26/2025



#82
4-Chlorotoluene
Concen: 20.038 ug/l
RT: 12.779 min Scan# 1814
Delta R.T. 0.000 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49


Tgt Ion: 91 Resp: 151121
Ion Ratio Lower Upper
91 100
126 33.7 17.0 51.0





#83
 tert-Butylbenzene
 Concen: 20.009 ug/l
 RT: 12.999 min Scan# 1850
 Delta R.T. 0.000 min
 Lab File: VY021613.D
 Acq: 24 Mar 2025 12:49




Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0324SBSD01



Tgt Ion:119 Resp: 152139
 Ion Ratio Lower Upper
 119 100
 91 65.3 33.2 99.6
 134 26.2 13.5 40.5




Manual Integrations
 APPROVED


Reviewed By :Mahesh Dadoda 03/26/2025
 Supervised By :Semsettin Yesilyurt 03/26/2025



#84
 1,2,4-Trimethylbenzene
 Concen: 20.052 ug/l
 RT: 13.041 min Scan# 1857
 Delta R.T. -0.006 min
 Lab File: VY021613.D
 Acq: 24 Mar 2025 12:49

Tgt Ion:105 Resp: 169692
 Ion Ratio Lower Upper
 105 100
 120 45.0 22.7 68.1









#85
 sec-Butylbenzene
 Concen: 20.099 ug/l
 RT: 13.175 min Scan# 1879
 Delta R.T. 0.000 min
 Lab File: VY021613.D
 Acq: 24 Mar 2025 12:49

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0324SBSD01

Tgt Ion:105 Resp: 227378
 Ion Ratio Lower Upper
 105 100
 134 19.9 9.9 29.7




Manual Integrations
 APPROVED


Reviewed By :Mahesh Dadoda 03/26/2025
 Supervised By :Semsettin Yesilyurt 03/26/2025



#86
 p-Isopropyltoluene
 Concen: 20.016 ug/l
 RT: 13.291 min Scan# 1898
 Delta R.T. -0.006 min
 Lab File: VY021613.D
 Acq: 24 Mar 2025 12:49


Tgt Ion:119 Resp: 186551
 Ion Ratio Lower Upper
 119 100
 134 26.2 13.1 39.1
 91 24.6 12.2 36.6





#87
1,3-Dichlorobenzene
Concen: 20.777 ug/l
RT: 13.285 min Scan# 1898
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49




Instrument : MSVOA_Y
ClientSampleId : VY0324SBSD01



Tgt Ion:146 Resp: 101479
Ion Ratio Lower Upper
146 100
111 39.1 20.0 60.0
148 63.0 31.8 95.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2025
Supervised By :Semsettin Yesilyurt 03/26/2025



#88
1,4-Dichlorobenzene
Concen: 20.697 ug/l
RT: 13.364 min Scan# 1910
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49

Tgt Ion:146 Resp: 98863
Ion Ratio Lower Upper
146 100
111 40.1 19.6 58.8
148 65.1 32.3 96.8





#89

n-Butylbenzene

Concen: 20.444 ug/l

RT: 13.614 min Scan# 1951

Delta R.T. -0.006 min

Lab File: VY021613.D

Acq: 24 Mar 2025 12:49

Instrument :

MSVOA_Y

ClientSampleId :





VY0324SBS01

Tgt Ion: 91	Resp: 17594
Ion Ratio	Lower Upper
91	100
92	51.4 26.1 78.3
134	25.1 12.6 37.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/26/2025
Supervised By :Semsettin Yesilyurt 03/26/2025



#90

Hexachloroethane

Concen: 20.538 ug/l




RT: 13.883 min Scan# 1995


Delta R.T. 0.000 min

Lab File: VY021613.D

Acq: 24 Mar 2025 12:49


Tgt Ion:117	Resp: 40836
Ion Ratio	Lower Upper
117	100
201	76.0 37.4 112.2





#91
 1,2-Dichlorobenzene
 Concen: 21.218 ug/l
 RT: 13.657 min Scan# 1958
 Delta R.T. -0.006 min
 Lab File: VY021613.D
 Acq: 24 Mar 2025 12:49




Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0324SBS01




Tgt Ion:146 Resp: 8882
 Ion Ratio Lower Upper
 146 100
 111 41.5 20.7 62.1
 148 62.5 32.1 96.5

Manual Integrations
 APPROVED



Reviewed By :Mahesh Dadoda 03/26/2025
 Supervised By :Semsettin Yesilyurt 03/26/2025




#92
 1,2-Dibromo-3-Chloropropane
 Concen: 22.479 ug/l
 RT: 14.273 min Scan# 2059
 Delta R.T. -0.006 min
 Lab File: VY021613.D
 Acq: 24 Mar 2025 12:49




Tgt Ion: 75 Resp: 6232
 Ion Ratio Lower Upper
 75 100
 155 77.1 43.5 130.5
 157 104.4 54.4 163.2





#93
 1,2,4-Trichlorobenzene
 Concen: 22.274 ug/l
 RT: 14.919 min Scan# 2165
 Delta R.T. -0.006 min
 Lab File: VY021613.D
 Acq: 24 Mar 2025 12:49




Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0324SBS01




Tgt Ion:180 Resp: 50930
 Ion Ratio Lower Upper
 180 100
 182 91.7 46.7 140.0
 145 29.2 15.6 46.8

Manual Integrations
 APPROVED



Reviewed By :Mahesh Dadoda 03/26/2025
 Supervised By :Semsettin Yesilyurt 03/26/2025




#94
 Hexachlorobutadiene
 Concen: 23.061 ug/l
 RT: 15.023 min Scan# 2182
 Delta R.T. -0.006 min
 Lab File: VY021613.D
 Acq: 24 Mar 2025 12:49



Tgt Ion:225 Resp: 33053
 Ion Ratio Lower Upper
 225 100
 223 61.9 31.4 94.2
 227 60.7 31.8 95.4









#95
Naphthalene
Concen: 22.756 ug/l
RT: 15.145 min Scan# 21
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49

Instrument :
MSVOA_Y
ClientSampleId :
VY0324SBSD01

Tgt Ion:128 Resp: 80284
Ion Ratio Lower Upper
128 100
127 13.0 10.6 15.8
129 11.6 8.6 13.0

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 03/26/2025
Supervised By :Semsettin Yesilyurt 03/26/2025



#96
1,2,3-Trichlorobenzene
Concen: 22.561 ug/l
RT: 15.328 min Scan# 2232
Delta R.T. -0.006 min
Lab File: VY021613.D
Acq: 24 Mar 2025 12:49

Tgt Ion:180 Resp: 43386
Ion Ratio Lower Upper
180 100
182 94.2 47.2 141.6
145 33.1 16.4 49.1

