

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032521\
 Data File : VY004202.D
 Acq On : 25 Mar 2021 15:35
 Operator : SY/MD
 Sample : M1812-05
 Misc : 7.32g/5mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-3

Quant Time: Mar 26 01:14:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030821S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 08 12:28:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	186270	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	303761	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	269576	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	125452	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	104196	56.68	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.36%
35) Dibromofluoromethane	7.728	113	94637	55.25	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.50%
50) Toluene-d8	10.185	98	358183	53.34	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.68%
62) 4-Bromofluorobenzene	12.483	95	122927	49.56	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	99.12%
Target Compounds						
16) Acetone	3.954	43	4603	10.37	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032521\
Data File : VY004202.D
Acq On : 25 Mar 2021 15:35
Operator : SY/MD
Sample : M1812-05
Misc : 7.32g/5mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-3

Quant Time: Mar 26 01:14:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030821S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 08 12:28:55 2021
Response via : Initial Calibration

