

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032521\
 Data File : VY004203.D
 Acq On : 25 Mar 2021 15:58
 Operator : SY/MD
 Sample : M1812-01RE
 Misc : 6.82g/5mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-1RE

Quant Time: Mar 26 01:14:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030821S.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 08 12:28:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	184832	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	298519	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	270612	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	120484	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	101918	55.88	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 111.76%	
35) Dibromofluoromethane	7.728	113	94029	55.86	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 111.72%	
50) Toluene-d8	10.185	98	356443	54.01	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 108.02%	
62) 4-Bromofluorobenzene	12.483	95	122005	50.06	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 100.12%	
Target Compounds						
						Qvalue
16) Acetone	3.972	43	3301	7.50	ug/l	91
25) 2-Butanone	6.996	43	4669	7.36	ug/l #	83
29) Tetrahydrofuran	7.356	42	2387	5.79	ug/l #	64

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032521\
Data File : VY004203.D
Acq On : 25 Mar 2021 15:58
Operator : SY/MD
Sample : M1812-01RE
Misc : 6.82g/5mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-1RE

Quant Time: Mar 26 01:14:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030821S.M
Quant Title : SW846 8260
QLast Update : Mon Mar 08 12:28:55 2021
Response via : Initial Calibration

