



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032521\
Data File : VY004193.D
Acq On : 25 Mar 2021 09:20
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030821S.M
Title : SW846 8260
Last Update : Mon Mar 08 12:28:55 2021



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	19.6	15670	PASS
75	95	30	60	53.4	42787	PASS
95	95	100	100	100.0	80107	PASS
96	95	5	9	6.6	5300	PASS
173	174	0.00	2	1.0	677	PASS
174	95	50	100	84.1	67397	PASS
175	174	5	9	7.4	4985	PASS
176	174	95	101	96.7	65181	PASS
177	176	5	9	6.9	4482	PASS