

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032524\
 Data File : VY017750.D
 Acq On : 25 Mar 2024 12:14
 Operator : SY/MD
 Sample : P1842-01RE
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 EO-01-03222024-ARE

Quant Time: Mar 26 03:34:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032224S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 23 04:03:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	13677	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.704	114	18591	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.508	117	11703	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	2350	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.161	65	4118	34.561	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	69.120%
35) Dibromofluoromethane	7.740	113	4347	39.117	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	78.240%
50) Toluene-d8	10.191	98	19590	46.060	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	92.120%
62) 4-Bromofluorobenzene	12.496	95	3121	24.437	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	48.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032524\
Data File : VY017750.D
Acq On : 25 Mar 2024 12:14
Operator : SY/MD
Sample : P1842-01RE
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
EO-01-03222024-ARE

Quant Time: Mar 26 03:34:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032224S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 23 04:03:52 2024
Response via : Initial Calibration

