

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032524\
 Data File : VY017751.D
 Acq On : 25 Mar 2024 12:37
 Operator : SY/MD
 Sample : P1847-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SU-02-032224

Quant Time: Mar 26 03:34:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032224S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 23 04:03:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	143386	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.703	114	264214	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	224526	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	77274	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.161	65	73077	58.501	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 117.000%			
35) Dibromofluoromethane	7.734	113	84528	53.521	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 107.040%			
50) Toluene-d8	10.197	98	300623	49.734	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery = 99.460%			
62) 4-Bromofluorobenzene	12.495	95	76833	42.331	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery = 84.660%			
Target Compounds						
20) Methylene Chloride	4.741	84	12727	8.009	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032524\
Data File : VY017751.D
Acq On : 25 Mar 2024 12:37
Operator : SY/MD
Sample : P1847-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SU-02-032224

Quant Time: Mar 26 03:34:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032224S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 23 04:03:52 2024
Response via : Initial Calibration

