

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032524\
 Data File : VY017756.D
 Acq On : 25 Mar 2024 17:34
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 03/26/2024
 Supervised By :Semsettin Yesilyurt 03/26/2024

Quant Time: Mar 26 03:39:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032224S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 23 04:03:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	161673	50.000	ug/l	-0.02
34) 1,4-Difluorobenzene	8.697	114	256587	50.000	ug/l	-0.01
63) Chlorobenzene-d5	11.496	117	229791	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	111348	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	76759	54.498	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	= 109.000%	
35) Dibromofluoromethane	7.722	113	83029	54.134	ug/l	-0.02
Spiked Amount	50.000	Range	54 - 147	Recovery	= 108.260%	
50) Toluene-d8	10.185	98	314006	53.493	ug/l	-0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	= 106.980%	
62) 4-Bromofluorobenzene	12.489	95	100335	56.922	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 113.840%	

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.906	85	66301	47.502	ug/l	97
3) Chloromethane	2.113	50	100046	55.331	ug/l	97
4) Vinyl Chloride	2.253	62	122262	55.924	ug/l	99
5) Bromomethane	2.656	94	98052	57.441	ug/l	98
6) Chloroethane	2.796	64	83034	56.829	ug/l	99
7) Trichlorofluoromethane	3.131	101	127070	55.145	ug/l	99
8) Diethyl Ether	3.527	74	43828	57.691	ug/l	97
9) 1,1,2-Trichlorotrifluo...	3.905	101	80948	49.348	ug/l	98
10) Methyl Iodide	4.100	142	104833	52.281	ug/l	97
11) Tert butyl alcohol	4.954	59	27923	298.058	ug/l #	82
12) 1,1-Dichloroethene	3.881	96	79620	51.602	ug/l	99
13) Acrolein	3.729	56	35062	310.484	ug/l	100
14) Allyl chloride	4.491	41	105657	50.528	ug/l	98
15) Acrylonitrile	5.161	53	92459	293.792	ug/l	100
16) Acetone	3.948	43	70336	285.508	ug/l	94
17) Carbon Disulfide	4.198	76	258029	49.978	ug/l	99
18) Methyl Acetate	4.478	43	36967	59.599	ug/l	97
19) Methyl tert-butyl Ether	5.228	73	194890	58.581	ug/l	100
20) Methylene Chloride	4.722	84	90955	50.766	ug/l	93
21) trans-1,2-Dichloroethene	5.228	96	91456	51.993	ug/l	99
22) Diisopropyl ether	6.125	45	244584	53.720	ug/l	96
23) Vinyl Acetate	6.064	43	759891	284.164	ug/l	97
24) 1,1-Dichloroethane	6.021	63	155049	51.812	ug/l	99
25) 2-Butanone	6.990	43	111850	266.726	ug/l	94
26) 2,2-Dichloropropane	6.984	77	123280	48.464	ug/l	98
27) cis-1,2-Dichloroethene	6.990	96	106294	53.423	ug/l	98
28) Bromochloromethane	7.344	49	62950	54.260	ug/l	98
29) Tetrahydrofuran	7.350	42	74301	281.736	ug/l	95
30) Chloroform	7.508	83	164707	53.092	ug/l	100
31) Cyclohexane	7.789	56	126887	45.550	ug/l	99
32) 1,1,1-Trichloroethane	7.710	97	138551	51.281	ug/l	98
36) 1,1-Dichloropropene	7.923	75	116576	49.198	ug/l	99
37) Ethyl Acetate	7.076	43	51747	56.833	ug/l	99
38) Carbon Tetrachloride	7.911	117	126033	50.798	ug/l	100
39) Methylcyclohexane	9.191	83	149952	49.298	ug/l	99
40) Benzene	8.167	78	369003	52.025	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032524\
 Data File : VY017756.D
 Acq On : 25 Mar 2024 17:34
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 03/26/2024
 Supervised By :Semsettin Yesilyurt 03/26/2024

Quant Time: Mar 26 03:39:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032224S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 23 04:03:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.332	41	30580m	67.992	ug/l	
42) 1,2-Dichloroethane	8.240	62	92061	54.189	ug/l	99
43) Isopropyl Acetate	8.277	43	96006	53.250	ug/l	97
44) Trichloroethene	8.947	130	99427	53.592	ug/l	97
45) 1,2-Dichloropropane	9.222	63	85115	52.667	ug/l	99
46) Dibromomethane	9.313	93	49856	55.361	ug/l	97
47) Bromodichloromethane	9.502	83	126081	54.753	ug/l	100
48) Methyl methacrylate	9.295	41	43732	52.134	ug/l	97
49) 1,4-Dioxane	9.307	88	11493	1311.599	ug/l #	99
51) 4-Methyl-2-Pentanone	10.075	43	263818	272.004	ug/l	97
52) Toluene	10.252	92	232792	53.114	ug/l	99
53) t-1,3-Dichloropropene	10.471	75	111609	55.421	ug/l	96
54) cis-1,3-Dichloropropene	9.935	75	135223	54.967	ug/l	97
55) 1,1,2-Trichloroethane	10.648	97	65983	57.275	ug/l	96
56) Ethyl methacrylate	10.514	69	88819	54.708	ug/l	97
57) 1,3-Dichloropropane	10.794	76	113237	57.066	ug/l	100
58) 2-Chloroethyl Vinyl ether	9.788	63	211586	259.662	ug/l	100
59) 2-Hexanone	10.837	43	182551	267.385	ug/l	96
60) Dibromochloromethane	10.990	129	89656	58.182	ug/l	98
61) 1,2-Dibromoethane	11.093	107	62904	58.160	ug/l	100
64) Tetrachloroethene	10.727	164	116330	54.741	ug/l	96
65) Chlorobenzene	11.520	112	259410	52.443	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.593	131	89410	54.127	ug/l	98
67) Ethyl Benzene	11.599	91	443931	52.256	ug/l	98
68) m/p-Xylenes	11.709	106	351700	106.721	ug/l	98
69) o-Xylene	12.032	106	162527	54.018	ug/l	100
70) Styrene	12.050	104	279750	55.866	ug/l	99
71) Bromoform	12.215	173	52012	59.788	ug/l #	99
73) Isopropylbenzene	12.337	105	417780	49.689	ug/l	98
74) N-amyl acetate	12.148	43	85197	49.329	ug/l	97
75) 1,1,2,2-Tetrachloroethane	12.587	83	68612	52.999	ug/l	100
76) 1,2,3-Trichloropropane	12.636	75	52269m	50.834	ug/l	
77) Bromobenzene	12.617	156	102945	53.778	ug/l	95
78) n-propylbenzene	12.678	91	504093	49.528	ug/l	99
79) 2-Chlorotoluene	12.764	91	281131	50.369	ug/l	98
80) 1,3,5-Trimethylbenzene	12.818	105	339075	51.517	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.386	75	25147	56.512	ug/l	98
82) 4-Chlorotoluene	12.861	91	286862	50.092	ug/l	98
83) tert-Butylbenzene	13.081	119	306982	51.540	ug/l	98
84) 1,2,4-Trimethylbenzene	13.123	105	334632	51.861	ug/l	98
85) sec-Butylbenzene	13.257	105	450164	50.428	ug/l	99
86) p-Isopropyltoluene	13.373	119	385425	51.999	ug/l	99
87) 1,3-Dichlorobenzene	13.373	146	202053	52.088	ug/l	99
88) 1,4-Dichlorobenzene	13.452	146	195142	52.403	ug/l	98
89) n-Butylbenzene	13.702	91	340119	50.720	ug/l	99
90) Hexachloroethane	13.965	117	75759	50.121	ug/l	99
91) 1,2-Dichlorobenzene	13.745	146	173210	53.903	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.361	75	10028	58.703	ug/l	94
93) 1,2,4-Trichlorobenzene	15.013	180	91949	59.069	ug/l	98
94) Hexachlorobutadiene	15.117	225	53608	53.651	ug/l	98
95) Naphthalene	15.239	128	160383	55.421	ug/l	99
96) 1,2,3-Trichlorobenzene	15.428	180	78561	61.126	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032524\
Data File : VY017756.D
Acq On : 25 Mar 2024 17:34
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 03/26/2024
Supervised By :Semsettin Yesilyurt 03/26/2024

Quant Time: Mar 26 03:39:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032224S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 23 04:03:52 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032524\
 Data File : VY017756.D
 Acq On : 25 Mar 2024 17:34
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1


Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC

Quant Time: Mar 26 03:39:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032224S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 23 04:03:52 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 03/26/2024
 Supervised By :Semsettin Yesilyurt 03/26/2024









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.795 min Scan# 91
Delta R.T. -0.018 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion:168 Resp: 16167
Ion Ratio Lower Upper
168 100
99 48.5 39.1 58.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2024
Supervised By :Semsettin Yesilyurt 03/26/2024



#2
Dichlorodifluoromethane
Concen: 47.502 ug/l
RT: 1.906 min Scan# 14
Delta R.T. -0.012 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34

Tgt Ion: 85 Resp: 66301
Ion Ratio Lower Upper
85 100
87 33.7 16.0 48.0









#3
Chloromethane
Concen: 55.331 ug/l
RT: 2.113 min Scan# 41
Delta R.T. -0.012 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 50 Resp: 100040
Ion Ratio Lower Upper
50 100
52 33.7 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2024
Supervised By :Semsettin Yesilyurt 03/26/2024



#4
Vinyl Chloride
Concen: 55.924 ug/l
RT: 2.253 min Scan# 71
Delta R.T. -0.012 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34


Tgt Ion: 62 Resp: 122262
Ion Ratio Lower Upper
62 100
64 31.8 26.1 39.1





#5
Bromomethane
Concen: 57.441 ug/l
RT: 2.656 min Scan# 11
Delta R.T. -0.012 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 94 Resp: 98052
Ion Ratio Lower Upper
94 100
96 91.0 74.1 111.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2024
Supervised By :Semsettin Yesilyurt 03/26/2024



#6
Chloroethane
Concen: 56.829 ug/l
RT: 2.796 min Scan# 160
Delta R.T. -0.018 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34


Tgt Ion: 64 Resp: 83034
Ion Ratio Lower Upper
64 100
66 31.8 26.1 39.1





#7
Trichlorofluoromethane
Concen: 55.145 ug/l
RT: 3.131 min Scan# 218
Delta R.T. -0.018 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion:101 Resp: 127070
Ion Ratio Lower Upper
101 100
103 60.7 49.3 73.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2024
Supervised By :Semsettin Yesilyurt 03/26/2024



#8
Diethyl Ether
Concen: 57.691 ug/l
RT: 3.527 min Scan# 280
Delta R.T. -0.025 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34

Tgt Ion: 74 Resp: 43828
Ion Ratio Lower Upper
74 100
45 85.5 41.3 124.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 49.348 ug/l
RT: 3.905 min Scan# 346
Delta R.T. -0.025 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion:101 Resp: 80948
Ion Ratio Lower Upper
101 100
85 44.3 34.5 51.7
151 79.3 61.7 92.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2024
Supervised By :Semsettin Yesilyurt 03/26/2024



#10
Methyl Iodide
Concen: 52.281 ug/l
RT: 4.100 min Scan# 374
Delta R.T. -0.018 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34


Tgt Ion:142 Resp: 104833
Ion Ratio Lower Upper
142 100
127 43.0 32.8 49.2
141 14.6 11.4 17.2





#11
Tert butyl alcohol
Concen: 298.058 ug/l
RT: 4.954 min Scan# 518
Delta R.T. -0.025 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 59 Resp: 2792
Ion Ratio Lower Upper
59 100
57 9.7 3.0 4.6

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 03/26/2024
Supervised By :Semsettin Yesilyurt 03/26/2024




#12
1,1-Dichloroethene
Concen: 51.602 ug/l
RT: 3.881 min Scan# 338
Delta R.T. -0.018 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34




Tgt Ion: 96 Resp: 79620
Ion Ratio Lower Upper
96 100
61 141.7 114.6 171.8
98 63.4 50.8 76.2





#13
Acrolein
Concen: 310.484 ug/l
RT: 3.729 min Scan# 317
Delta R.T. -0.025 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 56 Resp: 35062
Ion Ratio Lower Upper
56 100
55 69.5 55.7 83.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2024
Supervised By :Semsettin Yesilyurt 03/26/2024



#14
Allyl chloride
Concen: 50.528 ug/l
RT: 4.491 min Scan# 438
Delta R.T. -0.018 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34


Tgt Ion: 41 Resp: 105657
Ion Ratio Lower Upper
41 100
39 71.6 58.3 87.5
76 46.9 39.0 58.6





#15
Acrylonitrile
Concen: 293.792 ug/l
RT: 5.161 min Scan# 548
Delta R.T. -0.031 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 53 Resp: 92459
Ion Ratio Lower Upper
53 100
52 80.7 64.7 97.1
51 35.8 28.4 42.6

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 03/26/2024
Supervised By :Semsettin Yesilyurt 03/26/2024




#16
Acetone
Concen: 285.508 ug/l
RT: 3.948 min Scan# 349
Delta R.T. -0.025 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34



Tgt Ion: 43 Resp: 70336
Ion Ratio Lower Upper
43 100
58 36.2 31.7 47.5





#17

Carbon Disulfide

Concen: 49.978 ug/l

RT: 4.198 min Scan# 394

Delta R.T. -0.025 min

Lab File: VY017756.D

Acq: 25 Mar 2024 17:34

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 76 Resp: 258029

Ion Ratio Lower Upper

76 100





78 8.9 6.8 10.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/26/2024

Supervised By :Semsettin Yesilyurt 03/26/2024



#18

Methyl Acetate

Concen: 59.599 ug/l

RT: 4.478 min Scan# 436

Delta R.T. -0.031 min

Lab File: VY017756.D




Acq: 25 Mar 2024 17:34


Tgt Ion: 43 Resp: 36967

Ion Ratio Lower Upper

43 100


74 27.8 23.3 34.9





#19
Methyl tert-butyl Ether
Concen: 58.581 ug/l
RT: 5.228 min Scan# 51
Delta R.T. -0.025 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 73 Resp: 194890
Ion Ratio Lower Upper
73 100
57 21.5 17.3 25.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2024
Supervised By :Semsettin Yesilyurt 03/26/2024



#20
Methylene Chloride
Concen: 50.766 ug/l
RT: 4.722 min Scan# 476
Delta R.T. -0.025 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34


Tgt Ion: 84 Resp: 90955
Ion Ratio Lower Upper
84 100
49 115.0 84.4 126.6
51 35.3 26.0 39.0
86 65.2 49.5 74.3





#21
trans-1,2-Dichloroethene
Concen: 51.993 ug/l
RT: 5.228 min Scan# 51
Delta R.T. -0.025 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 96 Resp: 91456
Ion Ratio Lower Upper
96 100
61 121.6 97.5 146.3
98 61.2 50.5 75.7

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 03/26/2024
Supervised By :Semsettin Yesilyurt 03/26/2024




#22
Diisopropyl ether
Concen: 53.720 ug/l
RT: 6.125 min Scan# 706
Delta R.T. -0.018 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34




Tgt Ion: 45 Resp: 244584
Ion Ratio Lower Upper
45 100
43 56.8 42.2 63.2
87 32.4 25.0 37.6
59 13.1 10.5 15.7





#23
 Vinyl Acetate
 Concen: 284.164 ug/l
 RT: 6.064 min Scan# 61
 Delta R.T. -0.025 min
 Lab File: VY017756.D
 Acq: 25 Mar 2024 17:34




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC



Tgt Ion: 43 Resp: 75989
 Ion Ratio Lower Upper
 43 100
 86 12.3 10.6 16.0




Manual Integrations
 APPROVED


Reviewed By :Mahesh Dadoda 03/26/2024
 Supervised By :Semsettin Yesilyurt 03/26/2024



#24
 1,1-Dichloroethane
 Concen: 51.812 ug/l
 RT: 6.021 min Scan# 689
 Delta R.T. -0.025 min
 Lab File: VY017756.D
 Acq: 25 Mar 2024 17:34


Tgt Ion: 63 Resp: 155049
 Ion Ratio Lower Upper
 63 100
 98 7.8 4.0 12.2
 100 4.8 2.6 8.0





#25
2-Butanone
Concen: 266.726 ug/l
RT: 6.990 min Scan# 847
Delta R.T. -0.019 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 43 Resp: 111850
Ion Ratio Lower Upper
43 100
72 30.8 27.3 40.9

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 03/26/2024
Supervised By :Semsettin Yesilyurt 03/26/2024




#26
2,2-Dichloropropane
Concen: 48.464 ug/l
RT: 6.984 min Scan# 847
Delta R.T. -0.025 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34




Tgt Ion: 77 Resp: 123280
Ion Ratio Lower Upper
77 100
97 24.9 12.0 36.1





#27
cis-1,2-Dichloroethene
Concen: 53.423 ug/l
RT: 6.990 min Scan# 84
Delta R.T. -0.018 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 96 Resp: 106294
Ion Ratio Lower Upper
96 100
61 131.5 0.0 256.2
98 65.1 0.0 130.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2024
Supervised By :Semsettin Yesilyurt 03/26/2024



#28
Bromochloromethane
Concen: 54.260 ug/l
RT: 7.344 min Scan# 906
Delta R.T. -0.012 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34

Tgt Ion: 49 Resp: 62950
Ion Ratio Lower Upper
49 100
129 2.5 0.0 4.6
130 89.2 72.7 109.1









#29
Tetrahydrofuran
Concen: 281.736 ug/l
RT: 7.350 min Scan# 907
Delta R.T. -0.018 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 42 Resp: 7430
Ion Ratio Lower Upper
42 100
72 50.1 43.0 64.4
71 48.0 40.8 61.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2024
Supervised By :Semsettin Yesilyurt 03/26/2024



#30
Chloroform
Concen: 53.092 ug/l
RT: 7.508 min Scan# 933
Delta R.T. -0.025 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34


Tgt Ion: 83 Resp: 164707
Ion Ratio Lower Upper
83 100
85 65.1 52.2 78.4





#31
Cyclohexane
Concen: 45.550 ug/l
RT: 7.789 min Scan# 91
Delta R.T. -0.018 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 56 Resp: 12688
Ion Ratio Lower Upper
56 100
69 33.4 28.7 43.1
84 97.3 77.7 116.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2024
Supervised By :Semsettin Yesilyurt 03/26/2024



#32
1,1,1-Trichloroethane
Concen: 51.281 ug/l
RT: 7.710 min Scan# 966
Delta R.T. -0.012 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34

Tgt Ion: 97 Resp: 138551
Ion Ratio Lower Upper
97 100
99 64.0 50.3 75.5
61 42.0 32.6 48.8





#33

1,2-Dichloroethane-d4

Concen: 54.498 ug/l

RT: 8.149 min Scan# 1108

Delta R.T. -0.012 min

Lab File: VY017756.D

Acq: 25 Mar 2024 17:34

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 65 Resp: 76758

Ion Ratio Lower Upper

65 100





67 54.4 0.0 109.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/26/2024

Supervised By :Semsettin Yesilyurt 03/26/2024



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.697 min Scan# 1128

Delta R.T. -0.012 min

Lab File: VY017756.D

Acq: 25 Mar 2024 17:34




Tgt Ion:114 Resp: 256587


Ion Ratio Lower Upper

114 100

63 16.5 0.0 35.0


88 13.1 0.0 27.2





#35
Dibromofluoromethane
Concen: 54.134 ug/l
RT: 7.722 min Scan# 90
Delta R.T. -0.018 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion:113 Resp: 83029
Ion Ratio Lower Upper
113 100
111 103.5 82.2 123.4
192 20.3 15.9 23.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2024
Supervised By :Semsettin Yesilyurt 03/26/2024



#36
1,1-Dichloropropene
Concen: 49.198 ug/l
RT: 7.923 min Scan# 1001
Delta R.T. -0.018 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34

Tgt Ion: 75 Resp: 116576
Ion Ratio Lower Upper
75 100
110 36.4 19.0 57.0
77 31.5 25.3 37.9









#37
Ethyl Acetate
Concen: 56.833 ug/l
RT: 7.076 min Scan# 80
Delta R.T. -0.025 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 43 Resp: 5174
Ion Ratio Lower Upper
43 100
61 16.2 12.6 19.0
70 10.6 7.8 11.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2024
Supervised By :Semsettin Yesilyurt 03/26/2024



#38
Carbon Tetrachloride
Concen: 50.798 ug/l
RT: 7.911 min Scan# 999
Delta R.T. -0.012 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34


Tgt Ion:117 Resp: 126033
Ion Ratio Lower Upper
117 100
119 95.1 75.8 113.6
121 30.2 24.0 36.0





#39
Methylcyclohexane
Concen: 49.298 ug/l
RT: 9.191 min Scan# 1041
Delta R.T. -0.006 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 83 Resp: 149952
Ion Ratio Lower Upper
83 100
55 65.4 51.5 77.3
98 49.7 40.5 60.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2024
Supervised By :Semsettin Yesilyurt 03/26/2024



#40
Benzene
Concen: 52.025 ug/l
RT: 8.167 min Scan# 1041
Delta R.T. -0.012 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34


Tgt Ion: 78 Resp: 369003
Ion Ratio Lower Upper
78 100
77 24.4 19.5 29.3





#41
Methacrylonitrile
Concen: 67.992 ug/l m
RT: 7.332 min Scan# 904
Delta R.T. -0.000 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 41 Resp: 30580
Ion Ratio Lower Upper
41 100
39 0.0 48.0 72.0
67 0.0 75.1 112.7
52 0.0 29.8 44.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2024
Supervised By :Semsettin Yesilyurt 03/26/2024



#42
1,2-Dichloroethane
Concen: 54.189 ug/l
RT: 8.240 min Scan# 1053
Delta R.T. -0.012 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34

Tgt Ion: 62 Resp: 92061
Ion Ratio Lower Upper
62 100
98 10.5 0.0 22.0





#43

Isopropyl Acetate

Concen: 53.250 ug/l

RT: 8.277 min Scan# 1105

Delta R.T. -0.012 min

Lab File: VY017756.D

Acq: 25 Mar 2024 17:34

Instrument :

MSVOA_Y

ClientSampleId :

VSTDC050EC





Tgt Ion: 43 Resp: 96000

Ion	Ratio	Lower	Upper
43	100		
61	32.8	27.9	41.9
87	15.3	13.1	19.7

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/26/2024
Supervised By :Semsettin Yesilyurt 03/26/2024



#44

Trichloroethene

Concen: 53.592 ug/l

RT: 8.947 min Scan# 1169




Delta R.T. -0.012 min


Lab File: VY017756.D

Acq: 25 Mar 2024 17:34

Tgt Ion:130 Resp: 99427

Ion	Ratio	Lower	Upper
130	100		
95	91.4	0.0	188.0





#45

1,2-Dichloropropane

Concen: 52.667 ug/l

RT: 9.222 min Scan# 11

Delta R.T. -0.012 min

Lab File: VY017756.D

Acq: 25 Mar 2024 17:34

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 63 Resp: 8511

Ion Ratio Lower Upper





63 100

65 32.4 25.4 38.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/26/2024
Supervised By :Semsettin Yesilyurt 03/26/2024



#46

Dibromomethane

Concen: 55.361 ug/l

RT: 9.313 min Scan# 1229

Delta R.T. -0.006 min

Lab File: VY017756.D

Acq: 25 Mar 2024 17:34




Tgt Ion: 93 Resp: 49856


Ion Ratio Lower Upper

93 100

95 84.2 66.2 99.4


174 105.4 80.9 121.3





#47
Bromodichloromethane
Concen: 54.753 ug/l
RT: 9.502 min Scan# 1260
Delta R.T. -0.006 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 83 Resp: 12608
Ion Ratio Lower Upper
83 100
85 62.7 50.1 75.1
127 8.7 7.4 11.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2024
Supervised By :Semsettin Yesilyurt 03/26/2024



#48
Methyl methacrylate
Concen: 52.134 ug/l
RT: 9.295 min Scan# 1226
Delta R.T. -0.012 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34

Tgt Ion: 41 Resp: 43732
Ion Ratio Lower Upper
41 100
69 102.9 84.1 126.1
39 57.5 43.6 65.4





#49

1,4-Dioxane

Concen: 1311.599 ug/l

RT: 9.307 min Scan# 1149

Delta R.T. -0.006 min

Lab File: VY017756.D

Acq: 25 Mar 2024 17:34

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 88 Resp: 1149

Ion Ratio Lower Upper

88 100

43 29.4 0.0 0.0





58 61.0 49.3 73.9

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/26/2024

Supervised By :Semsettin Yesilyurt 03/26/2024



#50

Toluene-d8

Concen: 53.493 ug/l

RT: 10.185 min Scan# 1372

Delta R.T. -0.012 min

Lab File: VY017756.D




Acq: 25 Mar 2024 17:34


Tgt Ion: 98 Resp: 314006

Ion Ratio Lower Upper

98 100

100 66.1 52.0 78.0









#51
4-Methyl-2-Pentanone
Concen: 272.004 ug/l
RT: 10.075 min Scan# 1354
Delta R.T. -0.006 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 43 Resp: 263818
Ion Ratio Lower Upper
43 100
58 42.7 35.7 53.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2024
Supervised By :Semsettin Yesilyurt 03/26/2024



#52
Toluene
Concen: 53.114 ug/l
RT: 10.252 min Scan# 1383
Delta R.T. -0.006 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34

Tgt Ion: 92 Resp: 232792
Ion Ratio Lower Upper
92 100
91 172.7 139.0 208.6





#53
t-1,3-Dichloropropene
Concen: 55.421 ug/l
RT: 10.471 min Scan# 1419
Delta R.T. -0.006 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34

Instrument :

MSVOA_Y

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 75 Resp: 111609
Ion Ratio Lower Upper
75 100
77 32.0 24.1 36.1

Manual Integrations




APPROVED


Reviewed By :Mahesh Dadoda 03/26/2024
Supervised By :Semsettin Yesilyurt 03/26/2024



#54
cis-1,3-Dichloropropene
Concen: 54.967 ug/l
RT: 9.935 min Scan# 1331
Delta R.T. -0.006 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34


Tgt Ion: 75 Resp: 135223
Ion Ratio Lower Upper
75 100
77 31.8 25.4 38.0
39 40.3 29.8 44.6





#55
1,1,2-Trichloroethane
Concen: 57.275 ug/l
RT: 10.648 min Scan# 1448
Delta R.T. -0.006 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 97 Resp: 6598
Ion Ratio Lower Upper
97 100
83 87.1 67.5 101.3
85 55.3 42.9 64.3
99 63.9 46.6 70.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 03/26/2024
Supervised By :Semsettin Yesilyurt 03/26/2024




#56
Ethyl methacrylate
Concen: 54.708 ug/l
RT: 10.514 min Scan# 1426
Delta R.T. -0.006 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34



Tgt Ion: 69 Resp: 88819
Ion Ratio Lower Upper
69 100
41 55.2 42.3 63.5
39 30.1 22.9 34.3





#57

1,3-Dichloropropane

Concen: 57.066 ug/l

RT: 10.794 min Scan# 1473

Delta R.T. -0.006 min

Lab File: VY017756.D

Acq: 25 Mar 2024 17:34

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 76 Resp: 11323

Ion Ratio Lower Upper

76 100





78 32.6 26.2 39.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/26/2024

Supervised By :Semsettin Yesilyurt 03/26/2024



#58

2-Chloroethyl Vinyl ether

Concen: 259.662 ug/l

RT: 9.788 min Scan# 1307

Delta R.T. -0.007 min

Lab File: VY017756.D




Acq: 25 Mar 2024 17:34


Tgt Ion: 63 Resp: 211586

Ion Ratio Lower Upper

63 100

106 29.9 23.7 35.5









#59
2-Hexanone
Concen: 267.385 ug/l
RT: 10.837 min Scan# 1480
Delta R.T. -0.006 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 43 Resp: 18255
Ion Ratio Lower Upper
43 100
58 59.6 31.4 94.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2024
Supervised By :Semsettin Yesilyurt 03/26/2024



#60
Dibromochloromethane
Concen: 58.182 ug/l
RT: 10.990 min Scan# 1504
Delta R.T. -0.006 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34


Tgt Ion:129 Resp: 89656
Ion Ratio Lower Upper
129 100
127 76.7 39.0 117.0





#61
1,2-Dibromoethane
Concen: 58.160 ug/l
RT: 11.093 min Scan# 1109
Delta R.T. -0.012 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion:107 Resp: 62904
Ion Ratio Lower Upper
107 100
109 94.4 75.9 113.9

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 03/26/2024
Supervised By :Semsettin Yesilyurt 03/26/2024




#62
4-Bromofluorobenzene
Concen: 56.922 ug/l
RT: 12.489 min Scan# 1750
Delta R.T. -0.006 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34



Tgt Ion: 95 Resp: 100335
Ion Ratio Lower Upper
95 100
174 91.4 0.0 175.6
176 87.2 0.0 171.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.496 min Scan# 11

Delta R.T. -0.006 min

Lab File: VY017756.D

Acq: 25 Mar 2024 17:34

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion:117 Resp: 22979

Ion Ratio Lower Upper

117 100

82 51.0 42.4 63.6





119 31.4 25.9 38.9

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/26/2024

Supervised By :Semsettin Yesilyurt 03/26/2024



#64

Tetrachloroethene

Concen: 54.741 ug/l

RT: 10.727 min Scan# 1461

Delta R.T. -0.006 min

Lab File: VY017756.D

Acq: 25 Mar 2024 17:34

Tgt Ion:164 Resp: 116330




Ion Ratio Lower Upper


164 100

166 122.3 102.6 154.0

129 88.8 72.0 108.0


131 84.4 70.0 105.0





#65
Chlorobenzene
Concen: 52.443 ug/l
RT: 11.520 min Scan# 11
Delta R.T. -0.006 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion:112 Resp: 259410
Ion Ratio Lower Upper
112 100
114 32.1 26.4 39.6

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 03/26/2024
Supervised By :Semsettin Yesilyurt 03/26/2024




#66
1,1,1,2-Tetrachloroethane
Concen: 54.127 ug/l
RT: 11.593 min Scan# 1603
Delta R.T. -0.012 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34



Tgt Ion:131 Resp: 89410
Ion Ratio Lower Upper
131 100
133 96.1 47.8 143.3
119 65.8 34.1 102.2









#67
Ethyl Benzene
Concen: 52.256 ug/l
RT: 11.599 min Scan# 1604
Delta R.T. -0.006 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 91 Resp: 44393
Ion Ratio Lower Upper
91 100
106 31.8 24.5 36.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2024
Supervised By :Semsettin Yesilyurt 03/26/2024



#68
m/p-Xylenes
Concen: 106.721 ug/l
RT: 11.709 min Scan# 1622
Delta R.T. -0.006 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34

Tgt Ion:106 Resp: 351700
Ion Ratio Lower Upper
106 100
91 191.8 156.0 234.0









#69
o-Xylene
Concen: 54.018 ug/l
RT: 12.032 min Scan# 1675
Delta R.T. -0.012 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion:106 Resp: 16252
Ion Ratio Lower Upper
106 100
91 206.8 103.6 310.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2024
Supervised By :Semsettin Yesilyurt 03/26/2024



#70
Styrene
Concen: 55.866 ug/l
RT: 12.050 min Scan# 1678
Delta R.T. -0.006 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34

Tgt Ion:104 Resp: 279750
Ion Ratio Lower Upper
104 100
78 46.0 37.0 55.4
103 54.5 44.2 66.2





#71

Bromoform

Concen: 59.788 ug/l

RT: 12.215 min Scan# 1706

Delta R.T. -0.006 min

Lab File: VY017756.D

Acq: 25 Mar 2024 17:34

Instrument :

MSVOA_Y

ClientSampleId :





VSTDCCC050EC

Tgt Ion:173	Resp:	5201
Ion Ratio	Lower	Upper
173	100	
175	49.6	25.1 75.3
254	0.0	0.0 0.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/26/2024
Supervised By :Semsettin Yesilyurt 03/26/2024



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 13.434 min Scan# 1905


Delta R.T. -0.000 min

Lab File: VY017756.D

Acq: 25 Mar 2024 17:34

Tgt Ion:152	Resp:	111348
Ion Ratio	Lower	Upper
152	100	
115	55.5	28.2 84.7
150	172.8	0.0 345.6





#73

Isopropylbenzene

Concen: 49.689 ug/l

RT: 12.337 min Scan# 11

Delta R.T. -0.006 min

Lab File: VY017756.D

Acq: 25 Mar 2024 17:34

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion:105 Resp: 417780

Ion Ratio Lower Upper

105 100





120 27.5 13.4 40.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/26/2024

Supervised By :Semsettin Yesilyurt 03/26/2024



#74

N-ethyl acetate

Concen: 49.329 ug/l

RT: 12.148 min Scan# 1694

Delta R.T. -0.006 min

Lab File: VY017756.D

Acq: 25 Mar 2024 17:34

Tgt Ion: 43 Resp: 85197




Ion Ratio Lower Upper


43 100

70 49.2 41.4 62.2

55 27.9 23.3 34.9

61 26.4 22.6 33.8





#75

1,1,2,2-Tetrachloroethane

Concen: 52.999 ug/l

RT: 12.587 min Scan# 1766

Delta R.T. -0.006 min

Lab File: VY017756.D

Acq: 25 Mar 2024 17:34

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 83 Resp: 6861

Ion Ratio Lower Upper

83 100

131 10.7 5.0 14.9





85 64.0 32.0 96.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/26/2024

Supervised By :Semsettin Yesilyurt 03/26/2024



#76

1,2,3-Trichloropropane

Concen: 50.834 ug/l m

RT: 12.636 min Scan# 1774

Delta R.T. -0.006 min

Lab File: VY017756.D




Acq: 25 Mar 2024 17:34


Tgt Ion: 75 Resp: 52269

Ion Ratio Lower Upper

75 100

77 0.0 0.0 0.0





#77

Bromobenzene

Concen: 53.778 ug/l

RT: 12.617 min Scan# 1771

Delta R.T. -0.000 min

Lab File: VY017756.D

Acq: 25 Mar 2024 17:34

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC





Tgt Ion:156 Resp: 102945

Ion	Ratio	Lower	Upper
156	100		
77	165.5	87.1	261.3
158	95.8	48.6	145.9

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/26/2024
Supervised By :Semsettin Yesilyurt 03/26/2024



#78

n-propylbenzene

Concen: 49.528 ug/l

RT: 12.678 min Scan# 1781




Delta R.T. -0.006 min


Lab File: VY017756.D

Acq: 25 Mar 2024 17:34

Tgt Ion: 91 Resp: 504093


Ion	Ratio	Lower	Upper
91	100		
120	23.4	11.5	34.4





#79
2-Chlorotoluene
Concen: 50.369 ug/l
RT: 12.764 min Scan# 1796
Delta R.T. -0.006 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 91 Resp: 28113
Ion Ratio Lower Upper
91 100
126 36.8 17.8 53.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2024
Supervised By :Semsettin Yesilyurt 03/26/2024



#80
1,3,5-Trimethylbenzene
Concen: 51.517 ug/l
RT: 12.818 min Scan# 1804
Delta R.T. -0.006 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34

Tgt Ion:105 Resp: 339075
Ion Ratio Lower Upper
105 100
120 51.2 25.1 75.2









#81
trans-1,4-Dichloro-2-butene
Concen: 56.512 ug/l
RT: 12.386 min Scan# 1733
Delta R.T. -0.006 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 75 Resp: 2514
Ion Ratio Lower Upper
75 100
53 84.2 66.8 100.2
89 46.8 35.8 53.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2024
Supervised By :Semsettin Yesilyurt 03/26/2024



#82
4-Chlorotoluene
Concen: 50.092 ug/l
RT: 12.861 min Scan# 1811
Delta R.T. -0.006 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34

Tgt Ion: 91 Resp: 286862
Ion Ratio Lower Upper
91 100
126 36.4 17.6 52.9





#83

tert-Butylbenzene

Concen: 51.540 ug/l

RT: 13.081 min Scan# 1848

Delta R.T. -0.006 min

Lab File: VY017756.D

Acq: 25 Mar 2024 17:34

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion:119 Resp: 30698

Ion Ratio Lower Upper

119 100

91 60.6 31.3 93.8





134 26.6 13.4 40.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/26/2024

Supervised By :Semsettin Yesilyurt 03/26/2024



#84

1,2,4-Trimethylbenzene

Concen: 51.861 ug/l

RT: 13.123 min Scan# 1854

Delta R.T. -0.006 min

Lab File: VY017756.D




Acq: 25 Mar 2024 17:34


Tgt Ion:105 Resp: 334632

Ion Ratio Lower Upper

105 100

120 47.8 23.2 69.6





#85

sec-Butylbenzene

Concen: 50.428 ug/l

RT: 13.257 min Scan# 1876

Delta R.T. -0.006 min

Lab File: VY017756.D

Acq: 25 Mar 2024 17:34

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion:105 Resp: 450164

Ion Ratio Lower Upper

105 100





134 20.6 10.1 30.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/26/2024

Supervised By :Semsettin Yesilyurt 03/26/2024



#86

p-Isopropyltoluene

Concen: 51.999 ug/l

RT: 13.373 min Scan# 1895

Delta R.T. -0.006 min

Lab File: VY017756.D

Acq: 25 Mar 2024 17:34




Tgt Ion:119 Resp: 385425


Ion Ratio Lower Upper

119 100

134 26.8 13.1 39.3

91 21.7 10.7 32.1





#87

1,3-Dichlorobenzene

Concen: 52.088 ug/l

RT: 13.373 min Scan# 1896

Delta R.T. -0.006 min

Lab File: VY017756.D

Acq: 25 Mar 2024 17:34

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion:146 Resp: 20205

Ion Ratio Lower Upper

146 100

111 37.5 19.0 57.0





148 62.8 31.7 95.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/26/2024

Supervised By :Semsettin Yesilyurt 03/26/2024



#88

1,4-Dichlorobenzene

Concen: 52.403 ug/l

RT: 13.452 min Scan# 1908

Delta R.T. -0.006 min

Lab File: VY017756.D

Acq: 25 Mar 2024 17:34




Tgt Ion:146 Resp: 195142


Ion Ratio Lower Upper

146 100

111 35.9 18.4 55.4

148 62.6 32.4 97.2





#89

n-Butylbenzene

Concen: 50.720 ug/l

RT: 13.702 min Scan# 1950

Delta R.T. -0.006 min

Lab File: VY017756.D

Acq: 25 Mar 2024 17:34

Instrument :

MSVOA_Y

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 91	Resp: 340119
Ion Ratio	Lower Upper
91	100
92	52.3 26.6 79.7
134	26.6 12.9 38.7

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/26/2024
Supervised By :Semsettin Yesilyurt 03/26/2024



#90

Hexachloroethane

Concen: 50.121 ug/l




RT: 13.965 min Scan# 1992


Delta R.T. -0.006 min

Lab File: VY017756.D

Acq: 25 Mar 2024 17:34

Tgt Ion:117	Resp: 75759
Ion Ratio	Lower Upper
117	100
201	75.4 37.4 112.2









#91
1,2-Dichlorobenzene
Concen: 53.903 ug/l
RT: 13.745 min Scan# 1956
Delta R.T. -0.006 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion:146 Resp: 173210
Ion Ratio Lower Upper
146 100
111 38.1 19.9 59.7
148 63.5 31.8 95.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 03/26/2024
Supervised By :Semsettin Yesilyurt 03/26/2024



#92
1,2-Dibromo-3-Chloropropane
Concen: 58.703 ug/l
RT: 14.361 min Scan# 2057
Delta R.T. -0.006 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34

Tgt Ion: 75 Resp: 10028
Ion Ratio Lower Upper
75 100
155 94.7 48.6 145.8
157 114.9 62.7 188.2





#93

1,2,4-Trichlorobenzene

Concen: 59.069 ug/l

RT: 15.013 min Scan# 2164

Delta R.T. 0.000 min

Lab File: VY017756.D


Acq: 25 Mar 2024 17:34

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC






Tgt Ion:180	Resp:	91949
Ion Ratio	Lower	Upper
180	100	
182	93.3	47.9 143.8
145	28.9	14.2 42.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/26/2024
Supervised By :Semsettin Yesilyurt 03/26/2024



#94

Hexachlorobutadiene

Concen: 53.651 ug/l




RT: 15.117 min Scan# 2181


Delta R.T. -0.006 min

Lab File: VY017756.D

Acq: 25 Mar 2024 17:34


Tgt Ion:225	Resp:	53608
Ion Ratio	Lower	Upper
225	100	
223	64.4	30.4 91.2
227	64.0	31.9 95.9





#95
Naphthalene
Concen: 55.421 ug/l
RT: 15.239 min Scan# 21
Delta R.T. -0.006 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion:128 Resp: 16038
Ion Ratio Lower Upper
128 100
127 12.7 10.6 16.0
129 11.1 8.6 13.0

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 03/26/2024
Supervised By :Semsettin Yesilyurt 03/26/2024



#96
1,2,3-Trichlorobenzene
Concen: 61.126 ug/l
RT: 15.428 min Scan# 2232
Delta R.T. -0.006 min
Lab File: VY017756.D
Acq: 25 Mar 2024 17:34

Tgt Ion:180 Resp: 78561
Ion Ratio Lower Upper
180 100
182 93.4 47.6 142.9
145 30.3 15.4 46.1

