

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032624\
 Data File : VY017765.D
 Acq On : 26 Mar 2024 13:13
 Operator : SY/MD
 Sample : P1865-01
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VNJ-235

Quant Time: Mar 27 01:29:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032224S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 23 04:03:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	35270	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.697	114	47767	50.000	ug/l	-0.01
63) Chlorobenzene-d5	11.502	117	32310	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	8488	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	12137	39.500	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	79.000%
35) Dibromofluoromethane	7.728	113	13803	48.342	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.680%
50) Toluene-d8	10.191	98	50964	46.637	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	93.280%
62) 4-Bromofluorobenzene	12.489	95	9071	27.643	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	55.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032624\
Data File : VY017765.D
Acq On : 26 Mar 2024 13:13
Operator : SY/MD
Sample : P1865-01
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VNJ-235

Quant Time: Mar 27 01:29:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032224S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 23 04:03:52 2024
Response via : Initial Calibration

