

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032624\
 Data File : VY017772.D
 Acq On : 26 Mar 2024 16:10
 Operator : SY/MD
 Sample : P1869-01
 Misc : 4.37g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SOIL-1

Quant Time: Mar 27 01:31:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032224S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 23 04:03:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	127149	50.000	ug/l	-0.02
34) 1,4-Difluorobenzene	8.697	114	230544	50.000	ug/l	-0.01
63) Chlorobenzene-d5	11.496	117	207703	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	79740	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	69247	62.514	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	125.020%
35) Dibromofluoromethane	7.728	113	76590	55.577	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.160%
50) Toluene-d8	10.191	98	265491	50.337	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.680%
62) 4-Bromofluorobenzene	12.489	95	73419	46.357	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	92.720%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032624\
Data File : VY017772.D
Acq On : 26 Mar 2024 16:10
Operator : SY/MD
Sample : P1869-01
Misc : 4.37g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SOIL-1

Quant Time: Mar 27 01:31:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032224S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 23 04:03:52 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.795 min Scan# 91
Delta R.T. -0.018 min
Lab File: VY017772.D
Acq: 26 Mar 2024 16:10




Instrument :
MSVOA_Y
ClientSampleId :
SOIL-1


Tgt Ion:168 Resp: 127149
Ion Ratio Lower Upper
168 100
99 48.4 39.1 58.7



#33
1,2-Dichloroethane-d4
Concen: 62.514 ug/l
RT: 8.149 min Scan# 1038
Delta R.T. -0.012 min
Lab File: VY017772.D
Acq: 26 Mar 2024 16:10

Tgt Ion: 65 Resp: 69247
Ion Ratio Lower Upper
65 100
67 55.1 0.0 109.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.697 min Scan# 1130

Delta R.T. -0.012 min

Lab File: VY017772.D

Acq: 26 Mar 2024 16:10

Instrument :

MSVOA_Y

ClientSampleId :

SOIL-1





Tgt Ion:114 Resp: 230544

Ion Ratio Lower Upper

114 100

63 17.9 0.0 35.0

88 14.0 0.0 27.2



#35

Dibromofluoromethane

Concen: 55.577 ug/l

RT: 7.728 min Scan# 969

Delta R.T. -0.012 min

Lab File: VY017772.D

Acq: 26 Mar 2024 16:10




Tgt Ion:113 Resp: 76590


Ion Ratio Lower Upper

113 100

111 102.9 82.2 123.4

192 20.0 15.9 23.9









#50
Toluene-d8
Concen: 50.337 ug/l
RT: 10.191 min Scan# 11
Delta R.T. -0.006 min
Lab File: VY017772.D
Acq: 26 Mar 2024 16:10




Instrument :
MSVOA_Y
ClientSampleId :
SOIL-1


Tgt Ion: 98 Resp: 265491
Ion Ratio Lower Upper
98 100
100 65.3 52.0 78.0



#62
4-Bromofluorobenzene
Concen: 46.357 ug/l
RT: 12.489 min Scan# 1750
Delta R.T. -0.006 min
Lab File: VY017772.D
Acq: 26 Mar 2024 16:10

Tgt Ion: 95 Resp: 73419
Ion Ratio Lower Upper
95 100
174 93.0 0.0 175.6
176 90.7 0.0 171.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.496 min Scan# 11
Delta R.T. -0.006 min
Lab File: VY017772.D
Acq: 26 Mar 2024 16:10

Instrument :
MSVOA_Y
ClientSampleId :
SOIL-1

Tgt Ion:117 Resp: 207703
Ion Ratio Lower Upper
117 100
82 51.7 42.4 63.6
119 33.0 25.9 38.9



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.434 min Scan# 1905
Delta R.T. -0.000 min
Lab File: VY017772.D
Acq: 26 Mar 2024 16:10

Tgt Ion:152 Resp: 79740
Ion Ratio Lower Upper
152 100
115 54.7 28.2 84.7
150 155.8 0.0 345.6

