

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032723\
 Data File : VY013092.D
 Acq On : 27 Mar 2023 15:20
 Operator : KP/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Mar 28 01:44:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	84082	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	144743	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	135399	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	62023	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	56824	68.067	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 136.140%	
35) Dibromofluoromethane	7.716	113	50027	57.270	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 114.540%	
50) Toluene-d8	10.179	98	174087	54.514	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 109.020%	
62) 4-Bromofluorobenzene	12.483	95	54538	50.598	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 101.200%	
Target Compounds						
					Qvalue	
16) Acetone	3.942	43	6447	19.675	ug/l	94
52) Toluene	10.240	92	3008	1.205	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032723\
Data File : VY013092.D
Acq On : 27 Mar 2023 15:20
Operator : KP/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Mar 28 01:44:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration

