

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032723\
 Data File : VY013098.D
 Acq On : 27 Mar 2023 17:41
 Operator : KP/MD
 Sample : 02111-01
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 APCS-SOUTH

Quant Time: Mar 28 01:46:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	32063	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	53267	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	52075	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	24876	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	22486	70.634	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	141.260%
35) Dibromofluoromethane	7.716	113	20198	62.831	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	125.660%
50) Toluene-d8	10.179	98	65180	55.462	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	110.920%
62) 4-Bromofluorobenzene	12.483	95	20978	52.886	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	105.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032723\
Data File : VY013098.D
Acq On : 27 Mar 2023 17:41
Operator : KP/MD
Sample : 02111-01
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
APCS-SOUTH

Quant Time: Mar 28 01:46:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration

