

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032723\
 Data File : VY013099.D
 Acq On : 27 Mar 2023 18:04
 Operator : KP/MD
 Sample : 02111-02
 Misc : 4.99g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 APCS-EAST

Quant Time: Mar 28 01:47:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	508	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.691	114	891	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	705	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	48	50.000	ug/l	# 0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	510	101.115	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	202.220%#
35) Dibromofluoromethane	7.704	113	269	50.026	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.060%
50) Toluene-d8	10.179	98	1032	52.498	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	105.000%
62) 4-Bromofluorobenzene	12.489	95	203	30.595	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	61.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032723\
Data File : VY013099.D
Acq On : 27 Mar 2023 18:04
Operator : KP/MD
Sample : 02111-02
Misc : 4.99g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
APCS-EAST

Quant Time: Mar 28 01:47:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration

