

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032723\
 Data File : VY013102.D
 Acq On : 27 Mar 2023 19:15
 Operator : KP/MD
 Sample : 02111-05
 Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 APCS-WEST

Quant Time: Mar 28 01:48:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	12364	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	19403	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	18748	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	7133	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	9583	78.064	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	156.120%
35) Dibromofluoromethane	7.715	113	8288	70.779	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	141.560%
50) Toluene-d8	10.178	98	23770	55.527	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	111.060%
62) 4-Bromofluorobenzene	12.483	95	6685	46.267	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	92.540%
Target Compounds						
3) Chloromethane	2.107	50	252	1.730	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032723\
Data File : VY013102.D
Acq On : 27 Mar 2023 19:15
Operator : KP/MD
Sample : 02111-05
Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
APCS-WEST

Quant Time: Mar 28 01:48:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration

