



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032723\
 Data File : VY013084.D
 Acq On : 27 Mar 2023 10:10
 Operator : KP/MD
 Sample : BFB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Title : SW846 8260
 Last Update : Fri Mar 03 01:13:13 2023



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.9	9511	PASS
75	95	30	60	51.8	26032	PASS
95	95	100	100	100.0	50240	PASS
96	95	5	9	6.4	3234	PASS
173	174	0.00	2	1.4	651	PASS
174	95	50	100	89.7	45053	PASS
175	174	5	9	7.9	3560	PASS
176	174	95	101	97.2	43784	PASS
177	176	5	9	6.4	2792	PASS