

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032724\
 Data File : VY017784.D
 Acq On : 27 Mar 2024 16:22
 Operator : SY/MD
 Sample : P1894-04
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RB40858RT

Quant Time: Mar 28 02:18:43 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032224S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 23 04:03:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	78584	50.000	ug/l	-0.02
34) 1,4-Difluorobenzene	8.691	114	142405	50.000	ug/l	-0.02
63) Chlorobenzene-d5	11.496	117	124669	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	43398	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	42057	61.432	ug/l	-0.02
Spiked Amount	50.000	Range	50 - 163	Recovery	=	122.860%
35) Dibromofluoromethane	7.722	113	47027	55.246	ug/l	-0.02
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.500%
50) Toluene-d8	10.185	98	170193	52.240	ug/l	-0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	=	104.480%
62) 4-Bromofluorobenzene	12.489	95	42545	43.490	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	86.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032724\
Data File : VY017784.D
Acq On : 27 Mar 2024 16:22
Operator : SY/MD
Sample : P1894-04
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RB40858RT

Quant Time: Mar 28 02:18:43 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032224S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 23 04:03:52 2024
Response via : Initial Calibration

