

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY032822\  
 Data File : VY008076.D  
 Acq On : 28 Mar 2022 19:47  
 Operator : SY/MD  
 Sample : VSTDCCC050  
 Misc : 5.00g/5.0mL/MSVOA\_Y/SOIL  
 ALS Vial : 22 Sample Multiplier: 1

Instrument :  
 MSVOA\_Y  
 ClientSampleId :  
 VSTDCCC050EC

Manual Integrations  
 APPROVED

Reviewed By :John  
 Carlone

Quant Time: Mar 29 02:29:27 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y032522S.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Mar 26 03:24:51 2022  
 Response via : Initial Calibration

| Compound                     | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min)   |
|------------------------------|--------|-------|----------|----------|--------|------------|
| -----                        |        |       |          |          |        |            |
| Internal Standards           |        |       |          |          |        | 03/29/2022 |
| 1) Pentafluorobenzene        | 7.795  | 168   | 161325   | 50.000   | ug/l   | 0.00       |
| 34) 1,4-Difluorobenzene      | 8.691  | 114   | 281409   | 50.000   | ug/l   | 0.00       |
| 63) Chlorobenzene-d5         | 11.489 | 117   | 254159   | 50.000   | ug/l   | 0.00       |
| 72) 1,4-Dichlorobenzene-d4   | 13.428 | 152   | 124030   | 50.000   | ug/l   | 0.00       |
| System Monitoring Compounds  |        |       |          |          |        | 03/29/2022 |
| 33) 1,2-Dichloroethane-d4    | 8.142  | 65    | 82264    | 50.825   | ug/l   | 0.00       |
| Spiked Amount                | 50.000 | Range | 50 - 163 | Recovery | =      | 101.660%   |
| 35) Dibromofluoromethane     | 7.722  | 113   | 89684    | 50.831   | ug/l   | 0.00       |
| Spiked Amount                | 50.000 | Range | 54 - 147 | Recovery | =      | 101.660%   |
| 50) Toluene-d8               | 10.179 | 98    | 335764   | 50.512   | ug/l   | 0.00       |
| Spiked Amount                | 50.000 | Range | 49 - 140 | Recovery | =      | 101.020%   |
| 62) 4-Bromofluorobenzene     | 12.483 | 95    | 116598   | 49.718   | ug/l   | 0.00       |
| Spiked Amount                | 50.000 | Range | 25 - 144 | Recovery | =      | 99.440%    |
| Target Compounds             |        |       |          |          |        | Qvalue     |
| 2) Dichlorodifluoromethane   | 1.906  | 85    | 69526    | 47.808   | ug/l   | 97         |
| 3) Chloromethane             | 2.113  | 50    | 84109    | 47.495   | ug/l   | 97         |
| 4) Vinyl Chloride            | 2.253  | 62    | 110677   | 46.972   | ug/l   | 98         |
| 5) Bromomethane              | 2.656  | 94    | 88349    | 45.752   | ug/l   | 98         |
| 6) Chloroethane              | 2.796  | 64    | 74529    | 48.085   | ug/l   | 99         |
| 7) Trichlorofluoromethane    | 3.131  | 101   | 136424   | 45.770   | ug/l   | 94         |
| 8) Diethyl Ether             | 3.534  | 74    | 46619    | 48.298   | ug/l   | 74         |
| 9) 1,1,2-Trichlorotrifluo... | 3.905  | 101   | 86401    | 46.731   | ug/l   | 91         |
| 10) Methyl Iodide            | 4.094  | 142   | 102817   | 47.691   | ug/l   | 91         |
| 11) Tert butyl alcohol       | 4.954  | 59    | 38876    | 231.498  | ug/l # | 95         |
| 12) 1,1-Dichloroethene       | 3.875  | 96    | 82114    | 47.974   | ug/l   | 82         |
| 13) Acrolein                 | 3.735  | 56    | 25689    | 187.941  | ug/l   | 95         |
| 14) Allyl chloride           | 4.485  | 41    | 111245   | 49.398   | ug/l # | 90         |
| 15) Acrylonitrile            | 5.167  | 53    | 112148   | 243.923  | ug/l   | 98         |
| 16) Acetone                  | 3.948  | 43    | 78057    | 210.988  | ug/l # | 80         |
| 17) Carbon Disulfide         | 4.198  | 76    | 234507   | 48.352   | ug/l   | 100        |
| 18) Methyl Acetate           | 4.479  | 43    | 46131    | 47.040   | ug/l # | 78         |
| 19) Methyl tert-butyl Ether  | 5.222  | 73    | 242612   | 49.876   | ug/l   | 96         |
| 20) Methylene Chloride       | 4.716  | 84    | 97722    | 52.117   | ug/l # | 78         |
| 21) trans-1,2-Dichloroethene | 5.222  | 96    | 98197    | 49.547   | ug/l   | 84         |
| 22) Diisopropyl ether        | 6.125  | 45    | 248916   | 51.375   | ug/l # | 88         |
| 23) Vinyl Acetate            | 6.064  | 43    | 808941   | 269.455  | ug/l # | 85         |
| 24) 1,1-Dichloroethane       | 6.027  | 63    | 161503   | 50.350   | ug/l   | 98         |
| 25) 2-Butanone               | 6.984  | 43    | 133652   | 237.394  | ug/l # | 79         |
| 26) 2,2-Dichloropropane      | 6.984  | 77    | 152744   | 50.283   | ug/l   | 93         |
| 27) cis-1,2-Dichloroethene   | 6.990  | 96    | 117806   | 51.281   | ug/l   | 81         |
| 28) Bromochloromethane       | 7.338  | 49    | 43833    | 52.145   | ug/l # | 69         |
| 29) Tetrahydrofuran          | 7.350  | 42    | 88307    | 241.262  | ug/l # | 73         |
| 30) Chloroform               | 7.508  | 83    | 178915   | 51.514   | ug/l   | 98         |
| 31) Cyclohexane              | 7.789  | 56    | 133959   | 44.277   | ug/l # | 83         |
| 32) 1,1,1-Trichloroethane    | 7.704  | 97    | 162949   | 50.777   | ug/l   | 94         |
| 36) 1,1-Dichloropropene      | 7.917  | 75    | 134263   | 48.445   | ug/l   | 96         |
| 37) Ethyl Acetate            | 7.076  | 43    | 61704    | 47.659   | ug/l # | 89         |
| 38) Carbon Tetrachloride     | 7.899  | 117   | 145628   | 48.217   | ug/l   | 97         |
| 39) Methylcyclohexane        | 9.185  | 83    | 170248   | 45.510   | ug/l # | 84         |
| 40) Benzene                  | 8.161  | 78    | 390526   | 49.261   | ug/l   | 95         |

03/29/2022  
 Supervised By :Mahesh  
 Gadoda

03/29/2022

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY032822\  
 Data File : VY008076.D  
 Acq On : 28 Mar 2022 19:47  
 Operator : SY/MD  
 Sample : VSTDCCC050  
 Misc : 5.00g/5.0mL/MSVOA\_Y/SOIL  
 ALS Vial : 22 Sample Multiplier: 1

Instrument :  
 MSVOA\_Y  
 ClientSampleId :  
 VSTDCCC050EC

Manual Integrations  
 APPROVED

Reviewed By :John  
 Carlone

Quant Time: Mar 29 02:29:27 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y032522S.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Mar 26 03:24:51 2022  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc    | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|---------|--------|----------|
| 41) Methacrylonitrile         | 7.313  | 41   | 28684m   | 44.123  | ug/l   |          |
| 42) 1,2-Dichloroethane        | 8.240  | 62   | 106860   | 48.051  | ug/l   |          |
| 43) Isopropyl Acetate         | 8.277  | 43   | 120291   | 47.329  | ug/l # |          |
| 44) Trichloroethene           | 8.941  | 130  | 112010   | 48.860  | ug/l   |          |
| 45) 1,2-Dichloropropane       | 9.215  | 63   | 93320    | 49.947  | ug/l   |          |
| 46) Dibromomethane            | 9.307  | 93   | 59702    | 48.597  | ug/l   |          |
| 47) Bromodichloromethane      | 9.496  | 83   | 140018   | 49.843  | ug/l   |          |
| 48) Methyl methacrylate       | 9.295  | 41   | 51839    | 46.050  | ug/l # |          |
| 49) 1,4-Dioxane               | 9.295  | 88   | 15597    | 955.032 | ug/l # |          |
| 51) 4-Methyl-2-Pentanone      | 10.069 | 43   | 306803   | 239.338 | ug/l # |          |
| 52) Toluene                   | 10.246 | 92   | 260198   | 50.047  | ug/l   |          |
| 53) t-1,3-Dichloropropene     | 10.465 | 75   | 143417   | 49.805  | ug/l   |          |
| 54) cis-1,3-Dichloropropene   | 9.929  | 75   | 160721   | 49.305  | ug/l # |          |
| 55) 1,1,2-Trichloroethane     | 10.648 | 97   | 85411    | 50.350  | ug/l   |          |
| 56) Ethyl methacrylate        | 10.514 | 69   | 111056   | 50.189  | ug/l # |          |
| 57) 1,3-Dichloropropane       | 10.794 | 76   | 138051   | 49.724  | ug/l   |          |
| 58) 2-Chloroethyl Vinyl ether | 9.782  | 63   | 202823   | 221.226 | ug/l # |          |
| 59) 2-Hexanone                | 10.831 | 43   | 209978   | 234.090 | ug/l   |          |
| 60) Dibromochloromethane      | 10.983 | 129  | 104936   | 50.863  | ug/l   |          |
| 61) 1,2-Dibromoethane         | 11.093 | 107  | 83570    | 49.465  | ug/l   |          |
| 64) Tetrachloroethene         | 10.721 | 164  | 89964    | 46.087  | ug/l   |          |
| 65) Chlorobenzene             | 11.520 | 112  | 284215   | 49.595  | ug/l   |          |
| 66) 1,1,1,2-Tetrachloroethane | 11.593 | 131  | 108080   | 51.129  | ug/l   |          |
| 67) Ethyl Benzene             | 11.593 | 91   | 498365   | 49.617  | ug/l   |          |
| 68) m/p-Xylenes               | 11.703 | 106  | 397002   | 99.401  | ug/l   |          |
| 69) o-Xylene                  | 12.032 | 106  | 190391   | 49.788  | ug/l   |          |
| 70) Styrene                   | 12.044 | 104  | 322114   | 51.455  | ug/l   |          |
| 71) Bromoform                 | 12.209 | 173  | 64400    | 50.458  | ug/l # |          |
| 73) Isopropylbenzene          | 12.331 | 105  | 495164   | 48.367  | ug/l   |          |
| 74) N-amyl acetate            | 12.142 | 43   | 105097   | 46.700  | ug/l # |          |
| 75) 1,1,2,2-Tetrachloroethane | 12.587 | 83   | 97238    | 48.866  | ug/l   |          |
| 76) 1,2,3-Trichloropropane    | 12.636 | 75   | 86218m   | 57.025  | ug/l   |          |
| 77) Bromobenzene              | 12.611 | 156  | 116575   | 49.635  | ug/l   |          |
| 78) n-propylbenzene           | 12.672 | 91   | 581176   | 48.128  | ug/l   |          |
| 79) 2-Chlorotoluene           | 12.758 | 91   | 327287   | 48.858  | ug/l   |          |
| 80) 1,3,5-Trimethylbenzene    | 12.812 | 105  | 407907   | 48.636  | ug/l   |          |
| 81) trans-1,4-Dichloro-2-b... | 12.380 | 75   | 33185    | 48.971  | ug/l # |          |
| 82) 4-Chlorotoluene           | 12.855 | 91   | 332053   | 48.310  | ug/l   |          |
| 83) tert-Butylbenzene         | 13.075 | 119  | 377360   | 50.503  | ug/l   |          |
| 84) 1,2,4-Trimethylbenzene    | 13.123 | 105  | 397120   | 48.623  | ug/l   |          |
| 85) sec-Butylbenzene          | 13.257 | 105  | 525525   | 47.649  | ug/l   |          |
| 86) p-Isopropyltoluene        | 13.373 | 119  | 443077   | 48.272  | ug/l   |          |
| 87) 1,3-Dichlorobenzene       | 13.367 | 146  | 230288   | 49.456  | ug/l   |          |
| 88) 1,4-Dichlorobenzene       | 13.446 | 146  | 228685   | 49.246  | ug/l   |          |
| 89) n-Butylbenzene            | 13.696 | 91   | 385714   | 45.674  | ug/l   |          |
| 90) Hexachloroethane          | 13.965 | 117  | 84705    | 47.628  | ug/l   |          |
| 91) 1,2-Dichlorobenzene       | 13.739 | 146  | 204951   | 49.993  | ug/l   |          |
| 92) 1,2-Dibromo-3-Chloropr... | 14.361 | 75   | 13982    | 44.024  | ug/l   |          |
| 93) 1,2,4-Trichlorobenzene    | 15.007 | 180  | 110290   | 44.918  | ug/l   |          |
| 94) Hexachlorobutadiene       | 15.117 | 225  | 53580    | 41.249  | ug/l   |          |
| 95) Naphthalene               | 15.239 | 128  | 245423   | 45.219  | ug/l   |          |
| 96) 1,2,3-Trichlorobenzene    | 15.428 | 180  | 94828    | 44.635  | ug/l   |          |

03/29/2022  
 Supervised By :Mahesh  
 Dadoda

03/29/2022

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY032822\  
Data File : VY008076.D  
Acq On : 28 Mar 2022 19:47  
Operator : SY/MD  
Sample : VSTDCCC050  
Misc : 5.00g/5.0mL/MSVOA\_Y/SOIL  
ALS Vial : 22 Sample Multiplier: 1

Instrument :  
MSVOA\_Y  
ClientSampleId :  
VSTDCCC050EC

Manual Integrations  
APPROVED

Reviewed By :John  
Carlone

Quant Time: Mar 29 02:29:27 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y032522S.M  
Quant Title : SW846 8260  
QLast Update : Sat Mar 26 03:24:51 2022  
Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|

(#) = qualifier out of range (m) = manual integration (+) = signals summed

03/29/2022  
Supervised By :Mahesh  
Dadoda

03/29/2022


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY032822\  
Data File : VY008076.D  
Acq On : 28 Mar 2022 19:47  
Operator : SY/MD  
Sample : VSTDCCC050  
Misc : 5.00g/5.0mL/MSVOA\_Y/SOIL  
ALS Vial : 22 Sample Multiplier: 1


Instrument :  
MSVOA\_Y  
ClientSampleId :  
VSTDCCC050EC

Quant Time: Mar 29 02:29:27 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y032522S.M  
Quant Title : SW846 8260  
QLast Update : Sat Mar 26 03:24:51 2022  
Response via : Initial Calibration

Manual Integrations  
APPROVED


Reviewed By :John  
Carlone





#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 7.795 min Scan# 91  
Delta R.T. 0.006 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47



Instrument :  
MSVOA\_Y  
Client SampleId :  
VSTDCCC050EC



Tgt Ion:168 Resp: 16132  
Ion Ratio Lower Upper  
168 100  
99 54.4 46.9 70.3


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




03/29/2022  
Supervised By :Mahesh  
Dadoda



03/29/2022




#2  
Dichlorodifluoromethane  
Concen: 47.808 ug/l  
RT: 1.906 min Scan# 14  
Delta R.T. -0.000 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47




Tgt Ion: 85 Resp: 69526  
Ion Ratio Lower Upper  
85 100  
87 33.8 16.1 48.2





#3  
Chloromethane  
Concen: 47.495 ug/l  
RT: 2.113 min Scan# 41  
Delta R.T. 0.000 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47



Instrument :  
MSVOA\_Y  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 50 Resp: 84109  
Ion Ratio Lower Upper  
50 100  
52 32.6 24.6 37.0


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/29/2022  
Supervised By :Mahesh  
Dadoda

03/29/2022



#4  
Vinyl Chloride  
Concen: 46.972 ug/l  
RT: 2.253 min Scan# 71  
Delta R.T. 0.006 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47


Tgt Ion: 62 Resp: 110677  
Ion Ratio Lower Upper  
62 100  
64 32.5 25.0 37.6





#5  
Bromomethane  
Concen: 45.752 ug/l  
RT: 2.656 min Scan# 11  
Delta R.T. 0.013 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47



Instrument :  
MSVOA\_Y  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 94 Resp: 88349  
Ion Ratio Lower Upper  
94 100  
96 93.5 73.5 110.3


Manual Integrations  
**APPROVED**

Reviewed By :John  
Carlone






03/29/2022  
Supervised By :Mahesh  
Dadoda


03/29/2022



#6  
Chloroethane  
Concen: 48.085 ug/l  
RT: 2.796 min Scan# 160  
Delta R.T. 0.006 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47

Tgt Ion: 64 Resp: 74529  
Ion Ratio Lower Upper  
64 100  
66 32.1 25.4 38.0







#7  
Trichlorofluoromethane  
Concen: 45.770 ug/l  
RT: 3.131 min Scan# 214  
Delta R.T. 0.006 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47

Instrument :  
MSVOA\_Y  
ClientSampleId :  
VSTDCCC050EC

Tgt Ion:101 Resp: 136424  
Ion Ratio Lower Upper  
101 100  
103 64.5 48.0 72.0



Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/29/2022  
Supervised By :Mahesh  
Dadoda


03/29/2022



#8  
Diethyl Ether  
Concen: 48.298 ug/l  
RT: 3.534 min Scan# 281  
Delta R.T. 0.007 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47

Tgt Ion: 74 Resp: 46619  
Ion Ratio Lower Upper  
74 100  
45 82.8 55.1 165.5







#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 46.731 ug/l  
RT: 3.905 min Scan# 341  
Delta R.T. 0.006 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47

Instrument :  
MSVOA\_Y  
ClientSampleId :  
VSTDCCC050EC

Tgt Ion:101 Resp: 8640  
Ion Ratio Lower Upper  
101 100  
85 43.7 40.6 60.8  
151 74.3 64.9 97.3



Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/29/2022  
Supervised By :Mahesh  
Dadoda


03/29/2022



#10  
Methyl Iodide  
Concen: 47.691 ug/l  
RT: 4.094 min Scan# 373  
Delta R.T. 0.006 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47

Tgt Ion:142 Resp: 102817  
Ion Ratio Lower Upper  
142 100  
127 43.8 41.5 62.3  
141 14.1 11.9 17.9





#11

Tert butyl alcohol

Concen: 231.498 ug/l

RT: 4.954 min Scan# 513

Delta R.T. 0.006 min

Lab File: VY008076.D

Acq: 28 Mar 2022 19:47

Instrument :

MSVOA\_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 59 Resp: 38870

Ion Ratio Lower Upper

59 100



57 9.9 6.4 9.6

Manual Integrations

APPROVED

Reviewed By :John

Carlone





03/29/2022

Supervised By :Mahesh

Dadoda

03/29/2022



#12

1,1-Dichloroethene

Concen: 47.974 ug/l

RT: 3.875 min Scan# 337

Delta R.T. 0.006 min

Lab File: VY008076.D

Acq: 28 Mar 2022 19:47




Tgt Ion: 96 Resp: 82114


Ion Ratio Lower Upper

96 100

61 148.2 145.8 218.6


98 63.6 52.5 78.7





#13  
Acrolein  
Concen: 187.941 ug/l  
RT: 3.735 min Scan# 312  
Delta R.T. 0.012 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47


Instrument :  
MSVOA\_Y  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 56 Resp: 25689  
Ion Ratio Lower Upper  
56 100  
55 68.0 57.6 86.4


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/29/2022  
Supervised By :Mahesh  
Dadoda


03/29/2022



#14  
Allyl chloride  
Concen: 49.398 ug/l  
RT: 4.485 min Scan# 437  
Delta R.T. 0.006 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47


Tgt Ion: 41 Resp: 111245  
Ion Ratio Lower Upper  
41 100  
39 72.5 58.8 88.2  
76 48.0 25.1 37.7#





#15  
Acrylonitrile  
Concen: 243.923 ug/l  
RT: 5.167 min Scan# 547  
Delta R.T. 0.012 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47



Instrument :  
MSVOA\_Y  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 53 Resp: 112148  
Ion Ratio Lower Upper  
53 100  
52 81.3 67.0 100.4  
51 36.2 29.8 44.6


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/29/2022  
Supervised By :Mahesh  
Dadoda


03/29/2022



#16  
Acetone  
Concen: 210.988 ug/l  
RT: 3.948 min Scan# 349  
Delta R.T. 0.000 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47


Tgt Ion: 43 Resp: 78057  
Ion Ratio Lower Upper  
43 100  
58 37.9 22.0 33.0#





#17  
Carbon Disulfide  
Concen: 48.352 ug/l  
RT: 4.198 min Scan# 390  
Delta R.T. 0.006 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47



Instrument :  
MSVOA\_Y  
Client SampleId :  
VSTDCCC050EC



Tgt Ion: 76 Resp: 23450  
Ion Ratio Lower Upper  
76 100  
78 9.4 7.4 11.0


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




03/29/2022  
Supervised By :Mahesh  
Dadoda



03/29/2022




#18  
Methyl Acetate  
Concen: 47.040 ug/l  
RT: 4.479 min Scan# 436  
Delta R.T. 0.007 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47




Tgt Ion: 43 Resp: 46131  
Ion Ratio Lower Upper  
43 100  
74 30.8 16.6 24.8#





#19  
Methyl tert-butyl Ether  
Concen: 49.876 ug/l  
RT: 5.222 min Scan# 51  
Delta R.T. 0.006 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47



Instrument :  
MSVOA\_Y  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 73 Resp: 24261  
Ion Ratio Lower Upper  
73 100  
57 21.4 18.6 27.8


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




03/29/2022  
Supervised By :Mahesh  
Dadoda



03/29/2022



#20  
Methylene Chloride  
Concen: 52.117 ug/l  
RT: 4.716 min Scan# 475  
Delta R.T. 0.006 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47




Tgt Ion: 84 Resp: 97722  
Ion Ratio Lower Upper  
84 100  
49 104.6 114.4 171.6#  
51 33.2 35.3 52.9#  
86 63.2 53.7 80.5





#21  
trans-1,2-Dichloroethene  
Concen: 49.547 ug/l  
RT: 5.222 min Scan# 51  
Delta R.T. 0.006 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47



Instrument :  
MSVOA\_Y  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 96 Resp: 9819  
Ion Ratio Lower Upper  
96 100  
61 121.6 120.5 180.7  
98 62.9 49.8 74.8


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/29/2022  
Supervised By :Mahesh  
Dadoda

03/29/2022



#22  
Diisopropyl ether  
Concen: 51.375 ug/l  
RT: 6.125 min Scan# 706  
Delta R.T. 0.007 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47


Tgt Ion: 45 Resp: 248916  
Ion Ratio Lower Upper  
45 100  
43 55.6 49.3 73.9  
87 35.7 19.2 28.8#  
59 13.1 8.4 12.6#





#23  
 Vinyl Acetate  
 Concen: 269.455 ug/l  
 RT: 6.064 min Scan# 695  
 Delta R.T. 0.007 min  
 Lab File: VY008076.D  
 Acq: 28 Mar 2022 19:47


Instrument :  
 MSVOA\_Y  
 ClientSampleId :  
 VSTDCCC050EC



Tgt Ion: 43 Resp: 80894  
 Ion Ratio Lower Upper  
 43 100  
 86 14.2 7.1 10.7


Manual Integrations  
 APPROVED

Reviewed By :John  
 Carlone






03/29/2022  
 Supervised By :Mahesh  
 Dadoda


03/29/2022



#24  
 1,1-Dichloroethane  
 Concen: 50.350 ug/l  
 RT: 6.027 min Scan# 690  
 Delta R.T. 0.012 min  
 Lab File: VY008076.D  
 Acq: 28 Mar 2022 19:47


Tgt Ion: 63 Resp: 161503  
 Ion Ratio Lower Upper  
 63 100  
 98 7.5 3.6 10.8  
 100 5.3 2.0 5.8





#25  
2-Butanone  
Concen: 237.394 ug/l  
RT: 6.984 min Scan# 847  
Delta R.T. 0.006 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47



Instrument :  
MSVOA\_Y  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 43 Resp: 133652  
Ion Ratio Lower Upper  
43 100  
72 32.0 17.7 26.5


Manual Integrations  
**APPROVED**

Reviewed By :John  
Carlone






03/29/2022  
Supervised By :Mahesh  
Dadoda


03/29/2022



#26  
2,2-Dichloropropane  
Concen: 50.283 ug/l  
RT: 6.984 min Scan# 847  
Delta R.T. 0.006 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47


Tgt Ion: 77 Resp: 152744  
Ion Ratio Lower Upper  
77 100  
97 24.1 10.4 31.4





#27  
cis-1,2-Dichloroethene  
Concen: 51.281 ug/l  
RT: 6.990 min Scan# 847  
Delta R.T. 0.006 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47



Instrument :  
MSVOA\_Y  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 96 Resp: 117800  
Ion Ratio Lower Upper  
96 100  
61 124.6 0.0 319.8  
98 64.3 0.0 127.0


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/29/2022  
Supervised By :Mahesh  
Dadoda


03/29/2022



#28  
Bromochloromethane  
Concen: 52.145 ug/l  
RT: 7.338 min Scan# 905  
Delta R.T. 0.006 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47


Tgt Ion: 49 Resp: 43833  
Ion Ratio Lower Upper  
49 100  
129 2.7 0.0 2.2#  
130 95.2 55.4 83.0#





#29  
Tetrahydrofuran  
Concen: 241.262 ug/l  
RT: 7.350 min Scan# 906  
Delta R.T. 0.006 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47



Instrument :  
MSVOA\_Y  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 42 Resp: 88307  
Ion Ratio Lower Upper  
42 100  
72 55.0 30.3 45.5  
71 51.5 29.0 43.4


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




03/29/2022  
Supervised By :Mahesh  
Dadoda



03/29/2022




#30  
Chloroform  
Concen: 51.514 ug/l  
RT: 7.508 min Scan# 933  
Delta R.T. 0.006 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47




Tgt Ion: 83 Resp: 178915  
Ion Ratio Lower Upper  
83 100  
85 63.3 51.8 77.6





#31  
Cyclohexane  
Concen: 44.277 ug/l  
RT: 7.789 min Scan# 911  
Delta R.T. 0.006 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47


Instrument :  
MSVOA\_Y  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 56 Resp: 133959  
Ion Ratio Lower Upper  
56 100  
69 36.5 26.7 40.1  
84 103.8 67.2 100.8


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/29/2022  
Supervised By :Mahesh  
Dadoda


03/29/2022



#32  
1,1,1-Trichloroethane  
Concen: 50.777 ug/l  
RT: 7.704 min Scan# 965  
Delta R.T. 0.007 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47

Tgt Ion: 97 Resp: 162949  
Ion Ratio Lower Upper  
97 100  
99 64.1 50.7 76.1  
61 38.9 37.6 56.4





#33

1,2-Dichloroethane-d4

Concen: 50.825 ug/l

RT: 8.142 min Scan# 110

Delta R.T. 0.006 min

Lab File: VY008076.D

Acq: 28 Mar 2022 19:47

Instrument :

MSVOA\_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 65 Resp: 82264

Ion Ratio Lower Upper

65 100



67 53.1 0.0 99.4

Manual Integrations

APPROVED

Reviewed By :John

Carlone





03/29/2022

Supervised By :Mahesh

Dadoda

03/29/2022



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.691 min Scan# 1127

Delta R.T. 0.006 min

Lab File: VY008076.D

Acq: 28 Mar 2022 19:47




Tgt Ion:114 Resp: 281409

Ion Ratio Lower Upper

114 100

63 17.1 0.0 43.2


88 14.5 0.0 31.8





#35  
Dibromofluoromethane  
Concen: 50.831 ug/l  
RT: 7.722 min Scan# 911  
Delta R.T. 0.006 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47


Instrument :  
MSVOA\_Y  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 113 Resp: 89684  
Ion Ratio Lower Upper  
113 100  
111 103.4 82.6 124.0  
192 18.3 17.6 26.4



Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/29/2022  
Supervised By :Mahesh  
Dadoda


03/29/2022



#36  
1,1-Dichloropropene  
Concen: 48.445 ug/l  
RT: 7.917 min Scan# 1000  
Delta R.T. -0.000 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47


Tgt Ion: 75 Resp: 134263  
Ion Ratio Lower Upper  
75 100  
110 37.8 17.1 51.2  
77 30.9 24.4 36.6





#37  
Ethyl Acetate  
Concen: 47.659 ug/l  
RT: 7.076 min Scan# 861  
Delta R.T. 0.007 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47



Instrument :  
MSVOA\_Y  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 43 Resp: 61704  
Ion Ratio Lower Upper  
43 100  
61 16.2 9.7 14.5  
70 13.1 7.1 10.7


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/29/2022  
Supervised By :Mahesh  
Dadoda


03/29/2022



#38  
Carbon Tetrachloride  
Concen: 48.217 ug/l  
RT: 7.899 min Scan# 997  
Delta R.T. -0.000 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47

Tgt Ion:117 Resp: 145628  
Ion Ratio Lower Upper  
117 100  
119 93.8 77.9 116.9  
121 30.3 24.8 37.2





#39

Methylcyclohexane

Concen: 45.510 ug/l

RT: 9.185 min Scan# 11

Delta R.T. 0.006 min

Lab File: VY008076.D


Acq: 28 Mar 2022 19:47

Instrument :

MSVOA\_Y

ClientSampleId :

VSTDCCC050EC




Tgt Ion: 83 Resp: 17024

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 83  | 100   |       |       |
| 55  | 64.6  | 66.5  | 99.7  |
| 98  | 52.2  | 37.9  | 56.9  |


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone



03/29/2022  
Supervised By :Mahesh  
Dadoda

03/29/2022



#40

Benzene

Concen: 49.261 ug/l

RT: 8.161 min Scan# 1040




Delta R.T. -0.000 min


Lab File: VY008076.D

Acq: 28 Mar 2022 19:47

Tgt Ion: 78 Resp: 390526

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 78  | 100   |       |       |
| 77  | 23.0  | 20.4  | 30.6  |






#41  
Methacrylonitrile  
Concen: 44.123 ug/l m  
RT: 7.313 min Scan# 901  
Delta R.T. 0.006 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47

Instrument :

MSVOA\_Y

ClientSampleId :



VSTDCCC050EC



Tgt Ion: 41 Resp: 28684  
Ion Ratio Lower Upper  
41 100  
39 143.1 92.9 139.3  
67 0.0 0.0 0.0  
52 0.0 0.0 0.0


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




03/29/2022  
Supervised By :Mahesh  
Dadoda



03/29/2022



#42  
1,2-Dichloroethane  
Concen: 48.051 ug/l  
RT: 8.240 min Scan# 1053  
Delta R.T. 0.006 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47




Tgt Ion: 62 Resp: 106860  
Ion Ratio Lower Upper  
62 100  
98 11.4 0.0 14.8





#43  
Isopropyl Acetate  
Concen: 47.329 ug/l  
RT: 8.277 min Scan# 1059  
Delta R.T. 0.006 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47



Instrument :  
MSVOA\_Y  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 43 Resp: 12029  
Ion Ratio Lower Upper  
43 100  
61 33.2 20.2 30.2  
87 17.5 8.9 13.3


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/29/2022  
Supervised By :Mahesh  
Dadoda


03/29/2022



#44  
Trichloroethene  
Concen: 48.860 ug/l  
RT: 8.941 min Scan# 1168  
Delta R.T. 0.006 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47

Tgt Ion:130 Resp: 112010  
Ion Ratio Lower Upper  
130 100  
95 98.1 0.0 188.2





#45

1,2-Dichloropropane

Concen: 49.947 ug/l

RT: 9.215 min Scan# 1213

Delta R.T. 0.000 min

Lab File: VY008076.D

Acq: 28 Mar 2022 19:47

Instrument :

MSVOA\_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 63 Resp: 93320

Ion Ratio Lower Upper

63 100



65 31.9 23.0 34.6

Manual Integrations

APPROVED



Reviewed By :John

Carlone



03/29/2022  
Supervised By :Mahesh  
Dadoda

03/29/2022



#46

Dibromomethane

Concen: 48.597 ug/l

RT: 9.307 min Scan# 1228

Delta R.T. 0.006 min

Lab File: VY008076.D

Acq: 28 Mar 2022 19:47




Tgt Ion: 93 Resp: 59702


Ion Ratio Lower Upper

93 100

95 84.9 64.3 96.5


174 99.5 83.4 125.0





#47  
Bromodichloromethane  
Concen: 49.843 ug/l  
RT: 9.496 min Scan# 1259  
Delta R.T. -0.000 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47



Instrument :  
MSVOA\_Y  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 83 Resp: 140018  
Ion Ratio Lower Upper  
83 100  
85 65.1 51.9 77.9  
127 8.6 6.6 10.0


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




03/29/2022  
Supervised By :Mahesh  
Dadoda



03/29/2022




#48  
Methyl methacrylate  
Concen: 46.050 ug/l  
RT: 9.295 min Scan# 1226  
Delta R.T. 0.006 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47




Tgt Ion: 41 Resp: 51839  
Ion Ratio Lower Upper  
41 100  
69 108.1 58.2 87.2#  
39 59.5 47.4 71.0





#49  
1,4-Dioxane  
Concen: 955.032 ug/l  
RT: 9.295 min Scan# 1226  
Delta R.T. -0.000 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47



Instrument :  
MSVOA\_Y  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 88 Resp: 15597  
Ion Ratio Lower Upper  
88 100  
43 25.7 32.8 49.2  
58 56.6 59.8 89.6


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/29/2022  
Supervised By :Mahesh  
Dadoda

03/29/2022



#50  
Toluene-d8  
Concen: 50.512 ug/l  
RT: 10.179 min Scan# 1371  
Delta R.T. -0.000 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47


Tgt Ion: 98 Resp: 335764  
Ion Ratio Lower Upper  
98 100  
100 65.9 50.2 75.2





#51  
4-Methyl-2-Pentanone  
Concen: 239.338 ug/l  
RT: 10.069 min Scan# 1353  
Delta R.T. -0.000 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47



Instrument :  
MSVOA\_Y  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 43 Resp: 30680  
Ion Ratio Lower Upper  
43 100  
58 44.6 28.1 42.1


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




03/29/2022  
Supervised By :Mahesh  
Dadoda



03/29/2022




#52  
Toluene  
Concen: 50.047 ug/l  
RT: 10.246 min Scan# 1382  
Delta R.T. 0.006 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47




Tgt Ion: 92 Resp: 260198  
Ion Ratio Lower Upper  
92 100  
91 170.9 142.6 213.8





#53  
t-1,3-Dichloropropene  
Concen: 49.805 ug/l  
RT: 10.465 min Scan# 1418  
Delta R.T. 0.000 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47



Instrument :  
MSVOA\_Y  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 75 Resp: 14341  
Ion Ratio Lower Upper  
75 100  
77 30.4 25.4 38.2


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




03/29/2022  
Supervised By :Mahesh  
Dadoda



03/29/2022



#54  
cis-1,3-Dichloropropene  
Concen: 49.305 ug/l  
RT: 9.929 min Scan# 1330  
Delta R.T. -0.000 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47




Tgt Ion: 75 Resp: 160721  
Ion Ratio Lower Upper  
75 100  
77 33.0 25.6 38.4  
39 39.3 47.6 71.4#





#55  
1,1,2-Trichloroethane  
Concen: 50.350 ug/l  
RT: 10.648 min Scan# 1447  
Delta R.T. 0.006 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47

Instrument :  
MSVOA\_Y  
ClientSampleId :  
VSTDCCC050EC





Tgt Ion: 97 Resp: 8541

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 97  | 100   |       |       |
| 83  | 81.3  | 69.8  | 104.8 |
| 85  | 53.1  | 44.3  | 66.5  |
| 99  | 64.5  | 49.8  | 74.6  |


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




03/29/2022  
Supervised By :Mahesh  
Dadoda

03/29/2022






#56  
Ethyl methacrylate  
Concen: 50.189 ug/l  
RT: 10.514 min Scan# 1426  
Delta R.T. 0.006 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47



Tgt Ion: 69 Resp: 111056


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 69  | 100   |       |       |
| 41  | 53.1  | 63.4  | 95.0# |
| 39  | 31.4  | 35.8  | 53.8# |





#57  
1,3-Dichloropropane  
Concen: 49.724 ug/l  
RT: 10.794 min Scan# 1471  
Delta R.T. 0.006 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47



Instrument :  
MSVOA\_Y  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 76 Resp: 13805  
Ion Ratio Lower Upper  
76 100  
78 33.2 25.0 37.6


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/29/2022  
Supervised By :Mahesh  
Dadoda


03/29/2022



#58  
2-Chloroethyl Vinyl ether  
Concen: 221.226 ug/l  
RT: 9.782 min Scan# 1306  
Delta R.T. 0.000 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47


Tgt Ion: 63 Resp: 202823  
Ion Ratio Lower Upper  
63 100  
106 32.3 21.5 32.3#





#59  
2-Hexanone  
Concen: 234.090 ug/l  
RT: 10.831 min Scan# 1478  
Delta R.T. 0.000 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47



Instrument :  
MSVOA\_Y  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 43 Resp: 209978  
Ion Ratio Lower Upper  
43 100  
58 63.4 24.4 73.4


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/29/2022  
Supervised By :Mahesh  
Dadoda


03/29/2022



#60  
Dibromochloromethane  
Concen: 50.863 ug/l  
RT: 10.983 min Scan# 1503  
Delta R.T. 0.000 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47

Tgt Ion:129 Resp: 104936  
Ion Ratio Lower Upper  
129 100  
127 76.6 38.7 116.1





#61  
1,2-Dibromoethane  
Concen: 49.465 ug/l  
RT: 11.093 min Scan# 11  
Delta R.T. 0.006 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47

Instrument :

MSVOA\_Y



ClientSampleId :

VSTDCCC050EC

Tgt Ion:107 Resp: 83570  
Ion Ratio Lower Upper  
107 100  
109 93.6 75.4 113.2



Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/29/2022  
Supervised By :Mahesh  
Dadoda


03/29/2022



#62  
4-Bromofluorobenzene  
Concen: 49.718 ug/l  
RT: 12.483 min Scan# 1749  
Delta R.T. 0.000 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47


Tgt Ion: 95 Resp: 116598  
Ion Ratio Lower Upper  
95 100  
174 84.4 0.0 177.0  
176 80.8 0.0 175.8





#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 11.489 min Scan# 117  
Delta R.T. 0.000 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47



Instrument :  
MSVOA\_Y  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion:117 Resp: 254159  
Ion Ratio Lower Upper  
117 100  
82 51.5 43.0 64.4  
119 32.0 25.4 38.0


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




03/29/2022  
Supervised By :Mahesh  
Dadoda



03/29/2022




#64  
Tetrachloroethene  
Concen: 46.087 ug/l  
RT: 10.721 min Scan# 1460  
Delta R.T. 0.006 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47




Tgt Ion:164 Resp: 89964  
Ion Ratio Lower Upper  
164 100  
166 123.7 100.4 150.6  
129 95.4 71.4 107.0  
131 94.0 70.8 106.2





#65  
Chlorobenzene  
Concen: 49.595 ug/l  
RT: 11.520 min Scan# 1590  
Delta R.T. 0.006 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47



Instrument :  
MSVOA\_Y  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion:112 Resp: 284215  
Ion Ratio Lower Upper  
112 100  
114 32.1 25.8 38.6


Manual Integrations  
**APPROVED**

Reviewed By :John  
Carlone




03/29/2022  
Supervised By :Mahesh  
Dadoda



03/29/2022




#66  
1,1,1,2-Tetrachloroethane  
Concen: 51.129 ug/l  
RT: 11.593 min Scan# 1603  
Delta R.T. 0.006 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47




Tgt Ion:131 Resp: 108080  
Ion Ratio Lower Upper  
131 100  
133 95.3 47.9 143.6  
119 66.0 30.8 92.3





#67  
Ethyl Benzene  
Concen: 49.617 ug/l  
RT: 11.593 min Scan# 1603  
Delta R.T. 0.000 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47



Instrument :  
MSVOA\_Y  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 91 Resp: 49836  
Ion Ratio Lower Upper  
91 100  
106 33.1 24.5 36.7


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/29/2022  
Supervised By :Mahesh  
Dadoda


03/29/2022



#68  
m/p-Xylenes  
Concen: 99.401 ug/l  
RT: 11.703 min Scan# 1621  
Delta R.T. -0.000 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47


Tgt Ion:106 Resp: 397002  
Ion Ratio Lower Upper  
106 100  
91 192.5 166.6 249.8





#69  
o-Xylene  
Concen: 49.788 ug/l  
RT: 12.032 min Scan# 1675  
Delta R.T. 0.006 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47


Instrument :  
MSVOA\_Y  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion:106 Resp: 19039  
Ion Ratio Lower Upper  
106 100  
91 205.0 110.3 331.0


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/29/2022  
Supervised By :Mahesh  
Dadoda


03/29/2022



#70  
Styrene  
Concen: 51.455 ug/l  
RT: 12.044 min Scan# 1677  
Delta R.T. 0.000 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47


Tgt Ion:104 Resp: 322114  
Ion Ratio Lower Upper  
104 100  
78 45.3 42.4 63.6  
103 53.6 44.4 66.6





#71  
Bromoform  
Concen: 50.458 ug/l  
RT: 12.209 min Scan# 1704  
Delta R.T. -0.000 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47



Instrument :  
MSVOA\_Y  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion:173 Resp: 64400  
Ion Ratio Lower Upper  
173 100  
175 48.3 24.6 74.0  
254 0.0 0.0 0.0


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




03/29/2022  
Supervised By :Mahesh  
Dadoda



03/29/2022



#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 13.428 min Scan# 1904  
Delta R.T. 0.000 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47




Tgt Ion:152 Resp: 124030  
Ion Ratio Lower Upper  
152 100  
115 58.7 28.2 84.7  
150 175.4 0.0 347.6





#73  
Isopropylbenzene  
Concen: 48.367 ug/l  
RT: 12.331 min Scan# 11  
Delta R.T. -0.000 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47



Instrument :  
MSVOA\_Y  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion:105 Resp: 495164  
Ion Ratio Lower Upper  
105 100  
120 27.6 13.4 40.2


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




03/29/2022  
Supervised By :Mahesh  
Dadoda



03/29/2022




#74  
N-ethyl acetate  
Concen: 46.700 ug/l  
RT: 12.142 min Scan# 1693  
Delta R.T. -0.000 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47




Tgt Ion: 43 Resp: 105097  
Ion Ratio Lower Upper  
43 100  
70 53.9 28.7 43.1#  
55 25.8 17.3 25.9  
61 28.9 17.0 25.6#





#75  
1,1,2,2-Tetrachloroethane  
Concen: 48.866 ug/l  
RT: 12.587 min Scan# 1765  
Delta R.T. 0.006 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47



Instrument :  
MSVOA\_Y  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 83 Resp: 97238  
Ion Ratio Lower Upper  
83 100  
131 10.7 5.6 16.8  
85 64.9 32.3 96.8


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




03/29/2022  
Supervised By :Mahesh  
Dadoda



03/29/2022




#76  
1,2,3-Trichloropropane  
Concen: 57.025 ug/l m  
RT: 12.636 min Scan# 1774  
Delta R.T. 0.007 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47




Tgt Ion: 75 Resp: 86218  
Ion Ratio Lower Upper  
75 100  
77 0.0 0.0 0.0





#77  
Bromobenzene  
Concen: 49.635 ug/l  
RT: 12.611 min Scan# 11  
Delta R.T. 0.006 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47



Instrument :  
MSVOA\_Y  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion:156 Resp: 11657  
Ion Ratio Lower Upper  
156 100  
77 172.6 95.5 286.4  
158 96.5 48.0 144.0


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/29/2022  
Supervised By :Mahesh  
Dadoda

03/29/2022



#78  
n-propylbenzene  
Concen: 48.128 ug/l  
RT: 12.672 min Scan# 1780  
Delta R.T. 0.000 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47

Tgt Ion: 91 Resp: 581176  
Ion Ratio Lower Upper  
91 100  
120 24.0 11.3 33.9







#79  
2-Chlorotoluene  
Concen: 48.858 ug/l  
RT: 12.758 min Scan# 1794  
Delta R.T. -0.000 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47

Instrument :  
MSVOA\_Y  
ClientSampleId :  
VSTDCCC050EC

Tgt Ion: 91 Resp: 32728  
Ion Ratio Lower Upper  
91 100  
126 36.5 16.9 50.6



Manual Integrations  
**APPROVED**

Reviewed By :John  
Carlone






03/29/2022  
Supervised By :Mahesh  
Dadoda


03/29/2022



#80  
1,3,5-Trimethylbenzene  
Concen: 48.636 ug/l  
RT: 12.812 min Scan# 1803  
Delta R.T. 0.000 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47


Tgt Ion:105 Resp: 407907  
Ion Ratio Lower Upper  
105 100  
120 50.3 24.4 73.4





#81  
trans-1,4-Dichloro-2-butene  
Concen: 48.971 ug/l  
RT: 12.380 min Scan# 1732  
Delta R.T. -0.000 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47


Instrument :  
MSVOA\_Y  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 75 Resp: 33185  
Ion Ratio Lower Upper  
75 100  
53 78.0 85.3 127.9  
89 44.8 37.0 55.4


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/29/2022  
Supervised By :Mahesh  
Dadoda

03/29/2022



#82  
4-Chlorotoluene  
Concen: 48.310 ug/l  
RT: 12.855 min Scan# 1810  
Delta R.T. 0.000 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47


Tgt Ion: 91 Resp: 332053  
Ion Ratio Lower Upper  
91 100  
126 37.2 17.0 51.0





#83  
tert-Butylbenzene  
Concen: 50.503 ug/l  
RT: 13.075 min Scan# 119  
Delta R.T. -0.000 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47


Instrument :  
MSVOA\_Y  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion:119 Resp: 377360  
Ion Ratio Lower Upper  
119 100  
91 59.8 34.8 104.4  
134 26.8 13.3 39.9



Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/29/2022  
Supervised By :Mahesh  
Dadoda


03/29/2022



#84  
1,2,4-Trimethylbenzene  
Concen: 48.623 ug/l  
RT: 13.123 min Scan# 1854  
Delta R.T. 0.006 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47

Tgt Ion:105 Resp: 397120  
Ion Ratio Lower Upper  
105 100  
120 47.7 22.3 66.9







#85  
 sec-Butylbenzene  
 Concen: 47.649 ug/l  
 RT: 13.257 min Scan# 1875  
 Delta R.T. 0.006 min  
 Lab File: VY008076.D  
 Acq: 28 Mar 2022 19:47

Instrument :  
 MSVOA\_Y  
 ClientSampleId :  
 VSTDCCC050EC

Tgt Ion:105 Resp: 52552  
 Ion Ratio Lower Upper  
 105 100  
 134 21.2 10.2 30.4



Manual Integrations  
 APPROVED

Reviewed By :John  
 Carlone






03/29/2022  
 Supervised By :Mahesh  
 Dadoda


03/29/2022



#86  
 p-Isopropyltoluene  
 Concen: 48.272 ug/l  
 RT: 13.373 min Scan# 1895  
 Delta R.T. 0.006 min  
 Lab File: VY008076.D  
 Acq: 28 Mar 2022 19:47


Tgt Ion:119 Resp: 443077  
 Ion Ratio Lower Upper  
 119 100  
 134 27.0 13.2 39.5  
 91 22.1 12.4 37.2





#87  
1,3-Dichlorobenzene  
Concen: 49.456 ug/l  
RT: 13.367 min Scan# 1894  
Delta R.T. 0.000 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47


Instrument :  
MSVOA\_Y  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion:146 Resp: 230288  
Ion Ratio Lower Upper  
146 100  
111 38.5 19.5 58.5  
148 63.8 31.6 95.0


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




03/29/2022  
Supervised By :Mahesh  
Dadoda



03/29/2022



#88  
1,4-Dichlorobenzene  
Concen: 49.246 ug/l  
RT: 13.446 min Scan# 1907  
Delta R.T. 0.000 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47




Tgt Ion:146 Resp: 228685  
Ion Ratio Lower Upper  
146 100  
111 37.9 19.3 57.8  
148 63.6 32.6 97.8





#89  
n-Butylbenzene  
Concen: 45.674 ug/l  
RT: 13.696 min Scan# 1948  
Delta R.T. 0.000 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47



Instrument :  
MSVOA\_Y  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion: 91 Resp: 385714  
Ion Ratio Lower Upper  
91 100  
92 53.9 26.3 78.9  
134 27.9 13.0 38.9


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/29/2022  
Supervised By :Mahesh  
Dadoda


03/29/2022



#90  
Hexachloroethane  
Concen: 47.628 ug/l  
RT: 13.965 min Scan# 1992  
Delta R.T. 0.006 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47


Tgt Ion:117 Resp: 84705  
Ion Ratio Lower Upper  
117 100  
201 70.3 43.0 128.8





#91  
1,2-Dichlorobenzene  
Concen: 49.993 ug/l  
RT: 13.739 min Scan# 1955  
Delta R.T. 0.000 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47



Instrument :  
MSVOA\_Y  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion:146 Resp: 20495  
Ion Ratio Lower Upper  
146 100  
111 39.3 20.3 60.9  
148 63.4 32.3 96.9


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




03/29/2022  
Supervised By :Mahesh  
Dadoda



03/29/2022




#92  
1,2-Dibromo-3-Chloropropane  
Concen: 44.024 ug/l  
RT: 14.361 min Scan# 2057  
Delta R.T. 0.006 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47




Tgt Ion: 75 Resp: 13982  
Ion Ratio Lower Upper  
75 100  
155 100.5 38.0 113.9  
157 125.0 47.8 143.3





#93  
1,2,4-Trichlorobenzene  
Concen: 44.918 ug/l  
RT: 15.007 min Scan# 2163  
Delta R.T. 0.000 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47



Instrument :  
MSVOA\_Y  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion:180 Resp: 110290  
Ion Ratio Lower Upper  
180 100  
182 94.9 47.3 141.9  
145 31.6 14.6 43.8


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




03/29/2022  
Supervised By :Mahesh  
Dadoda



03/29/2022




#94  
Hexachlorobutadiene  
Concen: 41.249 ug/l  
RT: 15.117 min Scan# 2181  
Delta R.T. 0.006 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47




Tgt Ion:225 Resp: 53580  
Ion Ratio Lower Upper  
225 100  
223 62.2 30.8 92.3  
227 64.6 31.6 94.7





#95  
Naphthalene  
Concen: 45.219 ug/l  
RT: 15.239 min Scan# 21  
Delta R.T. -0.000 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47


Instrument :  
MSVOA\_Y  
ClientSampleId :  
VSTDCCC050EC



Tgt Ion:128 Resp: 24542  
Ion Ratio Lower Upper  
128 100  
127 12.8 11.0 16.4  
129 11.0 8.4 12.6


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




03/29/2022  
Supervised By :Mahesh  
Dadoda

03/29/2022



#96  
1,2,3-Trichlorobenzene  
Concen: 44.635 ug/l  
RT: 15.428 min Scan# 2232  
Delta R.T. 0.006 min  
Lab File: VY008076.D  
Acq: 28 Mar 2022 19:47



Tgt Ion:180 Resp: 94828  
Ion Ratio Lower Upper  
180 100  
182 94.7 47.3 141.9  
145 33.1 15.3 45.9

