



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032921\
 Data File : VY004213.D
 Acq On : 29 Mar 2021 15:48
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00g/5mL/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
 Title : SW846 8260
 Last Update : Tue Mar 30 06:38:14 2021



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.9	14584	PASS
75	95	30	60	54.5	40037	PASS
95	95	100	100	100.0	73413	PASS
96	95	5	9	6.7	4916	PASS
173	174	0.00	2	1.1	682	PASS
174	95	50	100	86.7	63635	PASS
175	174	5	9	7.8	4939	PASS
176	174	95	101	95.3	60650	PASS
177	176	5	9	6.6	4008	PASS