

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032923\
 Data File : VY013133.D
 Acq On : 29 Mar 2023 16:35
 Operator : KP/MD
 Sample : 02126-01
 Misc : 5.46g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RS-A

Quant Time: Mar 29 23:33:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	71872	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	126188	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	115085	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	42455	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	46179	64.713	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	129.420%	
35) Dibromofluoromethane	7.716	113	43087	56.579	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	113.160%	
50) Toluene-d8	10.179	98	150539	54.072	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	108.140%	
62) 4-Bromofluorobenzene	12.483	95	42247	44.959	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	89.920%	
Target Compounds						
16) Acetone	3.942	43	3846	13.731	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032923\
Data File : VY013133.D
Acq On : 29 Mar 2023 16:35
Operator : KP/MD
Sample : 02126-01
Misc : 5.46g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RS-A

Quant Time: Mar 29 23:33:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration

