

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032923\
 Data File : VY013132.D
 Acq On : 29 Mar 2023 13:42
 Operator : KP/MD
 Sample : 02117-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 NB-8-032823

Quant Time: Mar 29 23:32:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 03 01:13:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	18613	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	28959	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	26342	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	10152	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	11738	63.516	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	127.040%
35) Dibromofluoromethane	7.716	113	11139	63.736	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	127.480%
50) Toluene-d8	10.179	98	34362	53.782	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.560%
62) 4-Bromofluorobenzene	12.483	95	9411	43.640	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	87.280%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032923\
Data File : VY013132.D
Acq On : 29 Mar 2023 13:42
Operator : KP/MD
Sample : 02117-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
NB-8-032823

Quant Time: Mar 29 23:32:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y030223S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 03 01:13:13 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY013132.D
Acq: 29 Mar 2023 13:42




Instrument :
MSVOA_Y
ClientSampleId :
NB-8-032823


Tgt Ion:168 Resp: 18613
Ion Ratio Lower Upper
168 100
99 52.2 41.7 62.5



#33
1,2-Dichloroethane-d4
Concen: 63.516 ug/l
RT: 8.136 min Scan# 1036
Delta R.T. -0.007 min
Lab File: VY013132.D
Acq: 29 Mar 2023 13:42

Tgt Ion: 65 Resp: 11738
Ion Ratio Lower Upper
65 100
67 52.8 0.0 104.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.691 min Scan# 1127

Delta R.T. -0.000 min

Lab File: VY013132.D

Acq: 29 Mar 2023 13:42

Instrument :

MSVOA_Y

ClientSampleId :

NB-8-032823





Tgt Ion:114 Resp: 28959

Ion Ratio Lower Upper

114 100

63 20.2 0.0 37.4

88 16.5 0.0 28.6



#35

Dibromofluoromethane

Concen: 63.736 ug/l

RT: 7.716 min Scan# 967

Delta R.T. -0.000 min

Lab File: VY013132.D

Acq: 29 Mar 2023 13:42




Tgt Ion:113 Resp: 11139


Ion Ratio Lower Upper

113 100

111 101.5 82.2 123.2

192 20.9 16.8 25.2









#50
Toluene-d8
Concen: 53.782 ug/l
RT: 10.179 min Scan# 1371
Delta R.T. -0.000 min
Lab File: VY013132.D
Acq: 29 Mar 2023 13:42




Instrument :
MSVOA_Y
ClientSampleId :
NB-8-032823


Tgt Ion: 98 Resp: 34362
Ion Ratio Lower Upper
98 100
100 63.7 50.9 76.3



#62
4-Bromofluorobenzene
Concen: 43.640 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VY013132.D
Acq: 29 Mar 2023 13:42

Tgt Ion: 95 Resp: 9411
Ion Ratio Lower Upper
95 100
174 89.3 0.0 176.6
176 85.1 0.0 170.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.489 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY013132.D
Acq: 29 Mar 2023 13:42

Instrument :
MSVOA_Y
ClientSampleId :
NB-8-032823

Tgt Ion:117 Resp: 26342
Ion Ratio Lower Upper
117 100
82 57.0 45.4 68.0
119 33.1 25.6 38.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. -0.000 min
Lab File: VY013132.D
Acq: 29 Mar 2023 13:42

Tgt Ion:152 Resp: 10152
Ion Ratio Lower Upper
152 100
115 54.4 27.1 81.3
150 154.8 0.0 347.8

