

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032924\
 Data File : VY017796.D
 Acq On : 29 Mar 2024 17:22
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Mar 30 05:47:43 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032924S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 05:41:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.777	168	84	50.000	ug/l	#-0.02
34) 1,4-Difluorobenzene	8.685	114	681	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	1349	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	1230	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	941	1294.569	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 2589.140%#	
35) Dibromofluoromethane	7.716	113	618	157.047	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 314.100%#	
50) Toluene-d8	10.179	98	1543	104.970	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 209.940%#	
62) 4-Bromofluorobenzene	12.483	95	15289	3418.619	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 6837.240%#	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY032924\
Data File : VY017796.D
Acq On : 29 Mar 2024 17:22
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Mar 30 05:47:43 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032924S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 05:41:07 2024
Response via : Initial Calibration

