



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY033021\
 Data File : VY004221.D
 Acq On : 30 Mar 2021 08:26
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00g/5mL/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
 Title : SW846 8260
 Last Update : Tue Mar 30 06:38:14 2021



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.4	16143	PASS
75	95	30	60	53.5	42226	PASS
95	95	100	100	100.0	78963	PASS
96	95	5	9	6.6	5234	PASS
173	174	0.00	2	1.1	697	PASS
174	95	50	100	82.2	64891	PASS
175	174	5	9	7.4	4825	PASS
176	174	95	101	97.7	63424	PASS
177	176	5	9	5.9	3730	PASS