

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY033023\
 Data File : VY013147.D
 Acq On : 30 Mar 2023 12:31
 Operator : KP/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Mar 31 03:58:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y033023S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 31 03:45:49 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	104406	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	172243	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	157589	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	71231	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	61986	55.244	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 110.480%	
35) Dibromofluoromethane	7.716	113	57677	51.259	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 102.520%	
50) Toluene-d8	10.179	98	202963	47.780	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 95.560%	
62) 4-Bromofluorobenzene	12.483	95	64332	46.527	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 93.060%	
Target Compounds						
					Qvalue	
16) Acetone	3.948	43	937	2.731	ug/l #	87
20) Methylene Chloride	4.722	84	9813	3.265	ug/l #	88
93) 1,2,4-Trichlorobenzene	15.013	180	1846	1.463	ug/l	94
95) Naphthalene	15.239	128	4198	4.702	ug/l	98
96) 1,2,3-Trichlorobenzene	15.422	180	2033	1.828	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY033023\
Data File : VY013147.D
Acq On : 30 Mar 2023 12:31
Operator : KP/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Mar 31 03:58:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y033023S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 31 03:45:49 2023
Response via : Initial Calibration

