



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY033023\
Data File : VY013140.D
Acq On : 30 Mar 2023 08:50
Operator : KP/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y033023S.M
Title : SW846 8260
Last Update : Mon Apr 03 13:40:57 2023



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.1	10971	PASS
75	95	30	60	53.7	27893	PASS
95	95	100	100	100.0	51939	PASS
96	95	5	9	6.6	3424	PASS
173	174	0.00	2	1.7	796	PASS
174	95	50	100	88.2	45805	PASS
175	174	5	9	7.7	3527	PASS
176	174	95	101	95.9	43907	PASS
177	176	5	9	6.4	2792	PASS