

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY033122\
 Data File : VY008128.D
 Acq On : 31 Mar 2022 10:29
 Operator : SY/MD
 Sample : VY0331SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0331SBL01

Quant Time: Mar 31 14:21:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 26 03:24:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	125437	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	226935	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	211739	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	93596	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	69440	55.176	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	110.360%
35) Dibromofluoromethane	7.722	113	70975	49.883	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.760%
50) Toluene-d8	10.185	98	272573	50.848	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.700%
62) 4-Bromofluorobenzene	12.483	95	92730	49.032	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	98.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY033122\
Data File : VY008128.D
Acq On : 31 Mar 2022 10:29
Operator : SY/MD
Sample : VY0331SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0331SBL01

Quant Time: Mar 31 14:21:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 26 03:24:51 2022
Response via : Initial Calibration

