

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY033122\
 Data File : VY008148.D
 Acq On : 31 Mar 2022 19:02
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 01 04:35:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 26 03:24:51 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	87	0.00
2 T	Dichlorodifluoromethane	0.451	0.400	11.3	72	0.00
3 P	Chloromethane	0.549	0.479	12.8	72	0.00
4 C	Vinyl Chloride	0.730	0.644	11.8#	72	0.00
5 T	Bromomethane	0.598	0.577	3.5	84	0.00
6 T	Chloroethane	0.480	0.427	11.0	73	0.00
7 T	Trichlorofluoromethane	0.924	0.845	8.5	77	0.00
8 T	Diethyl Ether	0.299	0.298	0.3	82	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.573	0.530	7.5	76	0.00
10 T	Methyl Iodide	0.592	0.520	12.2	63	0.00
11 T	Tert butyl alcohol	0.052	0.047	9.6	73	0.00
12 CM	1,1-Dichloroethene	0.530	0.498	6.0#	78	0.00
13 T	Acrolein	0.042	0.039	7.1	82	0.00
14 T	Allyl chloride	0.698	0.646	7.4	75	0.00
15 T	Acrylonitrile	0.142	0.139	2.1	78	0.00
16 T	Acetone	0.115	0.095	17.4	64	0.00
17 T	Carbon Disulfide	1.503	1.334	11.2	72	0.00
18 T	Methyl Acetate	0.386	0.288	25.4#	76	0.00
19 T	Methyl tert-butyl Ether	1.508	1.477	2.1	80	0.00
20 T	Methylene Chloride	0.673	0.574	14.7	80	0.00
21 T	trans-1,2-Dichloroethene	0.614	0.593	3.4	80	0.00
22 T	Diisopropyl ether	1.502	1.470	2.1	79	0.00
23 T	Vinyl Acetate	0.930	0.968	-4.1	77	0.00
24 P	1,1-Dichloroethane	0.994	0.961	3.3	79	0.00
25 T	2-Butanone	0.174	0.161	7.5	72	0.00
26 T	2,2-Dichloropropane	0.941	0.884	6.1	76	0.00
27 T	cis-1,2-Dichloroethene	0.712	0.700	1.7	82	0.00
28 T	Bromochloromethane	0.322	0.252	21.7	78	0.00
29 T	Tetrahydrofuran	0.113	0.109	3.5	75	0.00
30 C	Chloroform	1.076	1.077	-0.1#	82	0.00
31 T	Cyclohexane	0.938	0.784	16.4	72	0.00
32 T	1,1,1-Trichloroethane	0.995	0.966	2.9	79	0.00
33 S	1,2-Dichloroethane-d4	0.502	0.479	4.6	81	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	89	0.00
35 S	Dibromofluoromethane	0.313	0.305	2.6	85	0.00
36 T	1,1-Dichloropropene	0.492	0.446	9.3	76	0.00
37 T	Ethyl Acetate	0.230	0.216	6.1	76	0.00
38 T	Carbon Tetrachloride	0.537	0.501	6.7	78	0.00
39 T	Methylcyclohexane	0.665	0.581	12.6	74	0.00
40 TM	Benzene	1.409	1.342	4.8	81	0.00
41 T	Methacrylonitrile	0.116	0.110	5.2	86	0.00
42 TM	1,2-Dichloroethane	0.395	0.375	5.1	80	0.00
43 T	Isopropyl Acetate	0.452	0.419	7.3	77	0.00
44 TM	Trichloroethene	0.407	0.387	4.9	81	0.00
45 C	1,2-Dichloropropane	0.332	0.316	4.8#	80	0.00
46 T	Dibromomethane	0.218	0.212	2.8	81	0.00
47 T	Bromodichloromethane	0.499	0.486	2.6	82	0.00
48 T	Methyl methacrylate	0.200	0.182	9.0	75	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY033122\
 Data File : VY008148.D
 Acq On : 31 Mar 2022 19:02
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 01 04:35:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 26 03:24:51 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	79	0.00
50 S	Toluene-d8	1.181	1.115	5.6	81	0.00
51 T	4-Methyl-2-Pentanone	0.228	0.215	5.7	76	0.00
52 CM	Toluene	0.924	0.904	2.2#	81	0.00
53 T	t-1,3-Dichloropropene	0.512	0.497	2.9	80	0.00
54 T	cis-1,3-Dichloropropene	0.579	0.562	2.9	80	0.00
55 T	1,1,2-Trichloroethane	0.301	0.303	-0.7	82	0.00
56 T	Ethyl methacrylate	0.393	0.386	1.8	77	0.00
57 T	1,3-Dichloropropane	0.493	0.481	2.4	81	0.00
58 T	2-Chloroethyl Vinyl ether	0.163	0.143	12.3	80	0.00
59 T	2-Hexanone	0.159	0.147	7.5	73	0.00
60 T	Dibromochloromethane	0.367	0.369	-0.5	82	0.00
61 T	1,2-Dibromoethane	0.300	0.299	0.3	81	0.00
62 S	4-Bromofluorobenzene	0.417	0.393	5.8	80	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	90	0.00
64 T	Tetrachloroethene	0.384	0.353	8.1	81	0.00
65 PM	Chlorobenzene	1.127	1.100	2.4	82	0.00
66 T	1,1,1,2-Tetrachloroethane	0.416	0.420	-1.0	83	0.00
67 C	Ethyl Benzene	1.976	1.887	4.5#	80	0.00
68 T	m/p-Xylenes	0.786	0.759	3.4	80	0.00
69 T	o-Xylene	0.752	0.735	2.3	80	0.00
70 T	Styrene	1.232	1.233	-0.1	81	0.00
71 P	Bromoform	0.251	0.258	-2.8	82	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
73 T	Isopropylbenzene	4.127	3.838	7.0	79	0.00
74 T	N-amyl acetate	0.907	0.798	12.0	75	0.00
75 P	1,1,2,2-Tetrachloroethane	0.802	0.768	4.2	80	0.00
76 T	1,2,3-Trichloropropane	0.610	0.560	8.2	82	0.00
77 T	Bromobenzene	0.947	0.921	2.7	82	0.00
78 T	n-propylbenzene	4.868	4.456	8.5	79	0.00
79 T	2-Chlorotoluene	2.700	2.516	6.8	80	0.00
80 T	1,3,5-Trimethylbenzene	3.381	3.166	6.4	80	0.00
81 T	trans-1,4-Dichloro-2-butene	0.273	0.250	8.4	75	0.00
82 T	4-Chlorotoluene	2.771	2.520	9.1	78	0.00
83 T	tert-Butylbenzene	3.012	2.835	5.9	79	0.00
84 T	1,2,4-Trimethylbenzene	3.292	3.077	6.5	80	0.00
85 T	sec-Butylbenzene	4.446	4.124	7.2	79	0.00
86 T	p-Isopropyltoluene	3.700	3.482	5.9	79	0.00
87 T	1,3-Dichlorobenzene	1.877	1.802	4.0	80	0.00
88 T	1,4-Dichlorobenzene	1.872	1.788	4.5	80	0.00
89 T	n-Butylbenzene	3.404	3.036	10.8	76	0.00
90 T	Hexachloroethane	0.717	0.673	6.1	79	0.00
91 T	1,2-Dichlorobenzene	1.653	1.601	3.1	80	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.128	0.114	10.9	75	0.00
93 T	1,2,4-Trichlorobenzene	0.990	0.892	9.9	77	0.00
94 T	Hexachlorobutadiene	0.524	0.452	13.7	77	0.00
95 T	Naphthalene	2.188	2.055	6.1	78	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY033122\
Data File : VY008148.D
Acq On : 31 Mar 2022 19:02
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Apr 01 04:35:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032522S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 26 03:24:51 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.856	0.776	9.3	78	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6