

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040121\
 Data File : VY004240.D
 Acq On : 01 Apr 2021 12:57
 Operator : SY/MD
 Sample : M1854-02
 Misc : 5.00g/5mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-1S1-(0-2)

Quant Time: Apr 02 02:09:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 30 06:38:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	196635	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	339208	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	313520	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	135873	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	122093	64.59	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 129.18%		
35) Dibromofluoromethane	7.728	113	104756	55.43	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 110.86%		
50) Toluene-d8	10.185	98	414172	58.26	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	= 116.52%		
62) 4-Bromofluorobenzene	12.483	95	137243	56.33	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	= 112.66%		
Target Compounds						
16) Acetone	3.966	43	3990	10.13	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040121\
Data File : VY004240.D
Acq On : 01 Apr 2021 12:57
Operator : SY/MD
Sample : M1854-02
Misc : 5.00g/5mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-1S1-(0-2)

Quant Time: Apr 02 02:09:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 30 06:38:14 2021
Response via : Initial Calibration

