

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040121\
 Data File : VY004241.D
 Acq On : 01 Apr 2021 13:20
 Operator : SY/MD
 Sample : M1854-05
 Misc : 6.16g/5mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-1D-(4-10)

Quant Time: Apr 02 02:09:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 30 06:38:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	181452	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	304216	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	273400	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	114236	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	108709	62.32	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	124.64%
35) Dibromofluoromethane	7.728	113	93442	55.13	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.26%
50) Toluene-d8	10.185	98	368762	57.84	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	115.68%
62) 4-Bromofluorobenzene	12.483	95	122087	55.87	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	111.74%
Target Compounds						
16) Acetone	3.960	43	6060	16.68	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040121\
Data File : VY004241.D
Acq On : 01 Apr 2021 13:20
Operator : SY/MD
Sample : M1854-05
Misc : 6.16g/5mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-1D-(4-10)

Quant Time: Apr 02 02:09:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 30 06:38:14 2021
Response via : Initial Calibration

