

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040121\
 Data File : VY004242.D
 Acq On : 01 Apr 2021 13:43
 Operator : SY/MD
 Sample : M1854-06
 Misc : 5.08g/5mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-1D1-(4-7.5)

Quant Time: Apr 02 02:09:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 30 06:38:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	214407	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	373437	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	351610	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	159214	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	133008	64.53	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 129.06%		
35) Dibromofluoromethane	7.728	113	116182	55.84	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 111.68%		
50) ToLuene-d8	10.185	98	460308	58.82	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	= 117.64%		
62) 4-Bromofluorobenzene	12.483	95	158966	59.27	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	= 118.54%		
Target Compounds						Qvalue
16) Acetone	3.967	43	3745	8.72	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040121\
Data File : VY004242.D
Acq On : 01 Apr 2021 13:43
Operator : SY/MD
Sample : M1854-06
Misc : 5.08g/5mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-1D1-(4-7.5)

Quant Time: Apr 02 02:09:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 30 06:38:14 2021
Response via : Initial Calibration

