

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040121\
 Data File : VY004248.D
 Acq On : 01 Apr 2021 16:27
 Operator : SY/MD
 Sample : M1900-04
 Misc : 5.40g/5mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CHRT-20111

Quant Time: Apr 02 02:11:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 30 06:38:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	67173	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	99577	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	67320	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	17405	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	24641	38.16	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	76.32%
35) Dibromofluoromethane	7.728	113	26732	48.18	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.36%
50) Toluene-d8	10.185	98	114183	54.72	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	109.44%
62) 4-Bromofluorobenzene	12.483	95	22155	30.98	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	61.96%
Target Compounds						
16) Acetone	3.954	43	3013	22.40	ug/l	# 65
84) 1,2,4-Trimethylbenzene	13.129	105	3649	3.51	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040121\
Data File : VY004248.D
Acq On : 01 Apr 2021 16:27
Operator : SY/MD
Sample : M1900-04
Misc : 5.40g/5mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CHRT-20111

Quant Time: Apr 02 02:11:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 30 06:38:14 2021
Response via : Initial Calibration

