

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040121\
 Data File : VY004235.D
 Acq On : 01 Apr 2021 10:02
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 02 02:06:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 30 06:38:14 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	125	0.00
2 T	Dichlorodifluoromethane	0.433	0.419	3.2	121	0.00
3 P	Chloromethane	0.584	0.603	-3.3	128	0.00
4 C	Vinyl Chloride	0.651	0.740	-13.7#	143	0.00
5 T	Bromomethane	0.358	0.315	12.0	113	0.00
6 T	Chloroethane	0.422	0.402	4.7	121	0.00
7 T	Trichlorofluoromethane	0.847	0.820	3.2	121	0.00
8 T	Diethyl Ether	0.272	0.274	-0.7	126	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.461	0.471	-2.2	130	0.00
10 T	Methyl Iodide	0.533	0.512	3.9	119	0.00
11 T	Tert butyl alcohol	0.039	0.042	-7.7	126	0.00
12 CM	1,1-Dichloroethene	0.452	0.455	-0.7#	126	0.00
13 T	Acrolein	0.037	0.030	18.9	123	0.00
14 T	Allyl chloride	0.721	0.853	-18.3	147	0.00
15 T	Acrylonitrile	0.129	0.139	-7.8	130	0.00
16 T	Acetone	0.100	0.082	18.0	103	0.00
17 T	Carbon Disulfide	1.189	1.326	-11.5	134	0.00
18 T	Methyl Acetate	0.321	0.345	-7.5	131	0.00
19 T	Methyl tert-butyl Ether	1.336	1.338	-0.1	123	0.00
20 T	Methylene Chloride	0.554	0.506	8.7	128	0.00
21 T	trans-1,2-Dichloroethene	0.508	0.525	-3.3	127	0.00
22 T	Diisopropyl ether	1.583	1.818	-14.8	145	0.00
23 T	Vinyl Acetate	1.065	1.181	-10.9	136	0.00
24 P	1,1-Dichloroethane	0.862	0.948	-10.0	137	0.00
25 T	2-Butanone	0.167	0.163	2.4	115	0.00
26 T	2,2-Dichloropropane	0.807	0.848	-5.1	132	0.00
27 T	cis-1,2-Dichloroethene	0.575	0.578	-0.5	125	0.00
28 T	Bromochloromethane	0.351	0.366	-4.3	139	0.00
29 T	Tetrahydrofuran	0.118	0.128	-8.5	132	0.00
30 C	Chloroform	0.917	0.913	0.4#	125	0.00
31 T	Cyclohexane	0.849	0.922	-8.6	143	0.00
32 T	1,1,1-Trichloroethane	0.839	0.830	1.1	124	0.00
33 S	1,2-Dichloroethane-d4	0.481	0.445	7.5	118	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	126	0.00
35 S	Dibromofluoromethane	0.279	0.255	8.6	118	0.00
36 T	1,1-Dichloropropene	0.456	0.476	-4.4	131	0.00
37 T	Ethyl Acetate	0.268	0.281	-4.9	131	0.00
38 T	Carbon Tetrachloride	0.397	0.430	-8.3	131	0.00
39 T	Methylcyclohexane	0.565	0.604	-6.9	134	0.00
40 TM	Benzene	1.294	1.346	-4.0	131	0.00
41 T	Methacrylonitrile	0.148	0.151	-2.0	132	0.00
42 TM	1,2-Dichloroethane	0.407	0.400	1.7	122	0.00
43 T	Isopropyl Acetate	0.507	0.545	-7.5	131	0.00
44 TM	Trichloroethene	0.412	0.373	9.5	115	0.00
45 C	1,2-Dichloropropane	0.320	0.346	-8.1#	136	0.00
46 T	Dibromomethane	0.188	0.182	3.2	120	0.00
47 T	Bromodichloromethane	0.449	0.456	-1.6	125	0.00
48 T	Methyl methacrylate	0.230	0.253	-10.0	131	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040121\
 Data File : VY004235.D
 Acq On : 01 Apr 2021 10:02
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 02 02:06:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 30 06:38:14 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.002	33.3#	123	0.00
50 S	Toluene-d8	1.048	0.983	6.2	123	0.00
51 T	4-Methyl-2-Pentanone	0.258	0.276	-7.0	128	0.00
52 CM	Toluene	0.832	0.853	-2.5#	128	0.00
53 T	t-1,3-Dichloropropene	0.478	0.494	-3.3	127	0.00
54 T	cis-1,3-Dichloropropene	0.534	0.567	-6.2	131	0.00
55 T	1,1,2-Trichloroethane	0.258	0.254	1.6	123	0.00
56 T	Ethyl methacrylate	0.353	0.370	-4.8	127	0.00
57 T	1,3-Dichloropropane	0.444	0.452	-1.8	126	0.00
58 T	2-Chloroethyl Vinyl ether	0.151	0.132	12.6	122	0.00
59 T	2-Hexanone	0.178	0.178	0.0	120	0.00
60 T	Dibromochloromethane	0.309	0.304	1.6	120	0.00
61 T	1,2-Dibromoethane	0.254	0.248	2.4	121	0.00
62 S	4-Bromofluorobenzene	0.359	0.334	7.0	121	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	124	0.00
64 T	Tetrachloroethene	0.492	0.422	14.2	109	0.00
65 PM	Chlorobenzene	0.997	1.003	-0.6	125	0.00
66 T	1,1,1,2-Tetrachloroethane	0.357	0.355	0.6	123	0.00
67 C	Ethyl Benzene	1.811	1.895	-4.6#	130	0.00
68 T	m/p-Xylenes	0.695	0.709	-2.0	126	0.00
69 T	o-Xylene	0.651	0.657	-0.9	126	0.00
70 T	Styrene	1.082	1.104	-2.0	126	0.00
71 P	Bromoform	0.203	0.203	0.0	119	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	122	0.00
73 T	Isopropylbenzene	3.653	3.875	-6.1	128	0.00
74 T	N-amyl acetate	1.068	1.212	-13.5	132	0.00
75 P	1,1,2,2-Tetrachloroethane	0.676	0.683	-1.0	119	0.00
76 T	1,2,3-Trichloropropane	0.454	0.522	-15.0	134	0.00
77 T	Bromobenzene	0.855	0.843	1.4	119	0.00
78 T	n-propylbenzene	4.226	4.546	-7.6	130	0.00
79 T	2-Chlorotoluene	2.322	2.468	-6.3	128	0.00
80 T	1,3,5-Trimethylbenzene	2.980	3.141	-5.4	127	0.00
81 T	trans-1,4-Dichloro-2-butene	0.246	0.284	-15.4	138	0.00
82 T	4-Chlorotoluene	2.444	2.604	-6.5	130	0.00
83 T	tert-Butylbenzene	2.575	2.681	-4.1	127	0.00
84 T	1,2,4-Trimethylbenzene	2.989	3.130	-4.7	127	0.00
85 T	sec-Butylbenzene	3.545	3.738	-5.4	128	0.00
86 T	p-Isopropyltoluene	3.298	3.396	-3.0	124	0.00
87 T	1,3-Dichlorobenzene	1.596	1.598	-0.1	121	0.00
88 T	1,4-Dichlorobenzene	1.576	1.601	-1.6	122	0.00
89 T	n-Butylbenzene	3.032	3.273	-7.9	129	0.00
90 T	Hexachloroethane	0.229	0.427	-86.5#	206#	0.00
91 T	1,2-Dichlorobenzene	1.432	1.419	0.9	119	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.126	0.117	7.1	116	0.00
93 T	1,2,4-Trichlorobenzene	1.018	0.984	3.3	116	0.00
94 T	Hexachlorobutadiene	0.541	0.529	2.2	117	0.00
95 T	Naphthalene	2.159	2.049	5.1	112	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040121\
Data File : VY004235.D
Acq On : 01 Apr 2021 10:02
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Apr 02 02:06:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 30 06:38:14 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.930	0.863	7.2	115	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6