

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040124\
 Data File : VY017801.D
 Acq On : 01 Apr 2024 11:31
 Operator : SY/MD
 Sample : VY0401SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0401SBL01

Quant Time: Apr 02 03:17:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032924S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 05:58:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	176069	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	323227	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	276594	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	98263	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	95762	62.853	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	125.700%
35) Dibromofluoromethane	7.722	113	103117	55.209	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.420%
50) Toluene-d8	10.185	98	373305	53.506	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	107.020%
62) 4-Bromofluorobenzene	12.489	95	93005	43.814	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	87.620%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040124\
Data File : VY017801.D
Acq On : 01 Apr 2024 11:31
Operator : SY/MD
Sample : VY0401SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0401SBL01

Quant Time: Apr 02 03:17:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032924S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 05:58:13 2024
Response via : Initial Calibration

