

Data Path : Z:\VOASRV\HPCHEM1\MSVOA Y\DATA\VY040220\
 Data File : VY002190.D
 Acq On : 02 Apr 2020 14:31
 Operator : SY/MD
 Sample : L2135-03
 Misc : 7.36G/5ML/MSVOA Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :

Quant Time: Apr 03 03:14:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y032620S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 07:11:05 2020
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.81	168	125097	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.71	114	246432	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.50	117	231216	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.44	152	102958	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.16	65	79623	57.58	ug/l	0.00
Spiked Amount	50.000		Recovery	=	115.16%	
35) Dibromofluoromethane	7.74	113	73145	52.97	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.94%	
50) Toluene-d8	10.19	98	307721	53.65	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.30%	
62) 4-Bromofluorobenzene	12.49	95	109482	54.57	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.14%	
Target Compounds				Qvalue		
16) Acetone	3.97	43	4244	10.923	ug/l	90
17) Carbon Disulfide	4.22	76	7458	1.656	ug/l	97
20) Methylene Chloride	4.75	84	6244	Below Cal	#	88
25) 2-Butanone	7.00	43	602	1.009	ug/l #	72
27) cis-1,2-Dichloroethene	7.01	96	8809	5.181	ug/l	89
31) Cyclohexane	7.80	56	6695	2.324	ug/l #	40
36) 1,1-Dichloropropene	7.81	75	10876	4.666	ug/l #	50
38) Carbon Tetrachloride	7.81	117	12913	6.530	ug/l #	16
39) Methylcyclohexane	9.19	83	35241	11.526	ug/l	97
48) Methyl methacrylate	9.26	41	1512	1.291	ug/l #	33
55) 1,1,2-Trichloroethane	10.63	97	12644	9.462	ug/l #	20
56) Ethyl methacrylate	10.53	69	4205	2.155	ug/l #	1
59) 2-Hexanone	10.81	43	15332	16.168	ug/l #	28
73) Isopropylbenzene	12.33	105	8615	1.128	ug/l	94
74) N-amyl acetate	12.13	43	20896	8.675	ug/l #	53
75) 1,1,2,2-Tetrachloroethane	12.56	83	2931	1.783	ug/l #	71
76) 1,2,3-Trichloropropane	12.49	75	54354	47.195	ug/l #	100
78) n-propylbenzene	12.68	91	15764	1.663	ug/l	99
79) 2-Chlorotoluene	12.68	91	15764	3.029	ug/l #	42
81) trans-1,4-Dichloro-2-buten	12.49	75	54354	91.434	ug/l #	18
82) 4-Chlorotoluene	12.68	91	15764	2.875	ug/l #	42
85) sec-Butylbenzene	13.26	105	35674	4.579	ug/l	90
89) n-Butylbenzene	13.70	91	22163	3.180	ug/l #	74
90) Hexachloroethane	13.98	117	3851	3.055	ug/l #	19
92) 1,2-Dibromo-3-Chloropropan	14.34	75	281	1.067	ug/l #	11
95) Naphthalene	15.25	128	7019	1.557	ug/l #	70


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA Y\DATA\VY040220\
Data File : VY002190.D
Acq On : 02 Apr 2020 14:31
Operator : SY/MD
Sample : L2135-03
Misc : 7.36G/5ML/MSVOA Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :

Quant Time: Apr 03 03:14:35 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_Y\METHODS\82Y032620S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 07:11:05 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.81 min Scan# 996

Delta R.T. 0.00 min

Lab File: VY002190.D

Acq: 02 Apr 2020 14:31


Instrument :
MSVOA_Y
ClientSampled :

Tot Ion:168 Resp: 125097

Ion Ratio Lower Upper

168 100

99 59.8 47.6 71.4



Abundance Ion 167.90 (167.60 to 168.60): VY002190.D (-947) (-)

Ion 98.90 (98.60 to 99.60): VY002190.D (-947) (-)

7.81

60000

50000

40000

30000



20000

10000

0

7.70 7.75 7.80 7.85 7.90

Time-->



#16

Acetone

Concen: 10.923 ug/l

RT: 3.97 min Scan# 366

Delta R.T. 0.01 min

Lab File: VY002190.D


Acq: 02 Apr 2020 14:31

Tgt Ion: 43 Resp: 4244

Ion Ratio Lower Upper

43 100

58 38.7 26.3 39.5



Abundance Ion 43.00 (42.70 to 43.70): VY002190.D (-316) (-)

Ion 58.00 (57.70 to 58.70): VY002190.D (-316) (-)

3.97

1500


1000


500

0

3.90 3.95 4.00 4.05

Time-->









#17
Carbon Disulfide
Concen: 1.656 ug/l
RT: 4.22 min Scan# 407
Delta R.T. 0.01 min
Lab File: VY002190.D
Acq: 02 Apr 2020 14:31




Instrument :
MSVOA_Y
ClientSampleId :


Tot Ion: 76 Resp: 7458
Ion Ratio Lower Upper
76 100
78 10.0 7.3 10.9



#20
Methylene Chloride
Concen: Below Cal
RT: 4.75 min Scan# 495
Delta R.T. 0.02 min
Lab File: VY002190.D
Acq: 02 Apr 2020 14:31

Tgt Ion: 84 Resp: 6244
Ion Ratio Lower Upper
84 100
49 111.0 102.2 153.4
51 31.2 31.2 46.8#
86 69.8 52.6 78.8





#25

2-Butanone

Concen: 1.009 ug/l

RT: 7.00 min Scan# 864

Delta R.T. -0.01 min

Lab File: VY002190.D

Acq: 02 Apr 2020 14:31

Instrument :

MSVOA_Y





ClientSampled :

Tot Ion: 43 Resp: 602

Ion Ratio Lower Upper

43 100

72 39.1 20.2 30.2#



#27

cis-1,2-Dichloroethene

Concen: 5.181 ug/l

RT: 7.01 min Scan# 865

Delta R.T. 0.01 min

Lab File: VY002190.D

Acq: 02 Apr 2020 14:31




Tgt Ion: 96 Resp: 8809


Ion Ratio Lower Upper

96 100

61 126.7 0.0 290.6

98 64.9 0.0 128.0





#31

Cyclohexane

Concen: 2.324 ug/l

RT: 7.80 min Scan# 995

Delta R.T. 0.01 min

Lab File: VY002190.D

Acq: 02 Apr 2020 14:31

Instrument :
MSVOA_Y
ClientSampleId :


Tot Ion: 56 Resp: 6695

Ion Ratio Lower Upper




56 100

69 101.7 25.8 38.8#

84 116.4 67.1 100.7#



Abundance Ion 56.00 (55.70 to 56.70): VY0
Ion 69.00 (68.70 to 69.70): VY0
Ion 84.00 (83.70 to 84.70): VY0



#33

1,2-Dichloroethane-d4

Concen: 57.576 ug/l

RT: 8.16 min Scan# 1054

Delta R.T. 0.00 min

Lab File: VY002190.D


Acq: 02 Apr 2020 14:31

Tgt Ion: 65 Resp: 79623



Ion Ratio Lower Upper


65 100

67 54.6 0.0 112.8



Abundance Ion 64.90 (64.60 to 65.60): VY0
Ion 66.90 (66.60 to 67.60): VY0









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.71 min Scan# 1143
 Delta R.T. 0.00 min
 Lab File: VY002190.D
 Acq: 02 Apr 2020 14:31




Instrument :
 MSVOA_Y
 ClientSampleId :


Tot Ion	Ratio	Lower	Upper
114	100		
63	20.5	0.0	41.4
88	15.6	0.0	33.0



#35
 Dibromofluoromethane
 Concen: 52.973 ug/l
 RT: 7.74 min Scan# 985
 Delta R.T. 0.01 min
 Lab File: VY002190.D
 Acq: 02 Apr 2020 14:31

Tgt Ion	Ratio	Lower	Upper
113	100		
111	103.2	84.4	126.6
192	17.4	14.6	21.8









#36
 1,1-Dichloropropene
 Concen: 4.666 ug/l
 RT: 7.81 min Scan# 996
 Delta R.T. -0.13 min
 Lab File: VY002190.D
 Acq: 02 Apr 2020 14:31




Instrument :
 MSVOA_Y
 ClientSampled :


Tot Ion	75	Resp	10876
Ion Ratio	100	Lower	Upper
75	100		
110	8.5	17.2	51.6#
77	0.5	24.7	37.1#



#38
 Carbon Tetrachloride
 Concen: 6.530 ug/l
 RT: 7.81 min Scan# 996
 Delta R.T. -0.11 min
 Lab File: VY002190.D
 Acq: 02 Apr 2020 14:31

Tgt Ion	117	Resp	12913
Ion Ratio <td>100</td> <td>Lower</td> <td>Upper</td>	100	Lower	Upper
117	100		
119	3.3	74.3	111.5#
121	0.0	23.3	34.9#









#39
Methylcyclohexane
Concen: 11.526 ug/l
RT: 9.19 min Scan# 1223
Delta R.T. 0.00 min
Lab File: VY002190.D
Acq: 02 Apr 2020 14:31




Instrument :
MSVOA_Y
ClientSampled :


Tot Ion: 83 Resp: 35241
Ion Ratio Lower Upper
83 100
55 73.8 61.2 91.8
98 49.8 38.2 57.2



#48
Methyl methacrylate
Concen: 1.291 ug/l
RT: 9.26 min Scan# 1234
Delta R.T. -0.04 min
Lab File: VY002190.D
Acq: 02 Apr 2020 14:31

Tgt Ion: 41 Resp: 1512
Ion Ratio Lower Upper
41 100
69 0.0 71.4 107.0#
39 38.1 41.0 61.4#





#50

Toluene-d8

Concen: 53.646 ug/l

RT: 10.19 min Scan# 1386

Delta R.T. 0.00 min

Lab File: VY002190.D

Acq: 02 Apr 2020 14:31

Instrument :

MSVOA_Y


ClientSampleId :

Tot Ion: 98 Resp: 307721

Ion Ratio Lower Upper

98 100

100 64.7 51.6 77.4



Abundance Ion 98.00 (97.70 to 98.70): VY0

Ion 100.00 (99.70 to 100.70): VY

10.19



150000

100000

50000

0

Time-->



#55

1,1,2-Trichloroethane

Concen: 9.462 ug/l

RT: 10.63 min Scan# 1458

Delta R.T. -0.03 min

Lab File: VY002190.D

Acq: 02 Apr 2020 14:31

Tgt Ion: 97 Resp: 12644


Ion Ratio Lower Upper

97 100

83 7.3 68.2 102.2#

85 1.7 42.9 64.3#

99 0.0 47.9 71.9#



Abundance Ion 96.90 (96.60 to 97.60): VY0

Ion 82.90 (82.60 to 83.60): VY0

Ion 84.90 (84.60 to 85.60): VY0

Ion 98.90 (98.60 to 99.60): VY0

10.63

8000


6000


4000

2000

0

Time-->





#56

Ethyl methacrylate

Concen: 2.155 ug/l

RT: 10.53 min Scan# 1442

Delta R.T. 0.01 min

Lab File: VY002190.D

Acq: 02 Apr 2020 14:31

Instrument :

MSVOA_Y

ClientSampled :





Tot Ion: 69 Resp: 4205

Ion Ratio Lower Upper

69 100

41 190.2 52.8 79.2#

39 96.5 25.0 37.6#



#59

2-Hexanone

Concen: 16.168 ug/l

RT: 10.81 min Scan# 1488

Delta R.T. -0.03 min

Lab File: VY002190.D




Acq: 02 Apr 2020 14:31


Tgt Ion: 43 Resp: 15332

Ion Ratio Lower Upper

43 100

58 3.3 27.9 83.6#





#62

4-Bromofluorobenzene

Concen: 54.574 ug/l

RT: 12.49 min Scan# 1763

Delta R.T. 0.00 min

Lab File: VY002190.D

Acq: 02 Apr 2020 14:31

Instrument :

MSVOA_Y

ClientSampleId :


Tot Ion: 95 Resp: 109482

Ion Ratio Lower Upper

95 100

174 70.7 0.0 143.4




176 67.9 0.0 137.0



Abundance Ion 94.90 (94.60 to 95.60): VY002190.D (-1755) (-)

Ion 173.80 (173.50 to 174.50): VY002190.D (-1755) (-)

Ion 175.80 (175.50 to 176.50): VY002190.D (-1755) (-)



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.50 min Scan# 1601

Delta R.T. 0.00 min

Lab File: VY002190.D

Acq: 02 Apr 2020 14:31


Tgt Ion: 117 Resp: 231216

Ion Ratio Lower Upper

117 100

82 56.4 45.4 68.2



119 32.3 25.3 37.9




Abundance Ion 116.90 (116.60 to 117.60): VY002190.D (-1593) (-)

Ion 82.00 (81.70 to 82.70): VY002190.D (-1593) (-)

Ion 118.90 (118.60 to 119.60): VY002190.D (-1593) (-)





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.44 min Scan# 1919

Delta R.T. 0.01 min

Lab File: VY002190.D

Acq: 02 Apr 2020 14:31

Instrument :

MSVOA_Y

ClientSampleId :


Tot Ion:152 Resp: 102958

Ion Ratio Lower Upper

152 100

115 61.0 31.2 93.6




150 157.0 0.0 354.8



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#73

Isopropylbenzene

Concen: 1.128 ug/l

RT: 12.33 min Scan# 1738

Delta R.T. -0.01 min

Lab File: VY002190.D


Acq: 02 Apr 2020 14:31

Tgt Ion:105 Resp: 8615

Ion Ratio Lower Upper



105 100


120 29.4 13.3 39.8



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V





#74

N-amvl acetate

Concen: 8.675 ug/l

RT: 12.13 min Scan# 1705

Delta R.T. -0.02 min

Lab File: VY002190.D

Acq: 02 Apr 2020 14:31

Instrument :

MSVOA_Y

ClientSampleId :

Tot Ion: 43 Resp: 20896





Ion Ratio Lower Upper

43 100

70 9.5 35.0 52.4#

55 7.2 18.7 28.1#

61 0.0 20.2 30.2#



#75

1,1,2,2-Tetrachloroethane

Concen: 1.783 ug/l

RT: 12.56 min Scan# 1775

Delta R.T. -0.03 min

Lab File: VY002190.D

Acq: 02 Apr 2020 14:31




Tgt Ion: 83 Resp: 2931


Ion Ratio Lower Upper

83 100

131 0.0 4.7 14.1#

85 41.5 32.4 97.0









#76
 1,2,3-Trichloropropane
 Concen: 47.195 ug/l
 RT: 12.49 min Scan# 1763
 Delta R.T. -0.15 min
 Lab File: VY002190.D
 Acq: 02 Apr 2020 14:31




Instrument :
 MSVOA_Y
 ClientSampleId :


Tot Ion: 75 Resp: 54354
 Ion Ratio Lower Upper
 75 100
 77 1.6 0.0 0.0#



#78
 n-propylbenzene
 Concen: 1.663 ug/l
 RT: 12.68 min Scan# 1795
 Delta R.T. 0.00 min
 Lab File: VY002190.D
 Acq: 02 Apr 2020 14:31

Tgt Ion: 91 Resp: 15764
 Ion Ratio Lower Upper
 91 100
 120 22.5 11.1 33.1









#79
 2-Chlorotoluene
 Concen: 3.029 ug/l
 RT: 12.68 min Scan# 1795
 Delta R.T. -0.09 min
 Lab File: VY002190.D
 Acq: 02 Apr 2020 14:31




Instrument :
 MSVOA_Y
 ClientSampleId :


Tot Ion: 91 Resp: 15764
 Ion Ratio Lower Upper
 91 100
 126 0.0 16.4 49.1#



#81
 trans-1,4-Dichloro-2-butene
 Concen: 91.434 ug/l
 RT: 12.49 min Scan# 1763
 Delta R.T. 0.10 min
 Lab File: VY002190.D
 Acq: 02 Apr 2020 14:31

Tgt Ion: 75 Resp: 54354
 Ion Ratio Lower Upper
 75 100
 53 1.1 65.0 97.4#
 89 0.0 33.4 50.0#





#82

4-Chlorotoluene

Concen: 2.875 ug/l

RT: 12.68 min Scan# 1795

Delta R.T. -0.18 min

Lab File: VY002190.D

Acq: 02 Apr 2020 14:31

Instrument :

MSVOA_Y


ClientSampleId :

Tot Ion: 91 Resp: 15764

Ion Ratio Lower Upper

91 100

126 0.0 16.3 48.9#




Abundance Ion 91.00 (90.70 to 91.70): VY002089.D (-1820) (-)

Ion 125.90 (125.60 to 126.60): VY002089.D (-1820) (-)

12.68

Time-->




Abundance

Ion 91.00 (90.70 to 91.70): VY002190.D (-1776) (-)

Ion 125.90 (125.60 to 126.60): VY002190.D (-1776) (-)

12.68

Time-->



#85

sec-Butylbenzene

Concen: 4.579 ug/l

RT: 13.26 min Scan# 1890

Delta R.T. 0.00 min

Lab File: VY002190.D


Acq: 02 Apr 2020 14:31

Tgt Ion: 105 Resp: 35674

Ion Ratio Lower Upper

105 100

134 23.9 9.7 29.1




Abundance Ion 105.00 (104.70 to 105.70): VY002089.D (-1882) (-)

Ion 134.00 (133.70 to 134.70): VY002089.D (-1882) (-)

13.26

Time-->




Abundance

Ion 105.00 (104.70 to 105.70): VY002190.D (-1841) (-)

Ion 134.00 (133.70 to 134.70): VY002190.D (-1841) (-)

13.26





Time-->



#89
n-Butylbenzene
Concen: 3.180 ug/l
RT: 13.70 min Scan# 1963
Delta R.T. 0.00 min
Lab File: VY002190.D
Acq: 02 Apr 2020 14:31




Instrument :
MSVOA_Y
ClientSampleId :


Tot Ion: 91 Resp: 22163
Ion Ratio Lower Upper
91 100
92 46.4 27.0 81.0
134 53.5 12.3 36.8#



#90
Hexachloroethane
Concen: 3.055 ug/l
RT: 13.98 min Scan# 2009
Delta R.T. 0.02 min
Lab File: VY002190.D
Acq: 02 Apr 2020 14:31

Tgt Ion: 117 Resp: 3851
Ion Ratio Lower Upper
117 100
201 0.0 31.6 94.8#









#92
 1,2-Dibromo-3-Chloropropane
 Concen: 1.067 ug/l
 RT: 14.34 min Scan# 2068
 Delta R.T. -0.02 min
 Lab File: VY002190.D
 Acq: 02 Apr 2020 14:31

Instrument :
 MSVOA_Y
 ClientSampleId :

Tot Ion: 75 Resp: 281
 Ion Ratio Lower Upper
 75 100
 155 18.1 41.6 124.9#
 157 0.0 53.4 160.3#



#95
 Naphthalene
 Concen: 1.557 ug/l
 RT: 15.25 min Scan# 2216
 Delta R.T. 0.00 min
 Lab File: VY002190.D
 Acq: 02 Apr 2020 14:31

Tgt Ion: 128 Resp: 7019
 Ion Ratio Lower Upper
 128 100
 127 0.0 9.8 14.8#
 129 0.0 8.6 13.0#

