


Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv040220\
 Data File : VY002181.D
 Acq On : 02 Apr 2020 10:32
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y032620S.M
 Title : SW846 8260
 Last Update : Fri Mar 27 07:11:05 2020



AutoFind: Scans 1764, 1765, 1766; Background Corrected with Scan 1755

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.5	28322	PASS
75	95	30	60	51.8	75272	PASS
95	95	100	100	100.0	145405	PASS
96	95	5	9	6.7	9724	PASS
173	174	0.00	2	0.9	982	PASS
174	95	50	100	72.0	104666	PASS
175	174	5	9	7.4	7752	PASS
176	174	95	101	97.2	101738	PASS
177	176	5	9	6.9	7049	PASS