

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040221\
 Data File : VY004266.D
 Acq On : 02 Apr 2021 16:35
 Operator : SY/MD
 Sample : M1909-02
 Misc : 5.96g/5mL/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 FS-SB-10-26-27

Quant Time: Apr 02 23:01:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 30 06:38:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	184336	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.703	114	332380	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.501	117	288008	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	167859	50.00	ug/l	# 0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.179	65	87578	49.42	ug/l	0.03
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.84%
35) Dibromofluoromethane	7.734	113	92432	49.91	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.82%
50) Toluene-d8	10.197	98	757751	108.79	ug/l	0.01
Spiked Amount	50.000	Range	49 - 140	Recovery	=	217.58%#
62) 4-Bromofluorobenzene	12.489	95	273146	114.42	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	228.84%#

Target Compounds						Qvalue
4) Vinyl Chloride	2.229	62	7275	3.03	ug/l	# 42
11) Tert butyl alcohol	4.978	59	523	3.63	ug/l	# 71
13) Acrolein	3.710	56	12445	92.37	ug/l	# 1
14) Allyl chloride	4.606	41	155045	58.36	ug/l	# 58
15) Acrylonitrile	5.179	53	2380	4.99	ug/l	# 47
16) Acetone	3.960	43	42241	114.42	ug/l	# 75
17) Carbon Disulfide	4.210	76	38656	8.82	ug/l	97
18) Methyl Acetate	4.576	43	56370	47.70	ug/l	# 52
20) Methylene Chloride	4.746	84	7221	1.55	ug/l	# 18
23) Vinyl Acetate	6.051	43	41283	10.52	ug/l	# 74
24) 1,1-Dichloroethane	6.057	63	3290	1.03	ug/l	# 82
25) 2-Butanone	7.002	43	28642	46.44	ug/l	# 62
28) Bromochloromethane	7.398	49	1293	1.00	ug/l	# 13
29) Tetrahydrofuran	7.368	42	3656	8.41	ug/l	# 33
30) Chloroform	7.539	83	28755	8.50	ug/l	# 1
31) Cyclohexane	7.782	56	1672700	534.11	ug/l	# 45
36) 1,1-Dichloropropene	8.191	75	1403200	462.47	ug/l	# 1
37) Ethyl Acetate	7.002	43	28642	16.09	ug/l	# 70
38) Carbon Tetrachloride	7.807	117	18659	7.07	ug/l	# 17
39) Methylcyclohexane	9.197	83	6538845	1740.56	ug/l	94
41) Methacrylonitrile	7.374	41	30790	31.22	ug/l	# 80
42) 1,2-Dichloroethane	8.179	62	859814	318.09	ug/l	# 74
43) Isopropyl Acetate	8.282	43	482292	143.11	ug/l	# 54
45) 1,2-Dichloropropane	9.258	63	7940	3.74	ug/l	# 1
46) Dibromomethane	9.270	93	12708	10.17	ug/l	# 1
47) Bromodichloromethane	9.617	83	62712	21.03	ug/l	# 44
48) Methyl methacrylate	9.252	41	450471	294.30	ug/l	# 39
49) 1,4-Dioxane	9.319	88	96	Below	Cal	# 29
51) 4-Methyl-2-Pentanone	9.977	43	4206115	2449.57	ug/l	# 42
53) t-1,3-Dichloropropene	10.258	75	329367	103.59	ug/l	# 1
55) 1,1,2-Trichloroethane	10.630	97	856658	500.35	ug/l	# 20
56) Ethyl methacrylate	10.538	69	348699	148.43	ug/l	# 1
58) 2-Chloroethyl Vinyl ether	9.776	63	14715	14.64	ug/l	# 45
59) 2-Hexanone	10.812	43	1478045	1252.09	ug/l	# 29
60) Dibromochloromethane	10.916	129	3618	1.76	ug/l	61
61) 1,2-Dibromoethane	11.099	107	3878	2.29	ug/l	# 2

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040221\
 Data File : VY004266.D
 Acq On : 02 Apr 2021 16:35
 Operator : SY/MD
 Sample : M1909-02
 Misc : 5.96g/5mL/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 FS-SB-10-26-27


Quant Time: Apr 02 23:01:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 30 06:38:14 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

65) Chlorobenzene	11.526	112	6340	1.10	ug/l	90
68) m/p-Xylenes	11.721	106	17952956	4482.64	ug/l #	1
69) o-Xylene	12.050	106	21881116	5839.08	ug/l #	1
70) Styrene	12.068	104	1781202	285.70	ug/l #	1
73) Isopropylbenzene	12.343	105	8938871	728.91	ug/l	98
74) N-ethyl acetate	12.135	43	903742	252.15	ug/l #	67
75) 1,1,2,2-Tetrachloroethane	12.538	83	52427	23.10	ug/l #	32
76) 1,2,3-Trichloropropane	12.611	75	14890	9.77	ug/l #	100
78) n-propylbenzene	12.684	91	12160307	857.06	ug/l	96
79) 2-Chlorotoluene	12.745	91	3817158	489.69	ug/l #	43
80) 1,3,5-Trimethylbenzene	12.824	105	16711084	1670.44	ug/l	89
81) trans-1,4-Dichloro-2-b...	12.343	75	116430	140.78	ug/l #	28
82) 4-Chlorotoluene	12.824	91	2493710	303.94	ug/l #	39
83) tert-Butylbenzene	13.135	119	5671371	656.08	ug/l #	59
84) 1,2,4-Trimethylbenzene	12.983	105	10040945	1000.65	ug/l	80
85) sec-Butylbenzene	13.263	105	614985	51.68	ug/l #	50
86) p-Isopropyltoluene	13.324	119	2632358	237.78	ug/l	97
89) n-Butylbenzene	13.714	91	1120070	110.03	ug/l #	66
90) Hexachloroethane	13.970	117	1130899	990.47	ug/l #	12
91) 1,2-Dichlorobenzene	13.794	146	13027	2.71	ug/l #	49
92) 1,2-Dibromo-3-Chloropr...	14.342	75	27568	65.09	ug/l #	3
95) Naphthalene	15.238	128	21048815	2903.64	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Apr 02 23:01:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 30 06:38:14 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 7.807 min Scan# 982
Delta R.T. 0.006 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35




Instrument :
MSVOA_Y
ClientSampleId :
FS-SB-10-26-27


Tgt Ion:168 Resp: 184336
Ion Ratio Lower Upper
168 100
99 57.7 42.6 64.0



#4
Vinyl Chloride
Concen: 3.03 ug/l
RT: 2.229 min Scan# 67
Delta R.T. -0.025 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35

Tgt Ion: 62 Resp: 7275
Ion Ratio Lower Upper
62 100
64 0.0 26.0 39.0#









#11
Tert butyl alcohol
Concen: 3.63 ug/l
RT: 4.978 min Scan# 518
Delta R.T. -0.000 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35




Instrument :
MSVOA_Y
ClientSampleId :
FS-SB-10-26-27


Tgt Ion: 59 Resp: 523
Ion Ratio Lower Upper
59 100
57 0.0 8.9 13.3#



#13
Acrolein
Concen: 92.37 ug/l
RT: 3.710 min Scan# 310
Delta R.T. -0.025 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35

Tgt Ion: 56 Resp: 12445
Ion Ratio Lower Upper
56 100
55 2152.9 55.0 82.4#





#14

Allyl chloride

Concen: 58.36 ug/l

RT: 4.606 min Scan# 457

Delta R.T. 0.115 min

Lab File: VY004266.D

Acq: 02 Apr 2021 16:35





Instrument :

MSVOA_Y

ClientSampleId :

FS-SB-10-26-27

Tgt Ion:	41	Resp:	155045
Ion	Ratio	Lower	Upper
41	100		
39	102.4	60.2	90.4#
76	0.0	30.3	45.5#



#15

Acrylonitrile

Concen: 4.99 ug/l




RT: 5.179 min Scan# 551


Delta R.T. -0.000 min

Lab File: VY004266.D

Acq: 02 Apr 2021 16:35

Tgt Ion:	53	Resp:	2380
Ion	Ratio	Lower	Upper
53	100		
52	15.9	65.3	97.9#
51	33.0	29.8	44.8









#16
Acetone
Concen: 114.42 ug/l
RT: 3.960 min Scan# 351
Delta R.T. 0.006 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35




Instrument :
MSVOA_Y
ClientSampleId :
FS-SB-10-26-27


Tgt Ion: 43 Resp: 42241
Ion Ratio Lower Upper
43 100
58 16.9 24.5 36.7#



#17
Carbon Disulfide
Concen: 8.82 ug/l
RT: 4.210 min Scan# 392
Delta R.T. 0.006 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35


Tgt Ion: 76 Resp: 38656
Ion Ratio Lower Upper
76 100
78 8.3 7.4 11.2








#18
Methyl Acetate
Concen: 47.70 ug/l
RT: 4.576 min Scan# 452
Delta R.T. 0.085 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35

Instrument :
MSVOA_Y
ClientSampleId :
FS-SB-10-26-27







Tgt Ion: 43 Resp: 56370
Ion Ratio Lower Upper
43 100
74 0.1 19.1 28.7#



#20
Methylene Chloride
Concen: 1.55 ug/l
RT: 4.746 min Scan# 480
Delta R.T. 0.011 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35

Tgt Ion: 84 Resp: 7221
Ion Ratio Lower Upper
84 100
49 128.3 101.5 152.3
51 105.9 31.5 47.3#
86 239.4 50.8 76.2#









#23
 Vinyl Acetate
 Concen: 10.52 ug/l
 RT: 6.051 min Scan# 694
 Delta R.T. -0.025 min
 Lab File: VY004266.D
 Acq: 02 Apr 2021 16:35




Instrument :
 MSVOA_Y
 ClientSampleId :
 FS-SB-10-26-27


Tgt Ion: 43 Resp: 41283
 Ion Ratio Lower Upper
 43 100
 86 1.0 8.5 12.7#



#24
 1,1-Dichloroethane
 Concen: 1.03 ug/l
 RT: 6.057 min Scan# 695
 Delta R.T. 0.024 min
 Lab File: VY004266.D
 Acq: 02 Apr 2021 16:35

Tgt Ion: 63 Resp: 3290
 Ion Ratio Lower Upper
 63 100
 98 0.0 3.5 10.4#
 100 0.0 2.3 6.9#





#25

2-Butanone

Concen: 46.44 ug/l

RT: 7.002 min Scan# 850

Delta R.T. 0.006 min

Lab File: VY004266.D

Acq: 02 Apr 2021 16:35

Instrument :

MSVOA_Y

ClientSampleId :





FS-SB-10-26-27

Tgt Ion: 43 Resp: 28642

Ion Ratio Lower Upper

43 100

72 6.8 21.0 31.6#



#28

Bromochloromethane

Concen: 1.00 ug/l

RT: 7.398 min Scan# 915

Delta R.T. 0.054 min

Lab File: VY004266.D

Acq: 02 Apr 2021 16:35




Tgt Ion: 49 Resp: 1293


Ion Ratio Lower Upper

49 100

129 0.0 0.0 4.2

130 0.0 62.5 93.7#





#29

Tetrahydrofuran

Concen: 8.41 ug/l

RT: 7.368 min Scan# 910





Delta R.T. 0.012 min

Lab File: VY004266.D

Acq: 02 Apr 2021 16:35

Tgt Ion:	42	Resp:	3656
Ion Ratio	Lower	Upper	
42	100		
72	0.0	35.1	52.7#
71	0.0	32.9	49.3#

Instrument :
MSVOA_Y
ClientSampled :
FS-SB-10-26-27



#30

Chloroform

Concen: 8.50 ug/l




RT: 7.539 min Scan# 938


Delta R.T. 0.018 min

Lab File: VY004266.D

Acq: 02 Apr 2021 16:35


Tgt Ion:	83	Resp:	28755
Ion Ratio	Lower	Upper	
83	100		
85	162.1	51.7	77.5#








#31
Cyclohexane
Concen: 534.11 ug/l
RT: 7.782 min Scan# 978
Delta R.T. -0.007 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35


Instrument :
MSVOA_Y
ClientSampleId :
FS-SB-10-26-27





Tgt Ion: 56 Resp: 1672700
Ion Ratio Lower Upper
56 100
69 7.0 26.6 40.0#
84 35.3 72.3 108.5#




#33
1,2-Dichloroethane-d4
Concen: 49.42 ug/l
RT: 8.179 min Scan# 1043
Delta R.T. 0.030 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35



Tgt Ion: 65 Resp: 87578
Ion Ratio Lower Upper
65 100
67 0.0 0.0 103.2









#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 8.703 min Scan# 1129
Delta R.T. 0.006 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35




Instrument :
MSVOA_Y
ClientSampleId :
FS-SB-10-26-27


Tgt Ion:114 Resp: 332380
Ion Ratio Lower Upper
114 100
63 24.3 0.0 40.4
88 15.5 0.0 30.8



#35
Dibromofluoromethane
Concen: 49.91 ug/l
RT: 7.734 min Scan# 970
Delta R.T. 0.006 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35

Tgt Ion:113 Resp: 92432
Ion Ratio Lower Upper
113 100
111 103.8 82.6 124.0
192 19.6 16.4 24.6









#36
1,1-Dichloropropene
Concen: 462.47 ug/l
RT: 8.191 min Scan# 1045
Delta R.T. 0.262 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35




Instrument :
MSVOA_Y
ClientSampleId :
FS-SB-10-26-27


Tgt Ion: 75 Resp: 1403200
Ion Ratio Lower Upper
75 100
110 0.0 18.1 54.3#
77 1102.0 25.1 37.7#



#37
Ethyl Acetate
Concen: 16.09 ug/l
RT: 7.002 min Scan# 850
Delta R.T. -0.086 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35

Tgt Ion: 43 Resp: 28642
Ion Ratio Lower Upper
43 100
61 0.0 10.2 15.2#
70 0.0 8.0 12.0#









#38
Carbon Tetrachloride
Concen: 7.07 ug/l
RT: 7.807 min Scan# 982
Delta R.T. -0.104 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35




Instrument :
MSVOA_Y
ClientSampleId :
FS-SB-10-26-27


Tgt Ion:117 Resp: 18659
Ion Ratio Lower Upper
117 100
119 5.9 75.6 113.4#
121 0.0 24.3 36.5#



#39
Methylcyclohexane
Concen: 1740.56 ug/l
RT: 9.197 min Scan# 1210
Delta R.T. 0.012 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35

Tgt Ion: 83 Resp: 6538845
Ion Ratio Lower Upper
83 100
55 81.7 60.0 90.0
98 48.1 40.2 60.2









#41
Methacrylonitrile
Concen: 31.22 ug/l
RT: 7.374 min Scan# 911
Delta R.T. 0.042 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35




Instrument :
MSVOA_Y
ClientSampled :
FS-SB-10-26-27


Tgt Ion: 41 Resp: 30790
Ion Ratio Lower Upper
41 100
39 106.3 103.4 155.2
67 0.0 0.0 0.0
52 28.1 0.0 0.0#



#42
1,2-Dichloroethane
Concen: 318.09 ug/l
RT: 8.179 min Scan# 1043
Delta R.T. -0.068 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35

Tgt Ion: 62 Resp: 859814
Ion Ratio Lower Upper
62 100
98 0.0 0.0 18.6









#43
Isopropyl Acetate
Concen: 143.11 ug/l
RT: 8.282 min Scan# 1060
Delta R.T. -0.000 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35




Instrument :
MSVOA_Y
ClientSampleId :
FS-SB-10-26-27


Tgt Ion: 43 Resp: 482292
Ion Ratio Lower Upper
43 100
61 0.0 21.9 32.9#
87 0.0 10.6 15.8#



#45
1,2-Dichloropropane
Concen: 3.74 ug/l
RT: 9.258 min Scan# 1220
Delta R.T. 0.036 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35

Tgt Ion: 63 Resp: 7940
Ion Ratio Lower Upper
63 100
65 272.7 23.7 35.5#





#46

Dibromomethane

Concen: 10.17 ug/l

RT: 9.270 min Scan# 1222

Delta R.T. -0.043 min

Lab File: VY004266.D

Acq: 02 Apr 2021 16:35

Tgt Ion: 93 Resp: 12708

Ion Ratio Lower Upper

93 100

95 1411.5 64.6 97.0#





174 0.0 82.7 124.1#

Instrument :

MSVOA_Y

ClientSampleId :

FS-SB-10-26-27



#47

Bromodichloromethane

Concen: 21.03 ug/l

RT: 9.617 min Scan# 1279

Delta R.T. 0.115 min

Lab File: VY004266.D

Acq: 02 Apr 2021 16:35




Tgt Ion: 83 Resp: 62712


Ion Ratio Lower Upper

83 100

85 16.0 50.3 75.5#

127 0.0 6.7 10.1#





#48

Methyl methacrylate

Concen: 294.30 ug/l

RT: 9.252 min Scan# 1219

Delta R.T. -0.043 min

Lab File: VY004266.D

Acq: 02 Apr 2021 16:35

Instrument :

MSVOA_Y

ClientSampleId :

FS-SB-10-26-27





Tgt Ion: 41 Resp: 450471

Ion Ratio Lower Upper

41 100

69 0.0 67.0 100.4#

39 48.2 49.0 73.6#



#50

Toluene-d8

Concen: 108.79 ug/l

RT: 10.197 min Scan# 1374

Delta R.T. 0.012 min

Lab File: VY004266.D




Acq: 02 Apr 2021 16:35


Tgt Ion: 98 Resp: 757751

Ion Ratio Lower Upper

98 100

100 34.3 52.5 78.7#









#51
4-Methyl-2-Pentanone
Concen: 2449.57 ug/l
RT: 9.977 min Scan# 1338
Delta R.T. -0.098 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35




Instrument :
MSVOA_Y
ClientSampleId :
FS-SB-10-26-27


Tgt Ion: 43 Resp: 4206115
Ion Ratio Lower Upper
43 100
58 3.4 30.6 45.8#



#53
t-1,3-Dichloropropene
Concen: 103.59 ug/l
RT: 10.258 min Scan# 1384
Delta R.T. -0.214 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35


Tgt Ion: 75 Resp: 329367
Ion Ratio Lower Upper
75 100
77 154.2 24.4 36.6#








#55
1,1,2-Trichloroethane
Concen: 500.35 ug/l
RT: 10.630 min Scan# 1445
Delta R.T. -0.019 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35

Instrument :
MSVOA_Y
ClientSampleId :
FS-SB-10-26-27







Tgt Ion: 97 Resp: 856658
Ion Ratio Lower Upper
97 100
83 10.1 70.6 105.8#
85 1.6 44.2 66.4#
99 0.3 50.5 75.7#



#56
Ethyl methacrylate
Concen: 148.43 ug/l
RT: 10.538 min Scan# 1430
Delta R.T. 0.024 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35

Tgt Ion: 69 Resp: 348699
Ion Ratio Lower Upper
69 100
41 203.4 54.8 82.2#
39 107.9 33.4 50.0#









#58
2-Chloroethyl Vinyl ether
Concen: 14.64 ug/l
RT: 9.776 min Scan# 1305
Delta R.T. -0.013 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35




Instrument :
MSVOA_Y
ClientSampleId :
FS-SB-10-26-27


Tgt Ion: 63 Resp: 14715
Ion Ratio Lower Upper
63 100
106 0.0 23.8 35.8#



#59
2-Hexanone
Concen: 1252.09 ug/l
RT: 10.812 min Scan# 1475
Delta R.T. -0.025 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35

Tgt Ion: 43 Resp: 1478045
Ion Ratio Lower Upper
43 100
58 3.0 26.7 80.1#









#60
Dibromochloromethane
Concen: 1.76 ug/l
RT: 10.916 min Scan# 1492
Delta R.T. -0.074 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35




Instrument :
MSVOA_Y
ClientSampleId :
FS-SB-10-26-27


Tgt Ion:129 Resp: 3618
Ion Ratio Lower Upper
129 100
127 43.9 38.9 116.7



#61
1,2-Dibromoethane
Concen: 2.29 ug/l
RT: 11.099 min Scan# 1522
Delta R.T. 0.006 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35

Tgt Ion:107 Resp: 3878
Ion Ratio Lower Upper
107 100
109 0.0 75.6 113.4#









#62
4-Bromofluorobenzene
Concen: 114.42 ug/l
RT: 12.489 min Scan# 1750
Delta R.T. 0.006 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35




Instrument :
MSVOA_Y
ClientSampleId :
FS-SB-10-26-27


Tgt Ion: 95 Resp: 273146
Ion Ratio Lower Upper
95 100
174 41.0 0.0 171.4
176 40.9 0.0 165.6



#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.501 min Scan# 1588
Delta R.T. 0.006 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35

Tgt Ion: 117 Resp: 288008
Ion Ratio Lower Upper
117 100
82 63.7 42.7 64.1
119 31.9 25.7 38.5









#65
Chlorobenzene
Concen: 1.10 ug/l
RT: 11.526 min Scan# 1592
Delta R.T. 0.006 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35




Instrument :
MSVOA_Y
ClientSampleId :
FS-SB-10-26-27


Tgt Ion:112 Resp: 6340
Ion Ratio Lower Upper
112 100
114 26.9 25.9 38.9



#68
m/p-Xylenes
Concen: 4482.64 ug/l
RT: 11.721 min Scan# 1624
Delta R.T. 0.018 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35

Tgt Ion:106 Resp:17952956
Ion Ratio Lower Upper
106 100
91 0.0 159.8 239.6#









#69
o-Xylene
Concen: 5839.08 ug/l
RT: 12.050 min Scan# 1678
Delta R.T. 0.018 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35




Instrument :
MSVOA_Y
ClientSampleId :
FS-SB-10-26-27


Tgt Ion:106 Resp:21881116
Ion Ratio Lower Upper
106 100
91 0.0 105.6 316.8#



#70
Styrene
Concen: 285.70 ug/l
RT: 12.068 min Scan# 1681
Delta R.T. 0.024 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35

Tgt Ion:104 Resp: 1781202
Ion Ratio Lower Upper
104 100
78 238.5 39.8 59.8#
103 216.6 43.3 64.9#









#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.434 min Scan# 1905
Delta R.T. 0.006 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35




Instrument :
MSVOA_Y
ClientSampleId :
FS-SB-10-26-27


Tgt Ion:152 Resp: 167859
Ion Ratio Lower Upper
152 100
115 0.0 29.0 87.1#
150 129.8 0.0 347.4



#73
Isopropylbenzene
Concen: 728.91 ug/l
RT: 12.343 min Scan# 1726
Delta R.T. 0.012 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35

Tgt Ion:105 Resp: 8938871
Ion Ratio Lower Upper
105 100
120 28.3 13.7 41.0









#74
N-amy1 acetate
Concen: 252.15 ug/l
RT: 12.135 min Scan# 1692
Delta R.T. -0.013 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35

Instrument :
MSVOA_Y
ClientSampleId :
FS-SB-10-26-27

Tgt Ion: 43 Resp: 903742




Ion	Ratio	Lower	Upper
43	100		
70	25.2	32.6	49.0#
55	37.9	18.0	27.0#
61	0.0	18.7	28.1#




#75
1,1,2,2-Tetrachloroethane
Concen: 23.10 ug/l
RT: 12.538 min Scan# 1758
Delta R.T. -0.049 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35

Tgt Ion: 83 Resp: 52427

Ion	Ratio	Lower	Upper
83	100		
131	0.0	5.6	16.8#
85	121.8	31.8	95.4#









#76
1,2,3-Trichloropropane
Concen: 9.77 ug/l
RT: 12.611 min Scan# 1770
Delta R.T. -0.025 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35




Instrument :
MSVOA_Y
ClientSampleId :
FS-SB-10-26-27


Tgt Ion: 75 Resp: 14890
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0



#78
n-propylbenzene
Concen: 857.06 ug/l
RT: 12.684 min Scan# 1782
Delta R.T. 0.012 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35

Tgt Ion: 91 Resp:12160307
Ion Ratio Lower Upper
91 100
120 25.8 11.9 35.5









#79
2-Chlorotoluene
Concen: 489.69 ug/l
RT: 12.745 min Scan# 1792
Delta R.T. -0.013 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35




Instrument :
MSVOA_Y
ClientSampleId :
FS-SB-10-26-27


Tgt Ion: 91 Resp: 3817158
Ion Ratio Lower Upper
91 100
126 2.1 17.5 52.6#



#80
1,3,5-Trimethylbenzene
Concen: 1670.44 ug/l
RT: 12.824 min Scan# 1805
Delta R.T. 0.012 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35

Tgt Ion: 105 Resp: 16711084
Ion Ratio Lower Upper
105 100
120 57.4 24.9 74.7









#81
trans-1,4-Dichloro-2-butene
Concen: 140.78 ug/l
RT: 12.343 min Scan# 1726
Delta R.T. -0.037 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35




Instrument :
MSVOA_Y
ClientSampleId :
FS-SB-10-26-27


Tgt Ion: 75 Resp: 116430
Ion Ratio Lower Upper
75 100
53 159.8 79.3 118.9#
89 113.0 39.4 59.2#



#82
4-Chlorotoluene
Concen: 303.94 ug/l
RT: 12.824 min Scan# 1805
Delta R.T. -0.031 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35

Tgt Ion: 91 Resp: 2493710
Ion Ratio Lower Upper
91 100
126 0.0 17.8 53.4#









#83
tert-Butylbenzene
Concen: 656.08 ug/l
RT: 13.135 min Scan# 1856
Delta R.T. 0.060 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35




Instrument :
MSVOA_Y
ClientSampleId :
FS-SB-10-26-27


Tgt Ion:119 Resp: 5671371
Ion Ratio Lower Upper
119 100
91 94.8 33.1 99.2
134 0.0 13.6 40.8#



#84
1,2,4-Trimethylbenzene
Concen: 1000.65 ug/l
RT: 12.983 min Scan# 1831
Delta R.T. -0.141 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35

Tgt Ion:105 Resp:10040945
Ion Ratio Lower Upper
105 100
120 32.9 22.9 68.7









#85
 sec-Butylbenzene
 Concen: 51.68 ug/l
 RT: 13.263 min Scan# 1877
 Delta R.T. 0.006 min
 Lab File: VY004266.D
 Acq: 02 Apr 2021 16:35




Instrument :
 MSVOA_Y
 ClientSampleId :
 FS-SB-10-26-27


Tgt Ion:105 Resp: 614985
 Ion Ratio Lower Upper
 105 100
 134 44.4 10.4 31.4#



#86
 p-Isopropyltoluene
 Concen: 237.78 ug/l
 RT: 13.324 min Scan# 1887
 Delta R.T. -0.049 min
 Lab File: VY004266.D
 Acq: 02 Apr 2021 16:35

Tgt Ion:119 Resp: 2632358
 Ion Ratio Lower Upper
 119 100
 134 27.3 13.3 39.9
 91 26.4 12.0 36.0









#89
n-Butylbenzene
Concen: 110.03 ug/l
RT: 13.714 min Scan# 1951
Delta R.T. 0.018 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35




Instrument :
MSVOA_Y
ClientSampleId :
FS-SB-10-26-27


Tgt Ion: 91 Resp: 1120070
Ion Ratio Lower Upper
91 100
92 70.3 26.5 79.4
134 0.0 13.6 40.7#



#90
Hexachloroethane
Concen: 990.47 ug/l
RT: 13.970 min Scan# 1993
Delta R.T. 0.005 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35

Tgt Ion: 117 Resp: 1130899
Ion Ratio Lower Upper
117 100
201 0.0 37.6 112.9#









#91
1,2-Dichlorobenzene
Concen: 2.71 ug/l
RT: 13.794 min Scan# 1964
Delta R.T. 0.054 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35




Instrument :
MSVOA_Y
ClientSampleId :
FS-SB-10-26-27


Tgt Ion:146 Resp: 13027
Ion Ratio Lower Upper
146 100
111 0.0 20.1 60.3#
148 30.5 31.6 94.7#



#92
1,2-Dibromo-3-Chloropropane
Concen: 65.09 ug/l
RT: 14.342 min Scan# 2054
Delta R.T. -0.019 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35

Tgt Ion: 75 Resp: 27568
Ion Ratio Lower Upper
75 100
155 12.2 46.3 138.8#
157 0.0 58.3 174.9#





#95
Naphthalene
Concen: 2903.64 ug/l
RT: 15.238 min Scan# 2201
Delta R.T. -0.000 min
Lab File: VY004266.D
Acq: 02 Apr 2021 16:35

Instrument :
MSVOA_Y
ClientSampleId :
FS-SB-10-26-27

Tgt Ion:128 Resp:21048815
Ion Ratio Lower Upper
128 100
127 21.1 10.1 15.1#
129 18.6 8.9 13.3#

