

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040221\
 Data File : VY004252.D
 Acq On : 02 Apr 2021 10:40
 Operator : SY/MD
 Sample : VY0402SBL01
 Misc : 5.00g/5mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0402SBL01

Quant Time: Apr 02 22:56:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 30 06:38:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	206606	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	349885	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	319096	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	142472	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	117161	58.99	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.98%
35) Dibromofluoromethane	7.728	113	105478	54.10	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.20%
50) Toluene-d8	10.185	98	428564	58.45	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	116.90%
62) 4-Bromofluorobenzene	12.489	95	139808	55.63	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	111.26%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040221\
Data File : VY004252.D
Acq On : 02 Apr 2021 10:40
Operator : SY/MD
Sample : VY0402SBL01
Misc : 5.00g/5mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0402SBL01

Quant Time: Apr 02 22:56:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 30 06:38:14 2021
Response via : Initial Calibration

