

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040221\
 Data File : VY004256.D
 Acq On : 02 Apr 2021 12:42
 Operator : SY/MD
 Sample : M1897-01
 Misc : 5.53g/5mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-1

Quant Time: Apr 02 22:58:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 30 06:38:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	159833	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	245917	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	202117	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	85001	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	74942	48.77	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	97.54%
35) Dibromofluoromethane	7.728	113	70187	51.22	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.44%
50) Toluene-d8	10.185	98	293860	57.02	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	114.04%
62) 4-Bromofluorobenzene	12.483	95	86978	49.24	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	98.48%
Target Compounds						
16) Acetone	3.960	43	2636	8.23	ug/l	Qvalue # 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040221\
Data File : VY004256.D
Acq On : 02 Apr 2021 12:42
Operator : SY/MD
Sample : M1897-01
Misc : 5.53g/5mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-1

Quant Time: Apr 02 22:58:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 30 06:38:14 2021
Response via : Initial Calibration

