

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040221\
 Data File : VY004257.D
 Acq On : 02 Apr 2021 13:06
 Operator : SY/MD
 Sample : M1897-03
 Misc : 7.39g/5mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-2

Quant Time: Apr 02 22:59:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 30 06:38:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	191559	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	329828	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	294386	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	114870	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	114908	62.40	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	124.80%
35) Dibromofluoromethane	7.728	113	101170	55.05	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.10%
50) Toluene-d8	10.185	98	398298	57.62	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	115.24%
62) 4-Bromofluorobenzene	12.483	95	127771	53.94	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	107.88%
Target Compounds						
16) Acetone	3.954	43	4531	11.81	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040221\
Data File : VY004257.D
Acq On : 02 Apr 2021 13:06
Operator : SY/MD
Sample : M1897-03
Misc : 7.39g/5mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-2

Quant Time: Apr 02 22:59:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 30 06:38:14 2021
Response via : Initial Calibration

