

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040221\
 Data File : VY004265.D
 Acq On : 02 Apr 2021 16:12
 Operator : SY/MD
 Sample : M1909-01
 Misc : 6.48g/5mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 FS-SB-10-5-10

Quant Time: Apr 02 23:01:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 30 06:38:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	185724	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	317097	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	276298	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	98553	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.154	65	114471	64.12	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 128.24%			
35) Dibromofluoromethane	7.734	113	98266	55.62	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 111.24%			
50) Toluene-d8	10.185	98	384525	57.86	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 115.72%			
62) 4-Bromofluorobenzene	12.483	95	114206	50.14	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 100.28%			
Target Compounds						Qvalue
16) Acetone	3.966	43	6822	18.34	ug/l	92
40) Benzene	8.173	78	73303	8.93	ug/l	98
52) Toluene	10.252	92	8613	1.63	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040221\
Data File : VY004265.D
Acq On : 02 Apr 2021 16:12
Operator : SY/MD
Sample : M1909-01
Misc : 6.48g/5mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
FS-SB-10-5-10

Quant Time: Apr 02 23:01:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032921S.M
Quant Title : SW846 8260
QLast Update : Tue Mar 30 06:38:14 2021
Response via : Initial Calibration

