

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040323\
 Data File : VY013180.D
 Acq On : 03 Apr 2023 16:55
 Operator : KP/MD
 Sample : 02165-05
 Misc : 5.79g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-5B

Quant Time: Apr 03 23:10:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y033023S.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 03 13:40:57 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	91210	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	146056	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	137215	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	65580	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	54871	55.978	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.960%
35) Dibromofluoromethane	7.716	113	49124	51.485	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.980%
50) Toluene-d8	10.179	98	172236	47.817	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	95.640%
62) 4-Bromofluorobenzene	12.483	95	52635	44.893	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	89.780%
Target Compounds						
						Qvalue
16) Acetone	3.942	43	8212	27.397	ug/l	97
20) Methylene Chloride	4.710	84	13129	8.198	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040323\
Data File : VY013180.D
Acq On : 03 Apr 2023 16:55
Operator : KP/MD
Sample : 02165-05
Misc : 5.79g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-5B

Quant Time: Apr 03 23:10:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y033023S.M
Quant Title : SW846 8260
QLast Update : Mon Apr 03 13:40:57 2023
Response via : Initial Calibration

