

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040323\
 Data File : VY013185.D
 Acq On : 03 Apr 2023 18:52
 Operator : KP/MD
 Sample : 02166-05
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HR-03-033123

Quant Time: Apr 03 23:12:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y033023S.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 03 13:40:57 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	3208	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	3430	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	4265	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	1466	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	2406	69.788	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 139.580%	
35) Dibromofluoromethane	7.710	113	1715	76.538	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 153.080%#	
50) Toluene-d8	10.179	98	3907	46.187	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 92.380%	
62) 4-Bromofluorobenzene	12.483	95	1424	51.718	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 103.440%	
Target Compounds						
20) Methylene Chloride	4.716	84	623	13.162	ug/l	# 77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040323\
Data File : VY013185.D
Acq On : 03 Apr 2023 18:52
Operator : KP/MD
Sample : 02166-05
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HR-03-033123

Quant Time: Apr 03 23:12:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y033023S.M
Quant Title : SW846 8260
QLast Update : Mon Apr 03 13:40:57 2023
Response via : Initial Calibration

