



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040323\
Data File : VY013164.D
Acq On : 03 Apr 2023 09:04
Operator : KP/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y033023S.M
Title : SW846 8260
Last Update : Mon Apr 03 13:40:57 2023



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	19.5	12931	PASS
75	95	30	60	51.3	34040	PASS
95	95	100	100	100.0	66392	PASS
96	95	5	9	6.9	4591	PASS
173	174	0.00	2	1.5	865	PASS
174	95	50	100	88.2	58579	PASS
175	174	5	9	7.7	4482	PASS
176	174	95	101	97.0	56843	PASS
177	176	5	9	7.2	4101	PASS