

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040423\
 Data File : VY013189.D
 Acq On : 04 Apr 2023 15:08
 Operator : KP/MD
 Sample : VY0404SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0404SBL01

Quant Time: Apr 05 02:15:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y033023S.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 03 13:40:57 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	73057	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	129172	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	116169	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	51704	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	52212	66.501	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	133.000%
35) Dibromofluoromethane	7.716	113	42821	50.746	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.500%
50) Toluene-d8	10.179	98	152684	47.929	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	95.860%
62) 4-Bromofluorobenzene	12.483	95	46515	44.859	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	89.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040423\
Data File : VY013189.D
Acq On : 04 Apr 2023 15:08
Operator : KP/MD
Sample : VY0404SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0404SBL01

Quant Time: Apr 05 02:15:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y033023S.M
Quant Title : SW846 8260
QLast Update : Mon Apr 03 13:40:57 2023
Response via : Initial Calibration

