

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040423\
 Data File : VY013193.D
 Acq On : 04 Apr 2023 16:41
 Operator : KP/MD
 Sample : 02170-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 EO-3-033123

Quant Time: Apr 05 01:20:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y033023S.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 03 13:40:57 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	5172	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	8091	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	5909	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	2708	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	3737	67.233	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	134.460%
35) Dibromofluoromethane	7.716	113	3198	60.504	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	121.000%
50) Toluene-d8	10.179	98	8970	44.954	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	89.900%
62) 4-Bromofluorobenzene	12.483	95	2475	38.106	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	76.220%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040423\
Data File : VY013193.D
Acq On : 04 Apr 2023 16:41
Operator : KP/MD
Sample : 02170-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
EO-3-033123

Quant Time: Apr 05 01:20:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y033023S.M
Quant Title : SW846 8260
QLast Update : Mon Apr 03 13:40:57 2023
Response via : Initial Calibration

