

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040423\
 Data File : VY013195.D
 Acq On : 04 Apr 2023 17:28
 Operator : KP/MD
 Sample : 02166-01
 Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HR-01-033123

Quant Time: Apr 05 01:20:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y033023S.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 03 13:40:57 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	78674	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	130717	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	128502	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	62955	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	53063	62.760	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	125.520%	
35) Dibromofluoromethane	7.710	113	46476	54.426	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	108.860%	
50) Toluene-d8	10.179	98	156032	48.401	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	96.800%	
62) 4-Bromofluorobenzene	12.477	95	52436	49.971	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	99.940%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.710	84	12384	9.528	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040423\
Data File : VY013195.D
Acq On : 04 Apr 2023 17:28
Operator : KP/MD
Sample : 02166-01
Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HR-01-033123

Quant Time: Apr 05 01:20:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y033023S.M
Quant Title : SW846 8260
QLast Update : Mon Apr 03 13:40:57 2023
Response via : Initial Calibration

