

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040423\
 Data File : VY013197.D
 Acq On : 04 Apr 2023 18:15
 Operator : KP/MD
 Sample : 02166-05
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HR-03-033123

Quant Time: Apr 05 01:21:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y033023S.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 03 13:40:57 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	74829	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	119294	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	120912	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	59895	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	51425	63.948	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	127.900%
35) Dibromofluoromethane	7.710	113	42645	54.722	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.440%
50) Toluene-d8	10.179	98	142056	48.285	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	96.580%
62) 4-Bromofluorobenzene	12.477	95	49371	51.556	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	103.120%
Target Compounds						
20) Methylene Chloride	4.710	84	12561	10.561	ug/l	Qvalue # 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040423\
Data File : VY013197.D
Acq On : 04 Apr 2023 18:15
Operator : KP/MD
Sample : 02166-05
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HR-03-033123

Quant Time: Apr 05 01:21:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y033023S.M
Quant Title : SW846 8260
QLast Update : Mon Apr 03 13:40:57 2023
Response via : Initial Calibration

