

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040424\
 Data File : VY017841.D
 Acq On : 04 Apr 2024 09:32
 Operator : SY/MD
 Sample : VY0404SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0404SBL01

Quant Time: Apr 05 00:42:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032924S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 05:58:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	138450	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.703	114	250696	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.508	117	215439	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.446	152	76506	50.000	ug/l	0.01

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.161	65	72555	60.560	ug/l	0.02
Spiked Amount	50.000	Range	50 - 163	Recovery	=	121.120%
35) Dibromofluoromethane	7.734	113	80827	55.795	ug/l	0.02
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.600%
50) Toluene-d8	10.197	98	293256	54.194	ug/l	0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	=	108.380%
62) 4-Bromofluorobenzene	12.501	95	72182	43.843	ug/l	0.01
Spiked Amount	50.000	Range	30 - 143	Recovery	=	87.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY040424\
Data File : VY017841.D
Acq On : 04 Apr 2024 09:32
Operator : SY/MD
Sample : VY0404SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0404SBL01

Quant Time: Apr 05 00:42:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y032924S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 05:58:13 2024
Response via : Initial Calibration

